



ADVANCED INDUSTRIAL RESOURCES, INC.

***COMPLIANCE TEST REPORT
ETHYLENE OXIDE
STERILIZATION CHAMBER EXHAUSTS
AT
CHEMENCE MEDICAL
PROJECT ID: KR-10590***

PREPARED FOR:



**CHEMENCE[®]
MEDICAL**

**CHEMENCE MEDICAL
200 TECHNOLOGY DRIVE
ALPHARETTA, GEORGIA 30005**

PREPARED BY:

**ADVANCED INDUSTRIAL RESOURCES, INC.
3407 NOVIS POINTE
ACWORTH, GEORGIA 30101**

OCTOBER 14, 2020



REPORT CERTIFICATION SHEET

Having conducted the Technical Review of this report, I hereby certify the data, information, results, and calculations in this report to be accurate and true according to the methods and procedures used.

Ross Winne
Technical Director
Advanced Industrial Resources

November 16, 2020

Date

Having written and prepared this report, I hereby certify that the data, information and results in this report to be correct and all inclusive of the necessary information required for a complete third-party review of the testing event.

Steven Haigh
Report Preparation Director
Advanced Industrial Resources

November 16, 2020

Date

Having supervised all aspects of the field testing, I hereby certify the equipment preparation, field sample collection procedures, and all equipment calibrations were conducted in accordance to the applicable methodologies.

Scott Gunnell
Field Project Supervisor
Advanced Industrial Resources

November 16, 2020

Date

TABLE OF CONTENTS

<u>1.0</u>	<u>INTRODUCTION</u>	<u>1</u>
1.1	SUMMARY OF TEST PROGRAM	1
1.2	KEY PERSONNEL	1
<u>2.0</u>	<u>PLANT AND SAMPLING LOCATION DESCRIPTIONS</u>	<u>2</u>
2.1	PROCESS & CONTROL EQUIPMENT DESCRIPTION	2
2.2	SAMPLING LOCATION	2
<u>3.0</u>	<u>SUMMARY AND DISCUSSION OF TEST RESULTS</u>	<u>3</u>
3.1	OBJECTIVES	3
3.2	FIELD TEST CHANGES AND PROBLEMS	3
3.3	PRESENTATION OF TEST RESULTS	3
3.4	PROCESS MONITORING	4
<u>4.0</u>	<u>SAMPLING AND ANALYTICAL PROCEDURES</u>	<u>5</u>
<u>5.0</u>	<u>DATA QUALITY OBJECTIVES</u>	<u>6</u>

LIST OF APPENDICES

APPENDIX A	TEST RESULTS
APPENDIX B	EXAMPLE CALCULATIONS AND NOMENCLATURE
APPENDIX C	FIELD DATA
APPENDIX D	LABORATORY REPORT
APPENDIX E	CALIBRATION DATA
APPENDIX F	PROCESS OPERATION DATA

1.0 INTRODUCTION

1.1 SUMMARY OF TEST PROGRAM

The Chemence Medical facility is located at 200 Technology Drive, Alpharetta, GA 30005. Chemence manufactures and distributes adhesive-based medical devices used by many of the largest medical centers around the globe. Chemence also operates medical product and device sterilization chambers using ethylene oxide (EO). Source emissions testing was conducted on one each of the two sets of ethylene oxide sterilization chamber exhausts: EOA and EOE. Similar chambers to EOA are EOB and EOC. EOD is a similar chamber to EOE but is not currently operating. This testing was conducted in response to an EPD letter of August 10, 2020 from Sean Taylor and follow-up conversations with Stephen Damaske and Daniel McCain of EPD regarding the ethylene oxide emissions from this source. Note that weekly batch production is between zero and 3 batches.

Testing was conducted on October 14, 2020, in accordance with the generally approved test protocol developed for projects of this type modified for the small duct diameters. All testing was conducted by Advanced Industrial Resources, Inc. (*AIR*) in accordance with approved USEPA Methods (i.e., 40 CFR 60 Appendix A, Methods 1, 2, 3, 4, and Modified TO-15).

1.2 KEY PERSONNEL

The key personnel who coordinated and this Test Report and their telephone numbers are:

Hugo Escobar, <i>Chemence Medical</i>	770-664-7078
Ross Winne, <i>AIR</i>	404-843-2100
Scott Wilson, <i>AIR</i>	800-224-5007

2.0 PLANT AND SAMPLING LOCATION DESCRIPTIONS

2.1 PROCESS & CONTROL EQUIPMENT DESCRIPTION

The Chemence Medical facility is located at 200 Technology Drive, Alpharetta, GA 30005. Plastic sealed tote baskets are used to hold products and devices for EO sterilization in batch processes after loading to the chambers. Batches typically run for 19 hours, followed by a purge cycle lasting 1 hour for EOA, EOB, and EOC, or 4 hours for EOE. During this purge cycle the gases are vented to the outside.

2.2 SAMPLING LOCATION

The exhaust sampling location for EOA was a 3-inch id section of PVC pipe modified to house a calibrated propeller vane anemometer for measuring the flow in ft/min. The A distance was 3 inches and the B distance was 6 inches to disturbances.

The exhaust sampling location for EOE was a 1.625-inch id section of PVC pipe modified to house a calibrated hot wire anemometer for measuring the flow in ft/min. The A distance was 6 inches and the B distance was 12 inches to disturbances.

3.0 SUMMARY AND DISCUSSION OF TEST RESULTS

3.1 OBJECTIVES

The purpose of the test program was to verify source emissions of ethylene oxide (EO) both in ppm and in lb/hr. The molecular weight (Mw) of EO is 44.05.

3.2 FIELD TEST CHANGES AND PROBLEMS

The testing was conducted in accordance with the general test protocol developed for projects of this type scaled to the smaller exhaust sizes. At the request of EPD, a duplicate sample was taken for each chamber utilizing 2 Summa cannisters simultaneously. No problems were encountered during testing that required deviation from the planned test protocol.

3.3 PRESENTATION OF TEST RESULTS

Emission concentrations and mass rates are summarized in Table 3-1. Reduced and tabulated data from the field-testing is included in Appendix A. The calculations and nomenclature used to reduce the data are presented in Appendix B. Actual raw field data sheets are presented in Appendix C.

Table 3-1

Source	Average Measured ppm (EO)	Average Measured lb/hr (EO)
EOA	52 ppm	0.02602589 lb/hr for the 1-hour cycle
EOE	405 ppm	0.06795603 lb/hr for the 4-hour cycle

EOA average emission concentration was determined to be **52 ppm (0.02603 lb/hr.)**
EOE average emission concentration was determined to be **405 ppm (0.06796 lb/hr.)**

3.4 PROCESS MONITORING

All essential process monitoring equipment was operating and data was recorded throughout the test period so as to allow necessary monitoring parameters and limits to be established, where applicable. Data, as made available, is presented in Appendix F.

4.0 SAMPLING AND ANALYTICAL PROCEDURES

Each of the two (2) ethylene oxide (EO) emissions performance tests was comprised of one (1) **duplicate** sampling repetition of four (4) hours for EOE and of one (1) **duplicate** 1-hour for EOA during the purge of each chamber, which followed a nineteen (19) hour sterilization cycle. All testing was according to the methodologies detailed in EPA Methods 1, 2, 3, 4, and Modified TO-15.

Performance testing for this source will be conducted according to the methodology in 40 *CFR* 60, Appendix A. Specifically:

- X Method 1 or 1A was used for the qualification of the location of sampling ports and for the determination of the number and positions of stack traverse points, as applicable to sample traverses.
- X Method 2 or 2A was employed for the determination of the stack gas velocity and volumetric flow rate during stack sampling using the Type “S” pitot tube or a calibrated anemometer.
- X Method 3 was used for the calculation of the density and dry molecular weight of the effluent stack gas.
- X Method 4 was used for the determination of moisture (or wet-bulb/dry-bulb).
- X Method TO-15 Modified SIM (Selective Ion Monitoring) – VOCs by GCMS in SIM Mode (reporting limit 0.5 ppbv). 6-liter summa cannisters to sample for 240-minute batch sample time using 4-hour flow controllers and for 60-minute batch using 1-hour flow controllers. Sample location will be in the centroid of the small diameter exhaust pipes, which are 1.625 inches for EOE and 3.0 inches for EOA.

Summa cannister samples were recovered on site by sealing and tagging them. All samples were stored upright in a closed shipping box until final laboratory analysis. All samples were ground transported via the *AIR* equipment resource vehicle and shipped for analysis to Eurofins Air Toxics, LLC, Folsom, CA. In order to limit the chain of custody, only essential *AIR* personnel were permitted access to these samples.

5.0 DATA QUALITY OBJECTIVES

The data quality objectives (DQOs) process is generally a seven-step iterative planning approach to ensure development of sampling designs for data collection activities that support decision making. The seven steps are as follows: (1) defining the problem; (2) stating decisions and alternative actions; (3) identifying inputs into the decision; (4) defining the study boundaries; (5) defining statistical parameters, specifying action levels, and developing action logic; (6) specifying acceptable error limits; and (7) selecting resource-effective sampling and analysis plan to meet the performance criteria. The first five steps are primarily focused on identifying qualitative criteria such as the type of data needed and defining how the data will be used. The sixth step defines quantitative criteria and the seventh step is used to develop a data collection design. In regards to emissions sampling, these steps have already been identified for typical monitoring parameters.

Monitoring methods presented in 40 *CFR* 60 Appendix A indicate the following regarding DQOs: Adherence to the requirements of this method will enhance the quality of the data obtained from air pollutant sampling methods. At a minimum, each method provides the following types of information: summary of method; equipment and supplies; reagents and standards; sample collection, preservation, storage, and transportation; quality control; calibration and standardization; analytical procedures, data analysis and calculations; and alternative procedures. These test methods have been designed and tested according to DQOs for emissions testing and analysis. These test methods have been specified and were followed to testing to ensure that DQOs were met for this project.

APPENDIX A

TEST RESULTS

Advanced Industrial Resources, Inc.

Test Results

Chemance Medical
Alpharetta, Georgia

		Units	EOA	EOA - Duplicate	EOE	EOE - Duplicate
Test Date			14-Oct-20	14-Oct-20	14-Oct-20	14-Oct-20
Start Time			10:33	10:33	10:22	10:22
End Time			11:33	11:33	14:22	14:22
P _{bar}	Barometric pressure	inches Hg	29.12	29.12	29.12	29.12
p _g	Gauge pressure of stack gas	inches H ₂ O	0.00	0.00	0.00	0.00
P _s	Pressure of stack gases	inches Hg	29.12	29.12	29.12	29.12
B _{ws,act}	Actual moisture	%	0.0175	0.0175	0.0181	0.0181
v _s	Velocity of stack gas	ft./min	1,664	1,664	1,915	1,915
v _s	Velocity of stack gas	ft./sec	28	28	32	32
D _s	Diameter of stack	in	3.00	3.00	1.625	1.625
A _s	Area of stack	ft ²	0.0491	0.0491	0.0144	0.0144
T _s	Temperature of stack gas	°R	565	565	569	569
Gas Stream Flow Rates						
Q _a	Vol. Flow rate of actual gas	cfm	82	82	28	28
Q _w	Vol. Flow rate of wet gas	scfm	74	74	25	25
Q _{sd}	Vol. Flow rate of dry gas	dscfm	73	73	24	24
Gas Stream Ethylene Oxide Concentrations - TO-15 SIM						
c _{EtO}	Conc. Of PM in dry stack gas	ppm	58.0	46.0	360.0	450.0
c _{EtO}	Conc. Of PM in dry stack gas	mg/dscm	106.2	84.3	659.4	824.2
c _{PM}	Conc. Of PM in dry stack gas	gr/dscf	0.0464	0.0368	0.2880	0.3600
Ethylene Oxide Mass Rates TO-15 SIM						
E _{EtO}	Emission rate of EtO	lb/hour	0.029	0.023	0.060	0.076

	ppm	ppm	lb/hr	lb/hr	Averages
EOA	58	58	0.02900615	0.02900615	0.02600551
EOA Dup	46	46	0.02300487	0.02300487	
EOE	360	360	0.06035624	0.06035624	0.06790077
EOE Dup	450	450	0.0754453	0.0754453	

APPENDIX B

EXAMPLE CALCULATIONS

AND

NOMENCLATURE

EXAMPLE CALCULATIONS

$$A_n = D_n^2 \pi / 4$$

$$A_s = D_s^2 \pi / 4$$

$$B_{ws} = V_{w(std)} / (V_{m(std)} + V_{w(std)})$$

$$c_{\text{analyte}} = (m_{\text{analyte}} / V_{m(std)}) (35.31466 \text{ ft}^3/\text{m}^3)$$

$$c'_{\text{analyte}} = (m_{\text{analyte}} / V_{m(std)}) (0.015432 \text{ gr/mg})$$

$$c_{\text{analyte}} = c'_{\text{analyte}} \text{MW}_{\text{analyte}} / 24.04 \text{ l/mol}$$

$$CC = t_{0.975} (S_d / n^{1/2})$$

$$d = 1/n (S_{di})$$

$$DE = (E_{\text{Inlet}} - E_{\text{Outlet}}) / E_{\text{Inlet}} \times 100\%$$

$$E_{\text{analyte}} = (m_{\text{analyte}} / V_{m(std)}) Q_{sd} (60 \text{ min/hr}) (2.2046 \times 10^{-6} \text{ lb./mg})$$

$$E_{\text{analyte}} = c_{\text{analyte}} Q_{sd} (60 \text{ min/hr}) (2.2046 \times 10^{-6} \text{ lb./mg})$$

$$I = 100 T_s (K_3 V_{lc} + Y_m V_m P_m / T_m) / (60 \theta v_s P_s A_n)$$

$$\text{where } K_3 = 0.002669 (\text{in. Hg ft}^3) / (\text{mL } ^\circ\text{R})$$

$$K_I = [(2.0084 \times 10^7 \Delta H_{@}) A_n (1 - B_{ws})]^2 (M_d / M_s) (T_m / T_s) (P_s / P_m)$$

$$M_d = 0.44 (\% \text{ CO}_2) + 0.32 (\% \text{ O}_2) + 0.28 (\% \text{ N}_2 + \% \text{ CO})$$

$$M_s = M_d (1 - B_{ws}) + M_w B_{ws}$$

$$P = Q_{sd} / \text{F-Factor} \times 60 \times (20.9 - \text{O}_2) / 20.9$$

$$P_m = P_{\text{bar}} + \Delta H / 13.6$$

$$P_s = P_{\text{bar}} + p_g / 13.6$$

$$Q_a = (60 \text{ s/min}) v_s A_s$$

$$Q_{sd} = (60 \text{ s/min}) (1 - B_{ws}) v_s A_s (T_{std} / T_s) (P_s / P_{std})$$

$$RA = [\text{Abs}(d) + \text{Abs}(CC)] / RM$$

$$S_d = [(S_{di}^2 - (S_{di})^2/n)/(n-1)]^{1/2}$$

$$T_m = t_m + 460^\circ$$

$$T_s = t_s + 460^\circ$$

$$V_{m(std)} = V_m Y_m (T_{std} / T_m) (P_m / P_{std})$$

$$V_{w(std)} = (V_{lc} \rho_w R T_{std}) / (M_w P_{std})$$

$$v_s = K_p C_p [\Delta p]^{1/2} [T_s / (P_s M_s)]^{1/2}$$

NOMENCLATURE

Symbol	Units	Description
Abs(x)	dimensionless	Absolute value of parameter x
A_n	ft ²	Area of the nozzle
A_s	ft ²	Area of the stack
B_{ws}	dimensionless	Volume proportion of water in the stack gas stream
C_p	dimensionless	Type S pitot tube coefficient
C_{anlyte}	mg/dscm	Concentration of analyte in dry stack gas, standardized
'C_{anlyte}	gr./dscf	Concentration of analyte in dry stack gas, standardized
'C_{anlyte}	ppm	Concentration of analyte in dry stack gas, standardized
CC	dimensionless	One-tailed 2.5% error confidence coefficient
d	ppm	Arithmetic mean of differences
d_i	ppm	Difference between individual CEM and reference method concentration value
D_n	inches	Internal diameter of the nozzle at the entrance orifice
D_s	inches	Internal diameter of the stack at sampling location
DE	percent	Destruction efficiency
DH	inches H ₂ O	Average pressure differential across the meter orifice
DH_@	inches H ₂ O	Orifice pressure differential that corresponds to 0.75 cfm of air at 68 °F and 29.92 inches of Hg
D_p	inches H ₂ O	Velocity head of stack gas
E_{anlyte}	lb./hour	Emission rate of analyte, time basis
I	percent	Isokinetic sampling ratio expressed as percentage
K_I	dimensionless	K-factor, ratio of DH to DP, ideal
K_p	ft[(lb/lb-mol)(in. Hg)] ^{1/2}	Type S pitot tube constant,
	s[(°R)(in. H ₂ O)] ^{1/2}	= 85.49
L_p	cfm	Measured post-test leakage rate of the sampling train
M_d	lb./lb.-mole	Molecular weight of gas at the DGM
M_s	lb./lb.-mole	Molecular weight of gas at the stack

NOMENCLATURE

Symbol	Units	Description
M_w	lb./lb.-mole	Molecular weight of water,
		= 18.0
m_{analyte}	mg	Mass of analyte in the sample
n	dimensionless	Number of data points
P	MMBtu	Fuel firing rate
P_{bar}	inches Hg	Barometric pressure at measurement site
P_{input}	tons/hour	Process dry mass input rate
p_g	inches H ₂ O	Gauge (static) pressure of stack gas
P_m	inches Hg	Absolute pressure of meter gases
P_s	inches Hg	Absolute pressure of stack gases
P_{std}	inches Hg	Standard absolute pressure
		= 29.92
Q_a	cfm	Volumetric flow rate of actual stack gas
Q_{sd}	dscfm	Volumetric flow rate of dry stack gas, standardized
R	(in. Hg)(ft ³)	Ideal gas constant,
	(lb-mole)(°R)	= 21.85
RA	percent	Relative accuracy
RE	percent	Removal efficiency
RM	ppm	Average reference method concentration
r_w	lb/mL	Density of water,
		= 0.002201
r_a	g/mL	Density of acetone,
		= 0.7899
S_d	dimensionless	Standard deviation
T_m	°R	Absolute temperature of dry gas meter
T_s	°R	Absolute temperature of stack gas
T_{std}	°R	Standard absolute temperature,
		= 528
$t_{0.975}$	dimensionless	2.5 percent error t-value
t_m	°F	Temperature of DGM
t_s	°F	Temperature of stack gas
q	minutes	Total sampling time

NOMENCLATURE

Symbol	Units	Description
V_{lc}	mL	Total volume of liquid collected
V_m	dcf	Volume of gas sample as measured by the DGM
$V_{m(std)}$	dscf	Volume of gas sample as measured by the DGM, standardized
$V_{w(std)}$	scf	Volume of water vapor in the gas sample, standardized
v_s	ft./sec	Velocity of stack gas
Y_m	dimensionless	DGM calibration coefficient
Y_c	dimensionless	DGM calibration check value
Y_w	dimensionless	Reference (wet) gas meter calibration coefficient
% CO ₂	percent	Percent CO ₂ by volume, dry basis
% O ₂	percent	Percent O ₂ by volume, dry basis
% N ₂	percent	Percent N ₂ by volume, dry basis

APPENDIX C

FIELD DATA

Advanced Industrial Resources, Inc.

Field Data

Chemance Medical
Alpharetta, Georgia

EOA			EOE		
Time	ft/min	T _s C°	Time	ft/min	T _s F°
10:33	1660	40.6	10:22	3250	109
10:38	1671	40.6	10:32	2270	
10:43	1665	40.7	10:42	1220	
10:48	1671	40.8	10:53	2100	
10:53	1660	40.8	11:03	2246	
10:58	1665	40.8	11:13	1526	
11:03	1655	40.7	11:22	2647	
11:08	1665	40.8	11:32	1200	
11:13	1665	40.7	11:42	2691	
11:18	1671	40.8	11:53	2660	
11:23	1670	40.8	12:04	2240	
11:28	1655	40.8	12:13	1240	
11:33	1660	40.8	12:22	2280	
Average:	1664.1	40.7	12:32	2150	
			12:42	2330	
			12:52	1040	
			13:02	2300	
			13:12	1155	
			13:22	1145	
			13:32	2220	
			13:42	1029	
			13:52	2365	
			14:02	1125	
			14:12	2315	
			14:22	1128	
			Average:	1914.9	109

<u>Input values:</u>		Run Number	
		EOA	EOE
T _{db}	F	103.8	106.1
T _{wb}	F	73.7	74.8
P _g	in H ₂ O	0.00	0.00
P _{bar}	in Hg	29.12	29.12
O ₂	%	21.0	21.0
CO ₂	%	0.0	0.0

<u>Calculated values:</u>			
P	in Hg	29.12	29.12
MW _{air}	lb/mol	28.84	28.84
P _{sat}	in Hg	0.84	0.87
p	in Hg	0.51	0.53
H	lb H ₂ O/lb air	0.0111	0.0115
B_{ws}		0.0175	0.0181

Note: % O₂ and % CO₂ are not important variables. Use 21% and 0% if values have not been measured.

Con ID

1 hour - 6L1822 - EOA

$P_{\text{bar}} = 29.12$

1 hour - 6L2491 - EOA

4 hours - 6L0682 - EOE

4 hours - 6L1011 - EOE

EOE

start - 1022

stop - 1422

EOA

start - 1032

stop - 1132

EOE

- 6L0682 - Initial Vacuum - 29.0" Hg

Final Vacuum - 9.5" Hg

- 6L1011 - Initial Vacuum - 29.0" Hg

Final Vacuum - 10" Hg

EOA

- 6L2491 - Initial Vacuum - ^{28 in}~~29.8~~" Hg

Final Vacuum - 7.5" Hg

6L1822 - Initial Vacuum - ^{28 in}~~29.4~~" Hg

Final Vacuum - 9.0" Hg

EOE $D_s = 1 \frac{5}{8}"$

EOA $D_s = 3"$

EDE

Time	Flow (ft./min.)	Vacuum
7:00	3250 ^{DM}	29" Hg
10:22	325 125	
10:32	2270	
10:42	1220	
10:53	34.3 2100	
11:03	25.5 2246	Jan
11:13	15.6 1526	
11:22	2647	
11:32	1200	23" Hg
11:42	2691	
11:53	2660	
12:04	2240	
12:13	1240	
12:22	2280	
12:32	2150	
12:42	2330	
12:52	1040	
13:02	2300	
13:12	1155	
13:22	1145	
13:32	2220	
13:42	1029	
13:52	2365	
14:02	1125	
14:12	2315	
14:22	1128	

$$P_{bar} = 29.12$$

$$W_b = 74.8^\circ F$$

$$D_b = 106.7^\circ F$$

106.1 same

Fan turns
on and off

103.2

$$A_s = 109.$$

$$D_s = 1.625$$

			$P_{bar} = 29.12$	
<u>Time</u>	<u>EOA</u> <u>Flow (ft./min.)</u>	<u>t_i (C)</u>		
1033	1660	40.6	28"Hg	40.1°C
1038	1671	40.6		$\omega_b = 73.7^\circ F$
1043	1665	40.7		$D_b = 103.8^\circ F$
1048	1671	40.8		
1053	1660	40.8		
1058	1665	40.8	19.5"Hg	
1103	1655	40.7		
1108	1665	40.8		
1113	1665	40.7	15 Hg	
1118	1671	40.8		
1123	1670	40.8	12.5	
1128	1655	40.8		
1133	1660	40.8		

$D_s \quad 3.0''$

APPENDIX D

LABORATORY REPORT

10/29/2020

Mr. Derek Stephens
Advanced Industrial Resources
3407 Novis Pointe

Acworth GA 30101

Project Name:
Project #: KR10590
Workorder #: 2010412

Dear Mr. Derek Stephens

The following report includes the data for the above referenced project for sample(s) received on 10/16/2020 at Eurofins Air Toxics LLC.

The data and associated QC analyzed by Modified TO-15 SIM are compliant with the project requirements or laboratory criteria with the exception of the deviations noted in the attached case narrative.

Thank you for choosing Eurofins Air Toxics LLC. for your air analysis needs. Eurofins Air Toxics Inc. is committed to providing accurate data of the highest quality. Please feel free to contact the Project Manager: Brian Whittaker at 916-985-1000 if you have any questions regarding the data in this report.

Regards,



Brian Whittaker
Project Manager

WORK ORDER #: 2010412

Work Order Summary

CLIENT:	Mr. Derek Stephens Advanced Industrial Resources 3407 Novis Pointe Acworth, GA 30101	BILL TO:	Mr. Derek Stephens Advanced Industrial Resources 3407 Novis Pointe Acworth, GA 30101
PHONE:	404-843-2100	P.O. #	
FAX:	404-845-0020	PROJECT #	KR10590
DATE RECEIVED:	10/16/2020	CONTACT:	Brian Whittaker
DATE COMPLETED:	10/29/2020		

<u>FRACTION #</u>	<u>NAME</u>	<u>TEST</u>	<u>RECEIPT VAC./PRES.</u>	<u>FINAL PRESSURE</u>
01A	KR10590-1 (EOE)	Modified TO-15 SIM	10.5 "Hg	5 psi
02A	KR10590-2 (EOE) DUP	Modified TO-15 SIM	9.5 "Hg	5 psi
03A	KR10590-3 (EOA)	Modified TO-15 SIM	8.5 "Hg	5 psi
04A	KR10590-4 (EOA) DUP	Modified TO-15 SIM	9.0 "Hg	5 psi
05A	Lab Blank	Modified TO-15 SIM	NA	NA
06A	CCV	Modified TO-15 SIM	NA	NA
07A	LCS	Modified TO-15 SIM	NA	NA
07AA	LCSD	Modified TO-15 SIM	NA	NA

CERTIFIED BY:



Technical Director

DATE: 10/29/20

Certification numbers: AZ Licensure AZ0775, FL NELAP – E87680, LA NELAP – 02089, NH NELAP - 209220, NJ NELAP - CA016, NY NELAP - 11291, TX NELAP - T104704434-20-16, UT NELAP – CA009332020-12, VA NELAP - 10615, WA NELAP - C935

Name of Accreditation Body: NELAP/ORELAP (Oregon Environmental Laboratory Accreditation Program)

Accreditation number: CA300005-014, Effective date: 10/18/2020, Expiration date: 10/17/2021.

Eurofins Air Toxics, LLC certifies that the test results contained in this report meet all requirements of the NELAC standards

This report shall not be reproduced, except in full, without the written approval of Eurofins Air Toxics, LLC.

180 BLUE RAVINE ROAD, SUITE B FOLSOM, CA - 95630

(916) 985-1000 . (800) 985-5955 . FAX (916) 351-8279

LABORATORY NARRATIVE
EPA TO-15 Ethylene oxide (SIM)
Advanced Industrial Resources
Workorder# 2010412

Four 6 Liter Summa Canister (100% Certified) samples were received on October 16, 2020. The laboratory performed analysis via EPA Method TO-15 using GC/MS in the SIM acquisition mode for the measurement of Ethylene oxide in ambient air.

Receiving Notes

The Chain of Custody (COC) information for samples KR10590-1 (EOE), KR10590-2 (EOE) DUP, KR10590-3 (EOA) and KR10590-4 (EOA) DUP did not match the entries on the sample tags with regard to sample identification. Therefore the information on the COC was used to process and report the samples.

Analytical Notes

Ethylene Oxide is not included on the laboratory's NELAP scope of accreditation for TO-15 SIM. However, TO-15 method and NELAP quality requirements were met.

Dilution was performed on samples KR10590-1 (EOE), KR10590-2 (EOE) DUP, KR10590-3 (EOA) and KR10590-4 (EOA) DUP due to the presence of high level target species.

Samples KR10590-1 (EOE), KR10590-2 (EOE) DUP, KR10590-3 (EOA) and KR10590-4 (EOA) DUP were analyzed 15 days after sample collection which was past the laboratory's SOP specified 14-day hold time, but within the method recommended 30-day hold time.

Definition of Data Qualifying Flags

Nine qualifiers may have been used on the data analysis sheets and indicates as follows:

B - Compound present in laboratory blank greater than reporting limit (background subtraction not performed).

J - Estimated value.

S - Saturated peak.

Q - Exceeds quality control limits.

U - Compound analyzed for but not detected above the reporting limit, LOD, or MDL value. See data page for project specific U-flag definition.

UJ- Non-detected compound associated with low bias in the CCV

N - The identification is based on presumptive evidence.

CN - See Case Narrative

File extensions may have been used on the data analysis sheets and indicates as follows:

a-File was requantified

b-File was quantified by a second column and detector

r1-File was requantified for the purpose of reissue

Summary of Detected Compounds MODIFIED EPA METHOD TO-15 GC/MS SIM

Client Sample ID: KR10590-1 (EOE)

Lab ID#: 2010412-01A

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Ethylene Oxide	1600	360000	3000	660000

Client Sample ID: KR10590-2 (EOE) DUP

Lab ID#: 2010412-02A

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Ethylene Oxide	1600	450000	2800	810000

Client Sample ID: KR10590-3 (EOA)

Lab ID#: 2010412-03A

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Ethylene Oxide	1500	58000	2700	100000

Client Sample ID: KR10590-4 (EOA) DUP

Lab ID#: 2010412-04A

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Ethylene Oxide	1500	46000	2800	83000



Air Toxics

Client Sample ID: KR10590-1 (EOE)

Lab ID#: 2010412-01A

MODIFIED EPA METHOD TO-15 GC/MS SIM

File Name:	30102822sim	Date of Collection:	10/14/20 2:22:00 PM
Dil. Factor:	33000	Date of Analysis:	10/28/20 11:51 PM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Ethylene Oxide	1600	360000	3000	660000

Container Type: 6 Liter Summa Canister (100% Certified)



Air Toxics

Client Sample ID: KR10590-2 (EOE) DUP

Lab ID#: 2010412-02A

MODIFIED EPA METHOD TO-15 GC/MS SIM

File Name:	30102823sim	Date of Collection:	10/14/20 2:22:00 PM
Dil. Factor:	31400	Date of Analysis:	10/29/20 12:22 AM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Ethylene Oxide	1600	450000	2800	810000

Container Type: 6 Liter Summa Canister (100% Certified)



Air Toxics

Client Sample ID: KR10590-3 (EOA)

Lab ID#: 2010412-03A

MODIFIED EPA METHOD TO-15 GC/MS SIM

File Name:	30102824sim	Date of Collection:	10/14/20 11:33:00 A
Dil. Factor:	29900	Date of Analysis:	10/29/20 12:52 AM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Ethylene Oxide	1500	58000	2700	100000

Container Type: 6 Liter Summa Canister (100% Certified)



Air Toxics

Client Sample ID: KR10590-4 (EOA) DUP

Lab ID#: 2010412-04A

MODIFIED EPA METHOD TO-15 GC/MS SIM

File Name:	30102825sim	Date of Collection:	10/14/20 11:33:00 A
Dil. Factor:	30600	Date of Analysis:	10/29/20 01:24 AM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Ethylene Oxide	1500	46000	2800	83000

Container Type: 6 Liter Summa Canister (100% Certified)



Air Toxics

Client Sample ID: Lab Blank

Lab ID#: 2010412-05A

MODIFIED EPA METHOD TO-15 GC/MS SIM

File Name: 30102805sim

Date of Collection: NA

Dil. Factor: 1.00

Date of Analysis: 10/28/20 11:47 AM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (ug/m3)	Amount (ug/m3)
Ethylene Oxide	0.050	Not Detected	0.090	Not Detected

Container Type: NA - Not Applicable



Air Toxics

Client Sample ID: CCV

Lab ID#: 2010412-06A

MODIFIED EPA METHOD TO-15 GC/MS SIM

File Name: 30102802sim

Date of Collection: NA

Dil. Factor: 1.00

Date of Analysis: 10/28/20 10:08 AM

Compound	%Recovery
----------	-----------

Ethylene Oxide	118
----------------	-----

Container Type: NA - Not Applicable



Air Toxics

Client Sample ID: LCS

Lab ID#: 2010412-07A

MODIFIED EPA METHOD TO-15 GC/MS SIM

File Name: 30102803sim

Date of Collection: NA

Dil. Factor: 1.00

Date of Analysis: 10/28/20 10:40 AM

Compound	%Recovery	Method Limits
Ethylene Oxide	111	70-130

Container Type: NA - Not Applicable



Air Toxics

Client Sample ID: LCSD

Lab ID#: 2010412-07AA

MODIFIED EPA METHOD TO-15 GC/MS SIM

File Name: 30102804sim

Date of Collection: NA

Dil. Factor: 1.00

Date of Analysis: 10/28/20 11:12 AM

Compound	%Recovery	Method Limits
Ethylene Oxide	105	70-130

Container Type: NA - Not Applicable

APPENDIX E

CALIBRATION DATA

**Accurate Solutions
Atlanta**

ISO/IEC 17025:2005 Accredited

Calibration Certificate**Issue Date: 4/27/2018**
Certificate No: 344684**Calibration Performed By:**Accurate Solutions - Atlanta, Inc.
825 Chance Rd, Suite 10
Marietta, GA 30066Phone: (770) 428-9400
Performed By: GALEN EVANS**For:** CHEMENCE MEDICAL
200 TECHNOLOGY DRIVE
ALPHARETTA, GA 30005Contact: HETAL PATEL
770-664-6732**Equipment Information:****Description: HOT WIRE ANEMOMETER**I.D.: 15100075
Manufacturer: AMPROBE
Model Number: TMA-20HW
Cal Date: 4/27/2018
Cal. Due Date: 4/30/2019Serial Number: 15100075
Temp./RH: 20°C / 49%RH
Cal. Procedure: COMPARATIVE
As Found: IN TOLERANCE
Calibration Result: PASS**Calibration Notes**

Seq.	Description	Standard	Tol. -	Tol. +	As Found	As Left	Unit
1	Applied Standard	70.0	52.0	88.0	68.0	68.0	FPM
2		150.0	132.0	168.0	151.2	151.2	FPM
3		806.0	788.0	824.0	799.8	799.8	FPM

Standards Used To Calibrate Equipment:

Company	I.D.	Description	Last Cal.	Cal. Due Date
ACCURATE SOLUTIONS	40-06-03532	AIR FLOW STANDARD	4/10/2018	4/30/2023

This Certificate may not be copied, except in full, without written consent of Accurate Solutions - Atlanta, Inc. Results listed relate only to the items that were tested and/or calibrated. "Due Date" is provided as a result of customer instruction. Any number of factors may cause the calibration item to drift out of calibration before the due date.

Accurate Solutions - Atlanta, Inc. maintains a quality system that is accredited by Perry Johnson Laboratory Accreditation, Inc. to meet or exceed the requirements set forth in ISO/IEC 17025:2005. The calibration results published in this certificate were obtained using equipment capable of producing results that are traceable through NIST to the International System of Units (SI). Calibrations meet or exceed a 4:1 TUR ratio unless otherwise noted. The uncertainty of measurement associated with the measurement result reported in this certificate is available upon request and was accounted for in making the decision of compliance or noncompliance with the relevant specification above. This certificate generated under digital signature by the below indicated technician certified competent in the company quality documents for the test(s) performed. The work was performed at the Accurate Solutions - Atlanta facility above.

Cert. #: 344684

Authorized Signature/Technician: GALEN EVANS

Date: 4/27/2018

**Accurate Solutions
Atlanta**

ISO/IEC 17025:2005 Accredited

Calibration Certificate**Issue Date: 11/10/2020**
Certificate No: 368950**Calibration Performed By:**Accurate Solutions - Atlanta, Inc.
825 Chance Rd, Suite 10
Marietta, GA 30066Phone: (770) 428-9400
Performed By: GALEN EVANS**For:** CHEMENCE MEDICAL
200 TECHNOLOGY DRIVE
ALPHARETTA, GA 30005Contact: RICHARD NARDINI
470-359-2896 X208**Equipment Information:****Description: HOT WIRE ANEMOMETER**I.D.: 15100075
Manufacturer: AMPROBE
Model Number: TMA-20HW
Cal Date: 11/10/2020
Cal. Due Date: 11/30/2021Serial Number: 15100075
Temp./RH: 23°C / 52%RH
Cal. Procedure: COMPARATIVE
As Found: IN TOLERANCE
Calibration Result: PASS**Calibration Notes** Tolerance is ± 18 up to 600 FPM, then 3% of reading.

Seq.	Description	Standard	Tol. -	Tol. +	As Found	As Left	Unit
1	Applied Standard	120.0	102.0	138.0	114.0	114.0	FPM
2		497.0	479.0	515.0	494.0	494.0	FPM
3		1996.0	1936.0	1956.0	1941.0	1941.0	FPM

Standards Used To Calibrate Equipment:

Company	I.D.	Description	Last Cal.	Cal. Due Date
ACCURATE SOLUTIONS	40-06-03532	AIR FLOW STANDARD	11/5/2020	11/30/2022

This Certificate may not be copied, except in full, without written consent of Accurate Solutions - Atlanta, Inc. Results listed relate only to the items that were tested and/or calibrated and the observations made during the test. "Due Date" is provided as a result of customer instruction. Any number of factors may cause the calibration item to drift out of calibration before the due date.

Accurate Solutions - Atlanta, Inc. maintains a quality system that is accredited by Perry Johnson Laboratory Accreditation, Inc. to meet or exceed the requirements set forth in ISO/IEC 17025:2017. The calibration results published in this certificate were obtained using equipment capable of producing results that are traceable through NIST to the International System of Units (SI). Calibrations meet or exceed a 4:1 TUR ratio unless otherwise noted or uncertainties are provided; Any uncertainty of measurement or TUR reported in this certificate is expressed as expanded uncertainty value at approximately the 95% confidence level using a coverage factor of $k=2$; Uncertainty computation does not include instrument under test resolution or repeatability; Any statement of conformity was made using a simple limit criteria not accounting for any uncertainties; or as determined by the contract review on file. This certificate generated under digital signature by the below indicated technician certified competent in the company quality documents for the test(s) performed. The work was performed at the Accurate Solutions - Atlanta facility indicated above.

Cert. #: 368950

Authorized Signature/Technician: GALEN EVANS

Date: 11/10/2020

Calibration Certificate

Issue Date: 11/10/2020
Certificate No: 368954



Calibration Performed By:

Accurate Solutions - Atlanta, Inc.
825 Chance Rd, Suite 10
Marietta, GA 30066

Phone: (770) 428-9400
Performed By: GALEN EVANS

For: CHEMENCE MEDICAL
200 TECHNOLOGY DRIVE
ALPHARETTA, GA 30005

Contact: RICHARD NARDINI
470-359-2896 X208

Equipment Information:

Description: VANE ANEMOMETER WITH IR THERMOMETER

I.D.: 160602500
Manufacturer: EXTECH
Model Number: AN200
Cal Date: 11/10/2020
Cal. Due Date: 11/30/2021

Serial Number: 160602500
Temp./RH: 23°C / 52%RH
Cal. Procedure: OEM PROC
As Found: IN TOLERANCE
Calibration Result: PASS

Calibration Notes

Seq.	Description	Standard	Tol. -	Tol. +	As Found	As Left	Unit
1	Air Flow, Observed Standard	120	76	164	118	118	FPM
2		875	809	941	882	882	FPM
3		1724	1632	1816	1663	1663	FPM
4	IR Temperature	25	23	27	25	25	°C
5		270	265	275	269	269	°C

Standards Used To Calibrate Equipment:

Company	I.D.	Description	Last Cal.	Cal. Due Date
ACCURATE SOLUTIONS	160122045	CALIBRATOR, IR	10/30/2020	10/31/2021
ACCURATE SOLUTIONS	40-06-03532	AIR FLOW STANDARD	11/5/2020	11/30/2022

This Certificate may not be copied, except in full, without written consent of Accurate Solutions - Atlanta, Inc. Results listed relate only to the items that were tested and/or calibrated and the observations made during the test. "Due Date" is provided as a result of customer instruction. Any number of factors may cause the calibration item to drift out of calibration before the due date.

Accurate Solutions - Atlanta, Inc. maintains a quality system that is accredited by Perry Johnson Laboratory Accreditation, Inc. to meet or exceed the requirements set forth in ISO/IEC 17025:2017. The calibration results published in this certificate were obtained using equipment capable of producing results that are traceable through NIST to the International System of Units (SI). Calibrations meet or exceed a 4:1 TUR ratio unless otherwise noted or uncertainties are provided; Any uncertainty of measurement or TUR reported in this certificate is expressed as expanded uncertainty value at approximately the 95% confidence level using a coverage factor of k=2; Uncertainty computation does not include instrument under test resolution or repeatability; Any statement of conformity was made using a simple limit criteria not accounting for any uncertainties; or as determined by the contract review on file. This certificate generated under digital signature by the below indicated technician certified competent in the company quality documents for the test(s) performed. The work was performed at the Accurate Solutions - Atlanta facility indicated above.

Cert. #: 368954

Authorized Signature/Technician: GALEN EVANS

Date: 11/10/2020

**Accurate Solutions
Atlanta**

ISO/IEC 17025:2005 Accredited

Calibration Certificate**Issue Date: 5/3/2018**
Certificate No: 344740**Calibration Performed By:**Accurate Solutions - Atlanta, Inc.
825 Chance Rd, Suite 10
Marietta, GA 30066Phone: (770) 428-9400
Performed By: GALEN EVANS**For:** CHEMENCE MEDICAL
200 TECHNOLOGY DRIVE
ALPHARETTA, GA 30005Contact: HETAL PATEL
770-664-6732**Equipment Information:****Description: VANE ANEMOMETER WITH IR THERMOMETER**

I.D.: 160602500

Manufacturer: EXTECH

Model Number: AN200

Cal Date: 4/27/2018

Cal. Due Date: 4/30/2019

Serial Number: 160602500

Temp./RH: 20°C / 49%RH

Cal. Procedure: OEM PROC

As Found: IN TOLERANCE

Calibration Result: PASS

Calibration Notes

Seq.	Description	Standard	Tol. -	Tol. +	As Found	As Left	Unit
1	Air Flow, Observed Standard	75	32	118	83	83	FPM
2		806	751	861	813	813	FPM
3		2008	1933	2083	2051	2051	FPM
4	IR Temperature	25	23	27	25	25	°C
5		250	245	255	247	247	°C

Standards Used To Calibrate Equipment:

Company	I.D.	Description	Last Cal.	Cal. Due Date
ACCURATE SOLUTIONS	160122045	CALIBRATOR, IR	8/17/2017	8/31/2018
ACCURATE SOLUTIONS	40-06-03532	AIR FLOW STANDARD	4/10/2018	4/30/2023

This Certificate may not be copied, except in full, without written consent of Accurate Solutions - Atlanta, Inc. Results listed relate only to the items that were tested and/or calibrated. "Due Date" is provided as a result of customer instruction. Any number of factors may cause the calibration item to drift out of calibration before the due date.

Accurate Solutions - Atlanta, Inc. maintains a quality system that is accredited by Perry Johnson Laboratory Accreditation, Inc. to meet or exceed the requirements set forth in ISO/IEC 17025:2005. The calibration results published in this certificate were obtained using equipment capable of producing results that are traceable through NIST to the International System of Units (SI). Calibrations meet or exceed a 4:1 TUR ratio unless otherwise noted. The uncertainty of measurement associated with the measurement result reported in this certificate is available upon request and was accounted for in making the decision of compliance or noncompliance with the relevant specification above. This certificate generated under digital signature by the below indicated technician certified competent in the company quality documents for the test(s) performed. The work was performed at the Accurate Solutions - Atlanta facility above.

Cert. #: 344740

Authorized Signature/Technician: GALEN EVANS

Date: 4/27/2018

APPENDIX F

PROCESS OPERATION DATA



Glue

LOT: 040482-R5

RUN # 100A94

Operation	Description	Date Completed	
4a	Operation 1 – Staging and Preconditioning *(Minimum 12 hours)	AP 19 Oct 2020	
	Operation 2 – EO Gas Consumable Preparation	AP 19 Oct 2020	
	Operation 3 – Tote Preparation and Bag Sealing	AP 19 Oct 2020	
	Operation 4 – Transfer and Load Time	AP 19 Oct 2020	
4b	Operation 5 – Sterilization/Aeration	AP 19 Oct 2020	
	Operation 6- Dosimeter Check	AP 19 Oct 2020	
4c			
PRODUCTION REVIEW COMPLETED BY/DATE: <i>David M. Davis 19 Oct 2020</i>			
Acceptance Criteria & Final Release			
The sterilization load/bag/tote contains appropriate QTY of devices.	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	The environment within the EO sterilization bag is ≥30% (absolute minimum at the start of the cycle during the warm-up phase is 24%RH	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
The average preconditioning environment is ≥12 hours at ≥68°F and ≥30%RH.	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	Devices are aerated within the sterilizer set to 50°C for ≥1 hour.	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
The average sterilizer cabinet temperature is 50°C± 3°.	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	All Dosimeters (for all bags) pass the calibration triangle.	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
The net EO delivered to each sterilization bag is 9.98g-12.00g	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	The Bis are negative (remain orange).	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
Two (2) Humidichips were placed into every sterilization bag.	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	The positive control turns positive (turbid and yellow).	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
The transfer time from 'white room' to final cartridge activation is ≤17 minutes.	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	The incubation time/temperature for the Bis is 7 days at 35-39°C.	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
The average environment within the EO sterilization bag is 50°C± 3°	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No		
Comments:	<input checked="" type="checkbox"/> N/A		
QA Release:	<i>0060 of 2020</i>		



Operation 1 - Staging for EO Sterilization

Minimum Quantity and Maximum Quantity Are Both 600 Pouches/Blisters per Tote
 Exofin XL min/ max is 99 pouches in each tote
 Exofin® Fusion min/max is 18 pouches in each tote and/or 45 pouches in each liner bag
 Ampoule boxes (Mastic/Glugone) min/max is 15 boxes per tote

Lot#:	040482-R5	Tote#:	1-10	Qty:	6,000	Dunnage Qty/ Tote#:	N/A <input checked="" type="checkbox"/>
Lot#:	N/A <input checked="" type="checkbox"/>	Tote#:	N/A <input checked="" type="checkbox"/>	Qty:	N/A <input checked="" type="checkbox"/>	Dunnage Qty/ Tote#:	N/A <input checked="" type="checkbox"/>
Lot#:	N/A <input checked="" type="checkbox"/>	Tote#:	N/A <input checked="" type="checkbox"/>	Qty:	N/A <input checked="" type="checkbox"/>	Dunnage Qty/ Tote#:	N/A <input checked="" type="checkbox"/>

Operation 1 - Preconditioning *(Minimum 12 hours)

Data Logger ID:	R17771	Calibration Due Date:	Dec 2020
-----------------	--------	-----------------------	----------

Data Logger Prepared as outlined in WI-MED-0018

Start Time:	2:55 am	Start Date:	13 Oct 2020	End Time:	2:55 pm 13 Oct 2020	End Date:	13 Oct 2020
Avg. Temperature (≥68°F):	76.26	9F	Avg. Relative Humidity (≥30%):	57.16	%		

Operator Signature/Date:	E. Lighter, R. Keller 13 Oct 2020	Checked By/Date:	M. B. B. 13 Oct 2020
QA Signature/Date:	J. H. H. 13 Oct 2020		


Operation 2 - Data Logger Setup (see WI-MED-0018)

Data Logger ID:	P70074	Calibration Due Date:	Apr 2021	Programmed Start Date/Time (include Quickstart):	13 Oct 2020 1:05 pm
Data Logger ID:	Q37941	Calibration Due Date:	Jul 2021	Programmed Start Date/Time (include Quickstart):	13 Oct 2020 1:05 pm

**Verify that totes are arranged as outlined in WI-MED-0018

☒ Pass ☐ Fail*

R17771 MultiChannel Preconditioning



Device Name: RHTemp1000IS
Device Description: Intrinsically Safe Temperature and Humidity Data Logger
Serial Number: R17771
Device ID: R17771 MultiChannel
Preconditioning
Lot Numbers:
 040482-R5 .7ml
 042482-R1 .7ml
 039482-R3 .36ml
 0085-F2 60 CM REG
Run Numbers:
 100A94
 100B94
 100C92

Date	Time	Channel 1	Channel 2
		Temperature (°F)	Humidity (% RH)
10/13/2020	2:55:49 AM	75.186	56.2
10/13/2020	3:00:49 AM	75.186	56.1
10/13/2020	3:05:49 AM	75.186	56.2
10/13/2020	3:10:49 AM	75.186	56.2
10/13/2020	3:15:49 AM	75.186	56.2
10/13/2020	3:20:49 AM	75.204	56.3
10/13/2020	3:25:49 AM	75.204	56.2
10/13/2020	3:30:49 AM	75.186	56.1
10/13/2020	3:35:49 AM	75.186	56.1
10/13/2020	3:40:49 AM	75.186	56
10/13/2020	3:45:49 AM	75.186	56
10/13/2020	3:50:49 AM	75.186	56
10/13/2020	3:55:49 AM	75.168	56.1
10/13/2020	4:00:49 AM	75.186	56.1
10/13/2020	4:05:49 AM	75.186	56.1
10/13/2020	4:10:49 AM	75.186	56
10/13/2020	4:15:49 AM	75.168	56
10/13/2020	4:20:49 AM	75.168	55.9
10/13/2020	4:25:49 AM	75.168	55.9
10/13/2020	4:30:49 AM	75.168	55.9
10/13/2020	4:35:49 AM	75.168	56
10/13/2020	4:40:49 AM	75.168	56
10/13/2020	4:45:49 AM	75.168	56
10/13/2020	4:50:49 AM	75.168	56.1
10/13/2020	4:55:49 AM	75.168	56
10/13/2020	5:00:49 AM	75.168	55.9
10/13/2020	5:05:49 AM	75.15	55.9
10/13/2020	5:10:49 AM	75.15	55.9
10/13/2020	5:15:49 AM	75.168	55.9
10/13/2020	5:20:49 AM	75.15	55.9
10/13/2020	5:25:49 AM	75.15	55.9
10/13/2020	5:30:49 AM	75.168	56
10/13/2020	5:35:49 AM	75.168	56
10/13/2020	5:40:49 AM	75.15	55.9
10/13/2020	5:45:49 AM	75.15	55.8
10/13/2020	5:50:49 AM	75.15	55.7
10/13/2020	5:55:49 AM	75.132	55.7
10/13/2020	6:00:49 AM	75.15	55.8
10/13/2020	6:05:49 AM	75.168	55.7
10/13/2020	6:10:49 AM	75.222	55.9
10/13/2020	6:15:49 AM	75.276	56.1
10/13/2020	6:20:49 AM	75.312	56.1
10/13/2020	6:25:49 AM	75.348	56.1
10/13/2020	6:30:49 AM	75.384	55.9
10/13/2020	6:35:49 AM	75.456	56.1
10/13/2020	6:40:49 AM	75.51	56.1
10/13/2020	6:45:49 AM	75.564	56.3
10/13/2020	6:50:49 AM	75.618	56.3
10/13/2020	6:55:49 AM	75.672	56.3
10/13/2020	7:00:49 AM	75.708	56.3
10/13/2020	7:05:49 AM	75.78	56.7
10/13/2020	7:10:49 AM	75.834	56.7
10/13/2020	7:15:49 AM	75.87	56.8
10/13/2020	7:20:49 AM	75.906	56.9

10/13/2020	7:25:49 AM	75.96	56.9
10/13/2020	7:30:49 AM	75.996	57
10/13/2020	7:35:49 AM	76.032	57.1
10/13/2020	7:40:49 AM	76.05	57
10/13/2020	7:45:49 AM	76.086	56.9
10/13/2020	7:50:49 AM	76.104	57
10/13/2020	7:55:49 AM	76.122	56.9
10/13/2020	8:00:49 AM	76.122	57.2
10/13/2020	8:05:49 AM	76.14	56.9
10/13/2020	8:10:49 AM	76.158	57
10/13/2020	8:15:49 AM	76.212	57.1
10/13/2020	8:20:49 AM	76.23	57.4
10/13/2020	8:25:49 AM	76.284	57.8
10/13/2020	8:30:49 AM	76.302	58
10/13/2020	8:35:49 AM	76.302	57.9
10/13/2020	8:40:49 AM	76.302	57.7
10/13/2020	8:45:49 AM	76.302	57.6
10/13/2020	8:50:49 AM	76.284	57.6
10/13/2020	8:55:49 AM	76.284	57.6
10/13/2020	9:00:49 AM	76.284	57.6
10/13/2020	9:05:49 AM	76.284	57.6
10/13/2020	9:10:49 AM	76.266	57.8
10/13/2020	9:15:49 AM	76.266	57.9
10/13/2020	9:20:49 AM	76.248	57.9
10/13/2020	9:25:49 AM	76.266	57.9
10/13/2020	9:30:49 AM	76.266	57.8
10/13/2020	9:35:49 AM	76.284	58
10/13/2020	9:40:49 AM	76.302	58.1
10/13/2020	9:45:49 AM	76.32	58
10/13/2020	9:50:49 AM	76.338	58.1
10/13/2020	9:55:49 AM	76.338	58.1
10/13/2020	10:00:49 AM	76.338	58
10/13/2020	10:05:49 AM	76.338	58
10/13/2020	10:10:49 AM	76.338	58.1
10/13/2020	10:15:49 AM	76.374	58.2
10/13/2020	10:20:49 AM	76.356	58.5
10/13/2020	10:25:49 AM	76.392	58.5
10/13/2020	10:30:49 AM	76.392	58.7
10/13/2020	10:35:49 AM	76.41	58.7
10/13/2020	10:40:49 AM	76.41	58.6
10/13/2020	10:45:49 AM	76.41	58.5
10/13/2020	10:50:49 AM	76.41	58.4
10/13/2020	10:55:49 AM	76.428	58.4
10/13/2020	11:00:49 AM	76.428	58.4
10/13/2020	11:05:49 AM	76.464	58.5
10/13/2020	11:10:49 AM	76.518	58.8
10/13/2020	11:15:49 AM	76.536	58.9
10/13/2020	11:20:49 AM	76.536	58.8
10/13/2020	11:25:49 AM	76.59	58.9
10/13/2020	11:30:49 AM	76.626	58.9
10/13/2020	11:35:49 AM	76.644	58.7
10/13/2020	11:40:49 AM	76.626	58.8
10/13/2020	11:45:49 AM	76.644	58.9
10/13/2020	11:50:49 AM	76.644	58.8
10/13/2020	11:55:49 AM	76.626	58.6
10/13/2020	12:00:49 PM	76.572	58.5
10/13/2020	12:05:49 PM	76.518	58.3
10/13/2020	12:10:49 PM	76.482	58.3
10/13/2020	12:15:49 PM	76.464	58.2
10/13/2020	12:20:49 PM	76.446	58.2
10/13/2020	12:25:49 PM	76.428	58.4
10/13/2020	12:30:49 PM	76.824	59.1
10/13/2020	12:35:49 PM	79.236	54.3
10/13/2020	12:40:49 PM	80.208	52.7
10/13/2020	12:45:49 PM	80.622	51.8
10/13/2020	12:50:49 PM	80.658	51.8
10/13/2020	12:55:49 PM	80.01	52.8
10/13/2020	1:00:49 PM	79.47	54
10/13/2020	1:05:49 PM	79.02	55
10/13/2020	1:10:49 PM	78.642	55.6
10/13/2020	1:15:49 PM	78.3	55.8
10/13/2020	1:20:49 PM	78.012	56.1
10/13/2020	1:25:49 PM	77.796	56.6
10/13/2020	1:30:49 PM	77.634	57
10/13/2020	1:35:49 PM	77.472	56.6
10/13/2020	1:40:49 PM	77.328	56.1
10/13/2020	1:45:49 PM	77.238	57.6
10/13/2020	1:50:49 PM	77.184	57.6
10/13/2020	1:55:49 PM	77.148	58.4
10/13/2020	2:00:49 PM	77.094	59.1
10/13/2020	2:05:49 PM	77.04	59.6
10/13/2020	2:10:49 PM	76.968	59
10/13/2020	2:15:49 PM	76.95	59.4
10/13/2020	2:20:49 PM	76.932	60.1
10/13/2020	2:25:49 PM	76.932	61.4
10/13/2020	2:30:49 PM	76.914	60.7
10/13/2020	2:35:49 PM	76.878	60.7
10/13/2020	2:40:49 PM	76.86	60.9
10/13/2020	2:45:49 PM	76.842	61.3
10/13/2020	2:50:49 PM	76.824	60.3
10/13/2020	2:55:49 PM	76.77	59.1

76.26252414

57.15655172

Document ID: SOP-MED-0064
Document Type: Device History Record



Name:
Version

EO Sterilization Router
2.0

Operation 2 – EO Gas Consumable Preparation

EO Gas Canister Pre-Weight Information Recorded on Pg. 5	Box labels are printed and match each tote.
EO Bag Information Recorded on Pg. 4	PCDs are placed in the center of the load.
EO Gas Canister Labeled and mounted to tote (s) with product.	Positive Control same Lot number as PCDs in totes
EZ Test Bis not damaged, Labeled, inside PCD and sealed.	Dosimeter is labeled and placed in the center of the load.
Humidichips (2) placed inside Humiditube inside the tote.	Equipment Logbooks filled out.
2 Data loggers placed into the front of the tote (hottest and coolest locations):	
Control Number (location): 050A94	Serial Number: P70074
Control Number (location): 090A94	Serial Number: Q37941
Operator Signature/date: Elizabeth Royler 13Oct2020	Checked By/date: Marnie C. Bell 13Oct2020
Operation 3 - Tote Preparation and Bag Sealing	
EO Bag Sealer Asset ID: 54774	EO Bag Sealer Calibration Due Date: Feb 2021
QA Line Clearance – Proceed to Bag Sealing	
All steps in Operation 2 are complete and all consumables are loaded into each tote	
QA Signature/Date: Zahra McHenry 13Oct2020	
Operation 4 - Transfer & Load Time	
Time left "white room" area (b): 4:10pm	Sterilizer Temp (50°C ± 3°): 49.5 °C
Printer Paper/Ink Checked: <input checked="" type="checkbox"/>	
Check to ensure each tote remains vacuum sealed prior to loading into the chamber:	
Time Last Cartridge Activated (a): 4:13pm	Total Transfer Time (c) (a-b) Not to exceed 17 minutes: 3 min
Time/Date that cycle started (countdown completed/doors locked):	4:14pm 13Oct2020
Operator Signature/Date: Marnie C. Bell 13Oct2020	Checked By/Date: Elizabeth Royler 13Oct2020

Document ID: SOP-MED-0064
Document Type: Device History Record



Name:
Version

EO Sterilization Router
2.0

Operation 5 - Sterilization / Aeration

Time sterilizer "UNLOAD" pressed (Minimum 16 hours from "Cycle Start"):	ID: 25am	Sterilize end time (After countdown/doors unlocked):	ID: 30am
Aeration start time (after countdown completed/doors locked):	ID: 33am		
Aeration End Time (when countdown is completed/doors unlocked):	11:33am	Date:	14Oct+2020

Operation 5 - EO Data Logger Downloading

Data Logger Offloaded as outlined in WI 09-014 - Record Batch Number and Control Number on Printouts

Data Logger ID:	P700074	Control Number:	050A04
Start Date/Time on Datalogger Printout:	13Oct+2020 4:10pm	Sterilizer Temp (50°C±3°):	49.45 °C
End Date/Time on Datalogger Printout:	14Oct+2020 11:35am	Avg. Relative Humidity Must Be ≥ 30% (for totes with product):	53.55 %
Data Logger ID:	Q37941	Control Number:	090A94
Start Date/Time on Datalogger Printout:	13Oct+2020 4:10pm	Sterilizer Temp (50°C±3°):	49.96 °C
End Date/Time on Datalogger Printout:	14Oct+2020 11:35am	Avg. Relative Humidity Must Be ≥ 30% (for totes with product):	55.72 %

BI Incubation


Bis/Dosimeters submitted to QC (submit P.2 with TR if multiple lots:

Attach EO Chamber Printout 50°C±3°	MIN:	49.9 °C	MAX:	50.2 °C	TR #:	BI:	20-2856	Attached:	<input checked="" type="checkbox"/>
------------------------------------	------	---------	------	---------	-------	-----	---------	-----------	-------------------------------------

Sterilization Consumable Lot Information

EOGas lot number(s):	191956	Exp. Date:	09Jun2024	Dosimeter lot number(s):	204517	Exp. Date:	12Feb2023
EOGas bag lot number(s):	EOKCM031090419	Exp. Date:	N/A ER1306+2020	Humidichip lot number(s):	2041610	Exp. Date:	19Feb2023
*BI lot number(s):	G-214	Exp. Date:	13Jun2021	*If multiple lots used, 1 PC from each lot must be present			
Scale/Balance ID:	0035805103	CAL Due:	Oct+2020				

P70074 MultiChannel 050A94



Device Name:	RHTemp1000IS	RHTemp1000IS
Device Description:	Intrinsically Safe Temperature and Humidity Data Logger	Intrinsically Safe Temperature and Humidity Data Logger
Serial Number:	P70074	P70074
Device ID:	050A94 MultiChannel	MultiChannel


Date	Time	Channel 1 Temperature (°C)	Channel 2 Humidity (% RH)
10/13/2020	4:10:41 PM	24.85	56.5
10/13/2020	4:15:41 PM	24.88	56.9
10/13/2020	4:20:41 PM	25.59	57.8
10/13/2020	4:25:41 PM	26.81	57.2
10/13/2020	4:30:41 PM	28.17	55.8
10/13/2020	4:35:41 PM	29.54	54.4
10/13/2020	4:40:41 PM	30.87	53.2
10/13/2020	4:45:41 PM	32.19	51.8
10/13/2020	4:50:41 PM	33.57	50.4
10/13/2020	4:55:41 PM	34.85	49.2
10/13/2020	5:00:41 PM	36.01	48.3
10/13/2020	5:05:41 PM	37.04	47.6
10/13/2020	5:10:41 PM	37.98	47.1
10/13/2020	5:15:41 PM	38.84	46.8
10/13/2020	5:20:41 PM	39.6	46.6
10/13/2020	5:25:41 PM	40.3	46.5
10/13/2020	5:30:41 PM	40.94	46.4
10/13/2020	5:35:41 PM	41.53	46.3
10/13/2020	5:40:41 PM	42.09	46.2
10/13/2020	5:45:41 PM	42.61	46.1
10/13/2020	5:50:41 PM	43.09	46.2
10/13/2020	5:55:41 PM	43.54	46.4
10/13/2020	6:00:41 PM	43.95	46.5
10/13/2020	6:05:41 PM	44.33	46.7
10/13/2020	6:10:41 PM	44.68	46.8
10/13/2020	6:15:41 PM	45.02	46.9
10/13/2020	6:20:41 PM	45.34	47.1
10/13/2020	6:25:41 PM	45.64	47.4
10/13/2020	6:30:41 PM	45.92	47.5
10/13/2020	6:35:41 PM	46.18	47.7
10/13/2020	6:40:41 PM	46.41	47.9
10/13/2020	6:45:41 PM	46.66	48.1
10/13/2020	6:50:41 PM	46.87	48.3
10/13/2020	6:55:41 PM	47.08	48.5
10/13/2020	7:00:41 PM	47.29	48.7
10/13/2020	7:05:41 PM	47.47	48.9
10/13/2020	7:10:41 PM	47.65	49.1
10/13/2020	7:15:41 PM	47.82	49.3
10/13/2020	7:20:41 PM	47.98	49.5
10/13/2020	7:25:41 PM	48.13	49.6
10/13/2020	7:30:41 PM	48.28	49.8
10/13/2020	7:35:41 PM	48.43	49.8
10/13/2020	7:40:41 PM	48.58	50
10/13/2020	7:45:41 PM	48.73	50.4
10/13/2020	7:50:41 PM	48.86	50.9
10/13/2020	7:55:41 PM	48.99	51.2
10/13/2020	8:00:41 PM	49.13	51.4
10/13/2020	8:05:41 PM	49.24	51.6
10/13/2020	8:10:41 PM	49.37	51.7
10/13/2020	8:15:41 PM	49.49	51.8
10/13/2020	8:20:41 PM	49.59	52
10/13/2020	8:25:41 PM	49.7	52
10/13/2020	8:30:41 PM	49.8	52.1
10/13/2020	8:35:41 PM	49.89	52.3
10/13/2020	8:40:41 PM	49.98	52.4
10/13/2020	8:45:41 PM	50.07	52.5
10/13/2020	8:50:41 PM	50.15	52.6
10/13/2020	8:55:41 PM	50.23	52.7
10/13/2020	9:00:41 PM	50.31	52.8
10/13/2020	9:05:41 PM	50.35	52.9
10/13/2020	9:10:41 PM	50.42	53
10/13/2020	9:15:41 PM	50.47	53.1
10/13/2020	9:20:41 PM	50.53	53.2
10/13/2020	9:25:41 PM	50.58	53.3
10/13/2020	9:30:41 PM	50.63	53.4
10/13/2020	9:35:41 PM	50.68	53.5
10/13/2020	9:40:41 PM	50.72	53.6

10/13/2020	9:45:41 PM	50.77	53.7
10/13/2020	9:50:41 PM	50.81	53.8
10/13/2020	9:55:41 PM	50.85	53.9
10/13/2020	10:00:41 PM	50.89	54
10/13/2020	10:05:41 PM	50.92	54.1
10/13/2020	10:10:41 PM	50.95	54.2
10/13/2020	10:15:41 PM	50.99	54.2
10/13/2020	10:20:41 PM	51.02	54.3
10/13/2020	10:25:41 PM	51.05	54.4
10/13/2020	10:30:41 PM	51.07	54.5
10/13/2020	10:35:41 PM	51.1	54.6
10/13/2020	10:40:41 PM	51.13	54.7
10/13/2020	10:45:41 PM	51.15	54.7
10/13/2020	10:50:41 PM	51.16	54.8
10/13/2020	10:55:41 PM	51.19	54.9
10/13/2020	11:00:41 PM	51.21	54.9
10/13/2020	11:05:41 PM	51.23	55
10/13/2020	11:10:41 PM	51.26	55.1
10/13/2020	11:15:41 PM	51.28	55.2
10/13/2020	11:20:41 PM	51.29	55.2
10/13/2020	11:25:41 PM	51.31	55.3
10/13/2020	11:30:41 PM	51.33	55.4
10/13/2020	11:35:41 PM	51.33	55.4
10/13/2020	11:40:41 PM	51.35	55.5
10/13/2020	11:45:41 PM	51.36	55.6
10/13/2020	11:50:41 PM	51.38	55.6
10/13/2020	11:55:41 PM	51.41	55.7
10/14/2020	12:00:41 AM	51.41	55.7
10/14/2020	12:05:41 AM	51.42	55.8
10/14/2020	12:10:41 AM	51.43	55.9
10/14/2020	12:15:41 AM	51.45	55.9
10/14/2020	12:20:41 AM	51.46	56
10/14/2020	12:25:41 AM	51.47	56
10/14/2020	12:30:41 AM	51.49	56
10/14/2020	12:35:41 AM	51.5	56.1
10/14/2020	12:40:41 AM	51.53	56.1
10/14/2020	12:45:41 AM	51.53	56.2
10/14/2020	12:50:41 AM	51.56	56.2
10/14/2020	12:55:41 AM	51.55	56.3
10/14/2020	1:00:41 AM	51.57	56.3
10/14/2020	1:05:41 AM	51.57	56.4
10/14/2020	1:10:41 AM	51.57	56.4
10/14/2020	1:15:41 AM	51.59	56.4
10/14/2020	1:20:41 AM	51.59	56.5
10/14/2020	1:25:41 AM	51.61	56.5
10/14/2020	1:30:41 AM	51.62	56.6
10/14/2020	1:35:41 AM	51.64	56.6
10/14/2020	1:40:41 AM	51.64	56.6
10/14/2020	1:45:41 AM	51.65	56.7
10/14/2020	1:50:41 AM	51.66	56.7
10/14/2020	1:55:41 AM	51.66	56.7
10/14/2020	2:00:41 AM	51.66	56.8
10/14/2020	2:05:41 AM	51.66	56.8
10/14/2020	2:10:41 AM	51.66	56.9
10/14/2020	2:15:41 AM	51.66	56.9
10/14/2020	2:20:41 AM	51.67	57
10/14/2020	2:25:41 AM	51.68	57
10/14/2020	2:30:41 AM	51.69	57.1
10/14/2020	2:35:41 AM	51.71	57.1
10/14/2020	2:40:41 AM	51.72	57.1
10/14/2020	2:45:41 AM	51.71	57.1
10/14/2020	2:50:41 AM	51.72	57.2
10/14/2020	2:55:41 AM	51.72	57.2
10/14/2020	3:00:41 AM	51.72	57.2
10/14/2020	3:05:41 AM	51.73	57.2
10/14/2020	3:10:41 AM	51.72	57.3
10/14/2020	3:15:41 AM	51.72	57.3
10/14/2020	3:20:41 AM	51.73	57.4
10/14/2020	3:25:41 AM	51.73	57.4
10/14/2020	3:30:41 AM	51.74	57.4
10/14/2020	3:35:41 AM	51.75	57.5
10/14/2020	3:40:41 AM	51.76	57.5
10/14/2020	3:45:41 AM	51.77	57.5
10/14/2020	3:50:41 AM	51.78	57.5

10/14/2020	3:55:41 AM	51.77	57.6
10/14/2020	4:00:41 AM	51.77	57.6
10/14/2020	4:05:41 AM	51.77	57.6
10/14/2020	4:10:41 AM	51.77	57.6
10/14/2020	4:15:41 AM	51.77	57.7
10/14/2020	4:20:41 AM	51.76	57.7
10/14/2020	4:25:41 AM	51.76	57.8
10/14/2020	4:30:41 AM	51.76	57.8
10/14/2020	4:35:41 AM	51.77	57.8
10/14/2020	4:40:41 AM	51.78	57.8
10/14/2020	4:45:41 AM	51.8	57.9
10/14/2020	4:50:41 AM	51.8	57.9
10/14/2020	4:55:41 AM	51.8	57.9
10/14/2020	5:00:41 AM	51.81	57.9
10/14/2020	5:05:41 AM	51.8	57.9
10/14/2020	5:10:41 AM	51.8	58
10/14/2020	5:15:41 AM	51.81	58
10/14/2020	5:20:41 AM	51.8	58
10/14/2020	5:25:41 AM	51.8	58.1
10/14/2020	5:30:41 AM	51.79	58.1
10/14/2020	5:35:41 AM	51.78	58.1
10/14/2020	5:40:41 AM	51.78	58.1
10/14/2020	5:45:41 AM	51.78	58.2
10/14/2020	5:50:41 AM	51.77	58.2
10/14/2020	5:55:41 AM	51.77	58.2
10/14/2020	6:00:41 AM	51.77	58.2
10/14/2020	6:05:41 AM	51.76	58.3
10/14/2020	6:10:41 AM	51.75	58.3
10/14/2020	6:15:41 AM	51.75	58.3
10/14/2020	6:20:41 AM	51.75	58.4
10/14/2020	6:25:41 AM	51.73	58.4
10/14/2020	6:30:41 AM	51.73	58.4
10/14/2020	6:35:41 AM	51.73	58.4
10/14/2020	6:40:41 AM	51.73	58.5
10/14/2020	6:45:41 AM	51.73	58.5
10/14/2020	6:50:41 AM	51.73	58.5
10/14/2020	6:55:41 AM	51.72	58.5
10/14/2020	7:00:41 AM	51.73	58.5
10/14/2020	7:05:41 AM	51.72	58.5

10/14/2020	7:10:41 AM	51.72	58.6
10/14/2020	7:15:41 AM	51.71	58.6
10/14/2020	7:20:41 AM	51.71	58.6
10/14/2020	7:25:41 AM	51.72	58.6
10/14/2020	7:30:41 AM	51.71	58.7
10/14/2020	7:35:41 AM	51.7	58.7
10/14/2020	7:40:41 AM	51.7	58.7
10/14/2020	7:45:41 AM	51.7	58.7
10/14/2020	7:50:41 AM	51.7	58.8
10/14/2020	7:55:41 AM	51.69	58.8
10/14/2020	8:00:41 AM	51.69	58.8
10/14/2020	8:05:41 AM	51.68	58.8
10/14/2020	8:10:41 AM	51.68	58.9
10/14/2020	8:15:41 AM	51.68	58.9
10/14/2020	8:20:41 AM	51.67	58.9
10/14/2020	8:25:41 AM	51.67	58.9
10/14/2020	8:30:41 AM	51.67	59
10/14/2020	8:35:41 AM	51.67	59
10/14/2020	8:40:41 AM	51.67	59
10/14/2020	8:45:41 AM	51.67	59
10/14/2020	8:50:41 AM	51.67	59
10/14/2020	8:55:41 AM	51.67	59
10/14/2020	9:00:41 AM	51.66	59
10/14/2020	9:05:41 AM	51.66	59.1
10/14/2020	9:10:41 AM	51.66	59.1
10/14/2020	9:15:41 AM	51.66	59.1
10/14/2020	9:20:41 AM	51.66	59.1
10/14/2020	9:25:41 AM	51.67	59.1
10/14/2020	9:30:41 AM	51.67	59.1
10/14/2020	9:35:41 AM	51.66	59.1
10/14/2020	9:40:41 AM	51.67	59.1
10/14/2020	9:45:41 AM	51.67	59.2
10/14/2020	9:50:41 AM	51.68	59.2
10/14/2020	9:55:41 AM	51.69	59.2
10/14/2020	10:00:41 AM	51.68	59.2
10/14/2020	10:05:41 AM	51.69	59.2
10/14/2020	10:10:41 AM	51.7	59.2
10/14/2020	10:15:41 AM	51.71	59.2
10/14/2020	10:20:41 AM	51.72	59.2
10/14/2020	10:25:41 AM	51.72	59.2
10/14/2020	10:30:41 AM	51.73	59.3
10/14/2020	10:35:41 AM	51.62	32.8
10/14/2020	10:40:41 AM	51.49	27.5
10/14/2020	10:45:41 AM	51.46	26.4
10/14/2020	10:50:41 AM	51.46	25.8
10/14/2020	10:55:41 AM	51.5	25.2
10/14/2020	11:00:41 AM	51.58	24.9
10/14/2020	11:05:41 AM	51.67	24.4
10/14/2020	11:10:41 AM	51.75	24.1
10/14/2020	11:15:41 AM	51.85	23.9
10/14/2020	11:20:41 AM	51.94	23.6
10/14/2020	11:25:41 AM	52.03	23.4
10/14/2020	11:30:41 AM	52.12	23.1
10/14/2020	11:35:41 AM	52.2	23.2
		49.44683761	53.54682906

Q37941 MultiChannel 090A94



Device Name:	RHTemp1000IS	RHTemp1000IS
Device Description:	Intrinsically Safe Temperature and Humidity Data Logger	Intrinsically Safe Temperature and Humidity Data Logger
Serial Number:	Q37941	Q37941
Device ID:	090A94 MultiChannel	MultiChannel

Date	Time	Channel 1 Temperature (°C)	Channel 2 Humidity (% RH)
10/13/2020	4:10:40 PM	24.57	58.2
10/13/2020	4:15:40 PM	24.62	59.2
10/13/2020	4:20:40 PM	25.91	58.8
10/13/2020	4:25:40 PM	27.59	57.4
10/13/2020	4:30:40 PM	29.37	56.1
10/13/2020	4:35:40 PM	31.12	55
10/13/2020	4:40:40 PM	32.8	54.1
10/13/2020	4:45:40 PM	34.4	53.2
10/13/2020	4:50:40 PM	35.88	52.3
10/13/2020	4:55:40 PM	37.28	51.4
10/13/2020	5:00:40 PM	38.56	50.4
10/13/2020	5:05:40 PM	39.7	50
10/13/2020	5:10:40 PM	40.71	49.8
10/13/2020	5:15:40 PM	41.59	49.7
10/13/2020	5:20:40 PM	42.37	49.6
10/13/2020	5:25:40 PM	43.06	49.4
10/13/2020	5:30:40 PM	43.68	48.9
10/13/2020	5:35:40 PM	44.2	48.6
10/13/2020	5:40:40 PM	44.68	48.4
10/13/2020	5:45:40 PM	45.11	48.6
10/13/2020	5:50:40 PM	45.48	48.9
10/13/2020	5:55:40 PM	45.83	49.1
10/13/2020	6:00:40 PM	46.15	49.3
10/13/2020	6:05:40 PM	46.44	49.5
10/13/2020	6:10:40 PM	46.73	49.8
10/13/2020	6:15:40 PM	47.01	50.1
10/13/2020	6:20:40 PM	47.28	50.4
10/13/2020	6:25:40 PM	47.52	50.9
10/13/2020	6:30:40 PM	47.72	51.5
10/13/2020	6:35:40 PM	47.91	52.1
10/13/2020	6:40:40 PM	48.08	52.3
10/13/2020	6:45:40 PM	48.24	52.5
10/13/2020	6:50:40 PM	48.4	52.7
10/13/2020	6:55:40 PM	48.57	52.9
10/13/2020	7:00:40 PM	48.72	53.1
10/13/2020	7:05:40 PM	48.86	53.3
10/13/2020	7:10:40 PM	48.99	53.5
10/13/2020	7:15:40 PM	49.1	53.7
10/13/2020	7:20:40 PM	49.21	53.9
10/13/2020	7:25:40 PM	49.32	54.1
10/13/2020	7:30:40 PM	49.45	54.2
10/13/2020	7:35:40 PM	49.58	54.3
10/13/2020	7:40:40 PM	49.71	54.4
10/13/2020	7:45:40 PM	49.82	54.5
10/13/2020	7:50:40 PM	49.92	54.6
10/13/2020	7:55:40 PM	50.02	54.7
10/13/2020	8:00:40 PM	50.11	54.8
10/13/2020	8:05:40 PM	50.22	54.9
10/13/2020	8:10:40 PM	50.33	55
10/13/2020	8:15:40 PM	50.44	55
10/13/2020	8:20:40 PM	50.53	55.1
10/13/2020	8:25:40 PM	50.6	55.1
10/13/2020	8:30:40 PM	50.67	55.2
10/13/2020	8:35:40 PM	50.73	55.3
10/13/2020	8:40:40 PM	50.8	55.4
10/13/2020	8:45:40 PM	50.87	55.5
10/13/2020	8:50:40 PM	50.95	55.6
10/13/2020	8:55:40 PM	51.01	55.6
10/13/2020	9:00:40 PM	51.06	55.6
10/13/2020	9:05:40 PM	51.1	55.7
10/13/2020	9:10:40 PM	51.13	55.8
10/13/2020	9:15:40 PM	51.16	55.9
10/13/2020	9:20:40 PM	51.18	56
10/13/2020	9:25:40 PM	51.21	56.1
10/13/2020	9:30:40 PM	51.26	56.2
10/13/2020	9:35:40 PM	51.31	56.3
10/13/2020	9:40:40 PM	51.35	56.3

10/13/2020	9:45:40 PM	51.38	56.4
10/13/2020	9:50:40 PM	51.4	56.4
10/13/2020	9:55:40 PM	51.42	56.5
10/13/2020	10:00:40 PM	51.43	56.6
10/13/2020	10:05:40 PM	51.44	56.7
10/13/2020	10:10:40 PM	51.46	56.8
10/13/2020	10:15:40 PM	51.5	56.9
10/13/2020	10:20:40 PM	51.53	56.9
10/13/2020	10:25:40 PM	51.55	56.9
10/13/2020	10:30:40 PM	51.56	57
10/13/2020	10:35:40 PM	51.57	57
10/13/2020	10:40:40 PM	51.57	57.1
10/13/2020	10:45:40 PM	51.57	57.2
10/13/2020	10:50:40 PM	51.58	57.2
10/13/2020	10:55:40 PM	51.59	57.3
10/13/2020	11:00:40 PM	51.62	57.4
10/13/2020	11:05:40 PM	51.65	57.4
10/13/2020	11:10:40 PM	51.66	57.4
10/13/2020	11:15:40 PM	51.67	57.5
10/13/2020	11:20:40 PM	51.68	57.5
10/13/2020	11:25:40 PM	51.67	57.5
10/13/2020	11:30:40 PM	51.67	57.6
10/13/2020	11:35:40 PM	51.67	57.7
10/13/2020	11:40:40 PM	51.69	57.8
10/13/2020	11:45:40 PM	51.71	57.8
10/13/2020	11:50:40 PM	51.73	57.8
10/13/2020	11:55:40 PM	51.74	57.8
10/14/2020	12:00:40 AM	51.74	57.8
10/14/2020	12:05:40 AM	51.73	57.9
10/14/2020	12:10:40 AM	51.73	58
10/14/2020	12:15:40 AM	51.73	58
10/14/2020	12:20:40 AM	51.73	58.1
10/14/2020	12:25:40 AM	51.74	58.1
10/14/2020	12:30:40 AM	51.76	58.2
10/14/2020	12:35:40 AM	51.79	58.2
10/14/2020	12:40:40 AM	51.81	58.1
10/14/2020	12:45:40 AM	51.82	58.1
10/14/2020	12:50:40 AM	51.82	58.2
10/14/2020	12:55:40 AM	51.81	58.2
10/14/2020	1:00:40 AM	51.81	58.3
10/14/2020	1:05:40 AM	51.8	58.3
10/14/2020	1:10:40 AM	51.79	58.4
10/14/2020	1:15:40 AM	51.8	58.5
10/14/2020	1:20:40 AM	51.81	58.5
10/14/2020	1:25:40 AM	51.84	58.5
10/14/2020	1:30:40 AM	51.87	58.5
10/14/2020	1:35:40 AM	51.88	58.4
10/14/2020	1:40:40 AM	51.88	58.5
10/14/2020	1:45:40 AM	51.87	58.5
10/14/2020	1:50:40 AM	51.86	58.5
10/14/2020	1:55:40 AM	51.85	58.6
10/14/2020	2:00:40 AM	51.84	58.6
10/14/2020	2:05:40 AM	51.82	58.7
10/14/2020	2:10:40 AM	51.82	58.7
10/14/2020	2:15:40 AM	51.82	58.8
10/14/2020	2:20:40 AM	51.85	58.8
10/14/2020	2:25:40 AM	51.87	58.8
10/14/2020	2:30:40 AM	51.89	58.7
10/14/2020	2:35:40 AM	51.9	58.7
10/14/2020	2:40:40 AM	51.89	58.8
10/14/2020	2:45:40 AM	51.89	58.8
10/14/2020	2:50:40 AM	51.88	58.8
10/14/2020	2:55:40 AM	51.87	58.9
10/14/2020	3:00:40 AM	51.86	58.9
10/14/2020	3:05:40 AM	51.84	59
10/14/2020	3:10:40 AM	51.83	59
10/14/2020	3:15:40 AM	51.84	59.1
10/14/2020	3:20:40 AM	51.85	59.1
10/14/2020	3:25:40 AM	51.88	59.1
10/14/2020	3:30:40 AM	51.9	59.1
10/14/2020	3:35:40 AM	51.92	59
10/14/2020	3:40:40 AM	51.93	59
10/14/2020	3:45:40 AM	51.92	59.1
10/14/2020	3:50:40 AM	51.91	59.1

10/14/2020	3:55:40 AM	51.9	59.1
10/14/2020	4:00:40 AM	51.89	59.2
10/14/2020	4:05:40 AM	51.88	59.2
10/14/2020	4:10:40 AM	51.87	59.2
10/14/2020	4:15:40 AM	51.86	59.3
10/14/2020	4:20:40 AM	51.85	59.4
10/14/2020	4:25:40 AM	51.86	59.4
10/14/2020	4:30:40 AM	51.87	59.4
10/14/2020	4:35:40 AM	51.9	59.4
10/14/2020	4:40:40 AM	51.93	59.3
10/14/2020	4:45:40 AM	51.95	59.3
10/14/2020	4:50:40 AM	51.96	59.2
10/14/2020	4:55:40 AM	51.96	59.3
10/14/2020	5:00:40 AM	51.95	59.3
10/14/2020	5:05:40 AM	51.94	59.3
10/14/2020	5:10:40 AM	51.92	59.4
10/14/2020	5:15:40 AM	51.91	59.4
10/14/2020	5:20:40 AM	51.9	59.5
10/14/2020	5:25:40 AM	51.89	59.5
10/14/2020	5:30:40 AM	51.88	59.5
10/14/2020	5:35:40 AM	51.87	59.6
10/14/2020	5:40:40 AM	51.86	59.6
10/14/2020	5:45:40 AM	51.85	59.6
10/14/2020	5:50:40 AM	51.84	59.6
10/14/2020	5:55:40 AM	51.83	59.7
10/14/2020	6:00:40 AM	51.83	59.7
10/14/2020	6:05:40 AM	51.82	59.7
10/14/2020	6:10:40 AM	51.82	59.7
10/14/2020	6:15:40 AM	51.81	59.8
10/14/2020	6:20:40 AM	51.8	59.8
10/14/2020	6:25:40 AM	51.79	59.8
10/14/2020	6:30:40 AM	51.78	59.8
10/14/2020	6:35:40 AM	51.78	59.8
10/14/2020	6:40:40 AM	51.78	59.9
10/14/2020	6:45:40 AM	51.78	59.9
10/14/2020	6:50:40 AM	51.78	59.9
10/14/2020	6:55:40 AM	51.78	59.9
10/14/2020	7:00:40 AM	51.77	59.9
10/14/2020	7:05:40 AM	51.77	59.9

10/14/2020	7:10:40 AM	51.76	59.9
10/14/2020	7:15:40 AM	51.75	59.9
10/14/2020	7:20:40 AM	51.75	59.9
10/14/2020	7:25:40 AM	51.74	60
10/14/2020	7:30:40 AM	51.73	60
10/14/2020	7:35:40 AM	51.73	60
10/14/2020	7:40:40 AM	51.72	60
10/14/2020	7:45:40 AM	51.72	60
10/14/2020	7:50:40 AM	51.71	60
10/14/2020	7:55:40 AM	51.71	60
10/14/2020	8:00:40 AM	51.7	60
10/14/2020	8:05:40 AM	51.7	60.1
10/14/2020	8:10:40 AM	51.7	60.1
10/14/2020	8:15:40 AM	51.69	60.1
10/14/2020	8:20:40 AM	51.69	60.1
10/14/2020	8:25:40 AM	51.69	60.1
10/14/2020	8:30:40 AM	51.68	60.1
10/14/2020	8:35:40 AM	51.68	60.1
10/14/2020	8:40:40 AM	51.68	60.1
10/14/2020	8:45:40 AM	51.67	60.1
10/14/2020	8:50:40 AM	51.67	60.1
10/14/2020	8:55:40 AM	51.67	60.2
10/14/2020	9:00:40 AM	51.67	60.2
10/14/2020	9:05:40 AM	51.67	60.2
10/14/2020	9:10:40 AM	51.67	60.2
10/14/2020	9:15:40 AM	51.67	60.2
10/14/2020	9:20:40 AM	51.67	60.2
10/14/2020	9:25:40 AM	51.68	60.2
10/14/2020	9:30:40 AM	51.68	60.2
10/14/2020	9:35:40 AM	51.69	60.2
10/14/2020	9:40:40 AM	51.69	60.2
10/14/2020	9:45:40 AM	51.69	60.2
10/14/2020	9:50:40 AM	51.7	60.2
10/14/2020	9:55:40 AM	51.71	60.2
10/14/2020	10:00:40 AM	51.72	60.2
10/14/2020	10:05:40 AM	51.72	60.2
10/14/2020	10:10:40 AM	51.73	60.2
10/14/2020	10:15:40 AM	51.74	60.2
10/14/2020	10:20:40 AM	51.74	60.1
10/14/2020	10:25:40 AM	51.74	60.1
10/14/2020	10:30:40 AM	51.75	60.3
10/14/2020	10:35:40 AM	51.6	33.1
10/14/2020	10:40:40 AM	51.47	31.4
10/14/2020	10:45:40 AM	51.41	30.8
10/14/2020	10:50:40 AM	51.39	30.3
10/14/2020	10:55:40 AM	51.4	29.9
10/14/2020	11:00:40 AM	51.41	29.2
10/14/2020	11:05:40 AM	51.44	28.8
10/14/2020	11:10:40 AM	51.47	28.4
10/14/2020	11:15:40 AM	51.49	28
10/14/2020	11:20:40 AM	51.52	27.7
10/14/2020	11:25:40 AM	51.54	27.3
10/14/2020	11:30:40 AM	51.56	27
10/14/2020	11:35:40 AM	51.6	26.9
		49.95666667	55.72008547



	Gas Pre-Weight (g)	Gas Post-Weight (g)	Net Weight Release (g)	Attach Pre and Post Weights Below:	Dosimeter Check	
Tote #1	88.38	77.73	10.65	Pre A 001:N + 88.380 g	1	▲ P F
Tote #2	88.24	77.75	10.49	002:N + 88.246 g	2	▲ P F
Tote #3	88.41	77.94	10.47	003:N + 88.413 g	3	▲ P F
Tote #4	88.36	77.83	10.53	004:N + 88.360 g	4	▲ P F
Tote #5	88.20	77.78	10.42	005:N + 88.208 g	5	▲ P F
Tote #6	88.25	77.82	10.43	006:N + 88.250 g	6	▲ P F
Tote #7	88.15	77.67	10.48	007:N + 88.155 g	7	▲ P F
Tote #8	88.26	77.76	10.50	008:N + 88.263 g	8	▲ P F
Tote #9	88.26	77.79	10.47	009:N + 88.261 g	9	▲ P F
Tote #10	88.33	77.81	10.52	010:N + 88.338 g	10	▲ P F

All EO Gas Post Weights Pass (9.98g-12.00g)
☒ Yes ☐ No (*If "No" notify Supervisor)

Operator/Date: Marcia Beltran 14 Oct 2020

Checked By/Date: Bre Atkinson 16 Oct 2020

Document ID: SOP-MED-0098
Document Type: DHR



Name:
Version

EOE Sterilization Router
3.0

LOT: 040482-R5/039482-R3/0085-F2 600 cm Reg

RUN # 360E21

Operation	Description	Date Verified
4a	Operation 1 – Staging and Preconditioning - Minimum 12 hours	HA 21 Oct 2020
	Operation 2 – EO Gas Consumable Preparation	HA 21 Oct 2020
	Operation 3 – Tote Preparation and Bag Sealing	HA 21 Oct 2020
4b	Operation 4 – Transfer and Load Time	HA 21 Oct 2020
4c	Operation 5 – Sterilization/Aeration	HA 21 Oct 2020
	Operation 6 – Dosimeter Check	HA 21 Oct 2020
PRODUCTION REVIEW COMPLETED BY/DATE: Rachel Stalder 21 Oct 2020		

Acceptance Criteria and Final Release			
The sterilization tote contains appropriate qty of devices between the minimum and maximum specification.	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	The environment within the EO sterilization bag is $\geq 30\%$	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
The average preconditioning environment is ≥ 12 hours at $\geq 68^\circ\text{F}$ and $\geq 30\%\text{RH}$.	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	Devices are aerated within the sterilizer set to 50°C for 24 hours.	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
Temperature Range is between 33°C to 55°C measured from the dataloggers in the totes.	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	All Dosimeters (for all bags) pass the calibration triangle.	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
The minimum sterilization time requirement is ≥ 19 Hours to a maximum sterilization time of ≤ 38 hours and 10 minutes.	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	The Bis are negative (remain orange).	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
Two (2) Humidchips were placed into every sterilization bag.	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	The positive control turns positive (turbid and yellow).	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
The sterilizer cabinet temperature is set to 50°C	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	The incubation time/temperature for the Bis is 7 days at $35\text{--}39^\circ\text{C}$.	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
EO gas release is $\geq 22\text{g}$ but $\leq 45\text{g}$	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No		
Comments: <input checked="" type="checkbox"/> N/A			
QA Release: 029 Oct 2020			

Document ID: SOP-MED-0098
Document Type: DHR

CHEMENCE[®]
MEDICAL

Name:
Version

EO Sterilization Router
2.0

Verify that totes are arranged as outlined in WI-MED-0063

Operation 1 - Staging for EO Sterilization

Small Blister Maximum Quantity Is 1,500 Blisters Per Tote

Exofin[®] Fusion Max Is 56 Pouches In Each Tote

The Minimum Quantity for Both Fusion and Small Blisters are 4 Blisters In Each Tote

Lot#: 040482-R5	Tote #: 1-5	Qty: 6,375	N/A	<input type="checkbox"/>
Lot#: 039482-R3	Tote #: 6-9	Qty: 5,098	N/A	<input type="checkbox"/>
Lot#: 6085-F2	Tote #: 10-18	Qty: 482	N/A	<input type="checkbox"/>
Lot#: Damage	Tote #: 19-27	Qty: 27,000	N/A	<input checked="" type="checkbox"/>
Lot#:	Tote #:	Qty:	N/A	<input checked="" type="checkbox"/>

Operation 1 - Preconditioning * (Minimum 12 hours) and Data Logger Setup

Preconditioning Room N/A ☐

Preconditioning Chamber N/A ☒

EO Room Data Logger	Data Logger ID	R17771	Calibration Due Date	Dec 2020
Tote #	Data Logger ID	R17757	Calibration Due Date	Dec 2020
Tote #	Data Logger ID	R17708	Calibration Due Date	Dec 2020

Precondition in Preconditioning Room

Start Time	12:30am	N/A <input type="checkbox"/>	Start Date	13Oct2020	N/A <input type="checkbox"/>	End Time	12:30pm	N/A <input type="checkbox"/>	End Date	13Oct2020	N/A <input type="checkbox"/>
Avg. Temperature ($\geq 68^{\circ}\text{F}$)	N/A <input type="checkbox"/>	75.73	$^{\circ}\text{F}$	Avg. Relative Humidity ($\geq 30\%$)	N/A <input type="checkbox"/>	56.49	99	ER13Oct2020	%		

If preconditioning is performed in the chamber then prepare all totes with consumables and have QA check operation 2 and 3.

Operator Signature/Date	Elizabeth Roeder 13Oct2020	Checked By/Date	Imanuela Beltran 13Oct2020
QA Signature/Date	Yahra McLenore 13Oct 2020		


* EE: ER 13Oct2020

☒ Pass ☐ Fail*

R17771 MultiChannel Preconditioning

Preconditioning
Lot Numbers:
040482-R5
039482-R3
0085-F2 60CM REG

Run Number:
360E21



Device Name: RHTemp1000IS
Device Description: Intrinsically Safe Temperature and Humidity Data Logger
Serial Number: R17771
Device ID: R17771
Preconditioning: MultiChannel
Lot Numbers:
 040482-R5
 039482-R3
 0085-F2 60CM REG
Run Number:
 360E 21

Date	Time	Channel 1 Temperature (°F)	Channel 2 Humidity (% RH)
10/13/2020	12:30:49 AM	75.258	56.7
10/13/2020	12:35:49 AM	75.258	56.7
10/13/2020	12:40:49 AM	75.24	56.6
10/13/2020	12:45:49 AM	75.24	56.6
10/13/2020	12:50:49 AM	75.24	56.5
10/13/2020	12:55:49 AM	75.24	56.6
10/13/2020	1:00:49 AM	75.24	56.6
10/13/2020	1:05:49 AM	75.24	56.6
10/13/2020	1:10:49 AM	75.24	56.7
10/13/2020	1:15:49 AM	75.24	56.6
10/13/2020	1:20:49 AM	75.24	56.5
10/13/2020	1:25:49 AM	75.222	56.5
10/13/2020	1:30:49 AM	75.24	56.5
10/13/2020	1:35:49 AM	75.222	56.4
10/13/2020	1:40:49 AM	75.222	56.5
10/13/2020	1:45:49 AM	75.222	56.5
10/13/2020	1:50:49 AM	75.24	56.5
10/13/2020	1:55:49 AM	75.222	56.5
10/13/2020	2:00:49 AM	75.222	56.3
10/13/2020	2:05:49 AM	75.204	56.3
10/13/2020	2:10:49 AM	75.204	56.3
10/13/2020	2:15:49 AM	75.204	56.3
10/13/2020	2:20:49 AM	75.204	56.2
10/13/2020	2:25:49 AM	75.204	56.2
10/13/2020	2:30:49 AM	75.204	56.2
10/13/2020	2:35:49 AM	75.204	56.3
10/13/2020	2:40:49 AM	75.204	56.3
10/13/2020	2:45:49 AM	75.186	56.2
10/13/2020	2:50:49 AM	75.186	56.1
10/13/2020	2:55:49 AM	75.186	56.2
10/13/2020	3:00:49 AM	75.186	56.1
10/13/2020	3:05:49 AM	75.186	56.2
10/13/2020	3:10:49 AM	75.186	56.2
10/13/2020	3:15:49 AM	75.186	56.2
10/13/2020	3:20:49 AM	75.204	56.3
10/13/2020	3:25:49 AM	75.204	56.2
10/13/2020	3:30:49 AM	75.186	56.1
10/13/2020	3:35:49 AM	75.186	56.1
10/13/2020	3:40:49 AM	75.186	56
10/13/2020	3:45:49 AM	75.186	56
10/13/2020	3:50:49 AM	75.186	56
10/13/2020	3:55:49 AM	75.168	56.1
10/13/2020	4:00:49 AM	75.186	56.1
10/13/2020	4:05:49 AM	75.186	56.1
10/13/2020	4:10:49 AM	75.186	56
10/13/2020	4:15:49 AM	75.168	56
10/13/2020	4:20:49 AM	75.168	55.9
10/13/2020	4:25:49 AM	75.168	55.9
10/13/2020	4:30:49 AM	75.168	55.9
10/13/2020	4:35:49 AM	75.168	56
10/13/2020	4:40:49 AM	75.168	56
10/13/2020	4:45:49 AM	75.168	56
10/13/2020	4:50:49 AM	75.168	56.1
10/13/2020	4:55:49 AM	75.168	56
10/13/2020	5:00:49 AM	75.168	55.9
10/13/2020	5:05:49 AM	75.15	55.9
10/13/2020	5:10:49 AM	75.15	55.9

10/13/2020	5:15:49 AM	75.168	55.9
10/13/2020	5:20:49 AM	75.15	55.9
10/13/2020	5:25:49 AM	75.15	55.9
10/13/2020	5:30:49 AM	75.168	56
10/13/2020	5:35:49 AM	75.168	56
10/13/2020	5:40:49 AM	75.15	55.9
10/13/2020	5:45:49 AM	75.15	55.8
10/13/2020	5:50:49 AM	75.15	55.7
10/13/2020	5:55:49 AM	75.132	55.7
10/13/2020	6:00:49 AM	75.15	55.8
10/13/2020	6:05:49 AM	75.168	55.7
10/13/2020	6:10:49 AM	75.222	55.9
10/13/2020	6:15:49 AM	75.276	56.1
10/13/2020	6:20:49 AM	75.312	56.1
10/13/2020	6:25:49 AM	75.348	56.1
10/13/2020	6:30:49 AM	75.384	55.9
10/13/2020	6:35:49 AM	75.456	56.1
10/13/2020	6:40:49 AM	75.51	56.1
10/13/2020	6:45:49 AM	75.564	56.3
10/13/2020	6:50:49 AM	75.618	56.3
10/13/2020	6:55:49 AM	75.672	56.3
10/13/2020	7:00:49 AM	75.708	56.3
10/13/2020	7:05:49 AM	75.78	56.7
10/13/2020	7:10:49 AM	75.834	56.7
10/13/2020	7:15:49 AM	75.87	56.8
10/13/2020	7:20:49 AM	75.906	56.9
10/13/2020	7:25:49 AM	75.96	56.9
10/13/2020	7:30:49 AM	75.996	57
10/13/2020	7:35:49 AM	76.032	57.1
10/13/2020	7:40:49 AM	76.05	57
10/13/2020	7:45:49 AM	76.086	56.9
10/13/2020	7:50:49 AM	76.104	57
10/13/2020	7:55:49 AM	76.122	56.9
10/13/2020	8:00:49 AM	76.122	57.2
10/13/2020	8:05:49 AM	76.14	56.9
10/13/2020	8:10:49 AM	76.158	57
10/13/2020	8:15:49 AM	76.212	57.1
10/13/2020	8:20:49 AM	76.23	57.4
10/13/2020	8:25:49 AM	76.284	57.8
10/13/2020	8:30:49 AM	76.302	58
10/13/2020	8:35:49 AM	76.302	57.9
10/13/2020	8:40:49 AM	76.302	57.7
10/13/2020	8:45:49 AM	76.302	57.6
10/13/2020	8:50:49 AM	76.284	57.6
10/13/2020	8:55:49 AM	76.284	57.6
10/13/2020	9:00:49 AM	76.284	57.6
10/13/2020	9:05:49 AM	76.284	57.6
10/13/2020	9:10:49 AM	76.266	57.8
10/13/2020	9:15:49 AM	76.266	57.9
10/13/2020	9:20:49 AM	76.248	57.9
10/13/2020	9:25:49 AM	76.266	57.9
10/13/2020	9:30:49 AM	76.266	57.8
10/13/2020	9:35:49 AM	76.284	58
10/13/2020	9:40:49 AM	76.302	58.1
10/13/2020	9:45:49 AM	76.32	58
10/13/2020	9:50:49 AM	76.338	58.1
10/13/2020	9:55:49 AM	76.338	58.1
10/13/2020	10:00:49 AM	76.338	58
10/13/2020	10:05:49 AM	76.338	58
10/13/2020	10:10:49 AM	76.338	58.1
10/13/2020	10:15:49 AM	76.374	58.2
10/13/2020	10:20:49 AM	76.356	58.5
10/13/2020	10:25:49 AM	76.392	58.5
10/13/2020	10:30:49 AM	76.392	58.7
10/13/2020	10:35:49 AM	76.41	58.7
10/13/2020	10:40:49 AM	76.41	58.6
10/13/2020	10:45:49 AM	76.41	58.5
10/13/2020	10:50:49 AM	76.41	58.4
10/13/2020	10:55:49 AM	76.428	58.4
10/13/2020	11:00:49 AM	76.428	58.4
10/13/2020	11:05:49 AM	76.464	58.5
10/13/2020	11:10:49 AM	76.518	58.8
10/13/2020	11:15:49 AM	76.536	58.9
10/13/2020	11:20:49 AM	76.536	58.8
10/13/2020	11:25:49 AM	76.59	58.9
10/13/2020	11:30:49 AM	76.626	58.9
10/13/2020	11:35:49 AM	76.644	58.7
10/13/2020	11:40:49 AM	76.626	58.8
10/13/2020	11:45:49 AM	76.644	58.9
10/13/2020	11:50:49 AM	76.644	58.8
10/13/2020	11:55:49 AM	76.626	58.6
10/13/2020	12:00:49 PM	76.572	58.5
10/13/2020	12:05:49 PM	76.518	58.3
10/13/2020	12:10:49 PM	76.482	58.3
10/13/2020	12:15:49 PM	76.464	58.2
10/13/2020	12:20:49 PM	76.446	58.2
10/13/2020	12:25:49 PM	76.428	58.4
10/13/2020	12:30:49 PM	76.824	59.1

75.72811034

56.98689655

Operation 2 - Data Logger Setup for Sterilization

Preconditioning Room N/A ☐

Preconditioning Chamber N/A ☒

Location #	010E21	Data Logger ID	R17757	Calibration Due Date	Dec 2020	Programmed Start Date/Time (include Quickstart)	13 Oct 2020 12:30 pm
Location #	170E21	Data Logger ID	R17768	Calibration Due Date	Dec 2020	Programmed Start Date/Time (include Quickstart)	13 Oct 2020 12:29 pm

Operation 2 - EO Gas Consumable Preparation

EO Gas Canister Pre-Weight Information	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	Box labels are printed and match each tote:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
EO Bag Information Recorded	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	PCDs are placed in the center of the load:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
EO Gas Canister Labeled and mounted to tote (s) with product	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	Positive Control is the same Lot number as PCDs in totes	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
EZ Test Bis not damaged, Labeled, inside PCD	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	Dosimeter is labeled and placed in the center of the load	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
Humidichips (2) placed inside Humiditube inside the tote	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	Equipment Logbooks filled out	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
2 Data loggers placed into the front of the tote (hottest and coolest locations)	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No		
Operator Signature/date:	<i>Elizabeth Rowler</i> 13 Oct 2020	Checked By/date:	<i>Manuela Beltran</i> 13 Oct 2020

Sterilization Consumable Lot Information

EO Gas lot number(s)	10072	Exp. Date	31 Jan 2023	Dosimeter lot number(s)	204517	Exp. Date	12 Feb 2023
EOG bag lot number(s)	N/A ER130CT2020	Exp. Date	N/A 13 Oct 2020	Humidichip lot number(s)	204610	Exp. Date	19 Feb 2023
*BI lot number(s)	G-214	Exp. Date	13 Jun 2021	*If multiple lots used, 1 PC from each lot must be present			
Scale/Balance ID:	0035805163	CAL Due:	Oct 2020				

Document ID: SOP-MED-0098
Document Type: DHR

CHEMENCE
MEDICAL

Name:
Version

EO Sterilization Router
2.0

Operation 3 - Tote Preparation and Bag Sealing

EO Bag Sealer Asset ID I - 28434

EO Bag Sealer Calibration Due Date

OCT 2020

QA Line Clearance - Proceed to bag sealing once QA gives approval

All steps in Operation 2 are complete and all consumables are loaded into each tote

QA Signature/Date:

Yahna McLenore 13 OCT 2020

Operation 4 - Transfer & Load Time

Preconditioning in Chamber N/A

☐ Yes ☐ No

Printer Paper/Ink Checked

☒ Yes ☐ No

Chamber Temperature is Set at 30°C

Start Date

Operator Signature/Date

Checked By/Date

End Time

End Date

Operator Signature/Date

Checked By/Date

Sterilization In Chamber

Time left "white room" area (b)

2:48pm N/A ☐

Sterilizer EOE Temp set to 50°C

☒ Yes ☐ No

Printer Paper/Ink Checked:

☒ Yes ☐ No

Activate all EO Cartridges

Time Last Cartridge Activated (a):

2:53pm

Total Transfer Time (a-b) should be ≤ 17 minutes:

8 (5) min

Time/Date that cycle started (doors Shut):

2:54pm 13 OCT 2020

Operator Signature/Date

Checked By/Date

Yahna McLenore 13 OCT 2020
Elyse R. Miller 20 OCT 2020 for 13 OCT 2020

Operation 5 - Sterilization / Aeration

Time sterilizer "UNLOAD" pressed
(Minimum 19 hours from "Cycle Start"):

10:19 am

Sterilize end time (After
countdown/doors unlocked):

10:25 am

Minimum Aeration time is \geq 4 hours

Aeration start time (after countdown completed/doors locked)

10:27 am

Aeration End Time (when countdown is completed/doors unlocked)

2:27 pm

Operator Signature/Date

Wanda LaBelle 14 OCT 2020

Checked By/Date

Wanda LaBelle 14 OCT 2020

Operation 6 - EO Data Logger Downloading

Data Logger Offloaded as outlined in WI-MED-0018 - Record Lot Number and location # (Control Number) on Printouts

Include the start date and time on data logger printout.

Include the end time on the data logger printout.

Average Relative Humidity Must Be $> 30\%$ (for totes with product)


Location #	Data Logger ID	R17757	Start Date/Time	End Date/Time	Avg. Relative Humidity	Sterilizer Temp	Avg. Relative Humidity	Sterilizer Temp
			Start Date/Time	End Date/Time	13 OCT 2020 2:54 pm			
010E21	R17757		14 OCT 2020 2:29 pm	13 OCT 2020 2:54 pm	40.93	40.93	55.78	50.95
170E21	R17768		14 OCT 2020 2:29 pm	14 OCT 2020 2:29 pm	50.95			

BI Incubation

Bis submitted to QC		<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No		TR #	BI Lot #	G-214	
Attach EO Chamber Printout (Preconditioning)		MIN	11/19 150C+2020	°C	MAX	11/19 150C+2020	°C
Attach EO Chamber Printout (Sterilization)		MIN	*50 45	°C	MAX	50	°C
						Printout Attached	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No

*EE: ER150C+2020
*EE: ER150C+2020

R17757 MultiChannel 010E21



—●— R17757 - Temperature (°C)
—■— R17757 - Humidity (% RH)

Device Name:	RHTemp1000IS	RHTemp1000IS
Device Description:	Intrinsically Safe Temperature and Humidity Data Logger	Intrinsically Safe Temperature and Humidity Data Logger
Serial Number:	R17757	R17757
Device ID:	010E21 MultiChannel	MultiChannel

Date	Time	Channel 1 Temperature (°C)	Channel 2 Humidity (% RH)
10/13/2020	2:54:44 PM	25.08	63.5
10/13/2020	2:59:44 PM	25.22	63.6
10/13/2020	3:04:44 PM	25.46	63.6
10/13/2020	3:09:44 PM	25.78	62.7
10/13/2020	3:14:44 PM	26.22	62.4
10/13/2020	3:19:44 PM	26.71	61.9
10/13/2020	3:24:44 PM	27.2	61.2
10/13/2020	3:29:44 PM	27.72	60.3
10/13/2020	3:34:44 PM	28.24	59.5
10/13/2020	3:39:44 PM	28.76	58.6
10/13/2020	3:44:44 PM	29.27	57.8
10/13/2020	3:49:44 PM	29.78	57
10/13/2020	3:54:44 PM	30.28	56.2
10/13/2020	3:59:44 PM	30.77	55.4
10/13/2020	4:04:44 PM	31.24	54.8
10/13/2020	4:09:44 PM	31.7	54.3
10/13/2020	4:14:44 PM	32.14	53.8
10/13/2020	4:19:44 PM	32.56	53.4
10/13/2020	4:24:44 PM	32.96	53.1
10/13/2020	4:29:44 PM	33.33	52.7
10/13/2020	4:34:44 PM	33.69	52.5
10/13/2020	4:39:44 PM	34.03	52.3
10/13/2020	4:44:44 PM	34.35	52.2
10/13/2020	4:49:44 PM	34.66	52.3
10/13/2020	4:54:44 PM	34.95	52.6
10/13/2020	4:59:44 PM	35.22	53
10/13/2020	5:04:44 PM	35.46	53.3
10/13/2020	5:09:44 PM	35.68	53.4
10/13/2020	5:14:44 PM	35.89	53.6
10/13/2020	5:19:44 PM	36.09	53.7
10/13/2020	5:24:44 PM	36.29	53.9
10/13/2020	5:29:44 PM	36.48	54
10/13/2020	5:34:44 PM	36.66	54.2
10/13/2020	5:39:44 PM	36.84	54.3
10/13/2020	5:44:44 PM	37.02	54.4
10/13/2020	5:49:44 PM	37.19	54.5
10/13/2020	5:54:44 PM	37.36	54.6
10/13/2020	5:59:44 PM	37.53	54.8
10/13/2020	6:04:44 PM	37.69	54.9
10/13/2020	6:09:44 PM	37.84	55.1
10/13/2020	6:14:44 PM	37.99	55.2
10/13/2020	6:19:44 PM	38.13	55.3
10/13/2020	6:24:44 PM	38.27	55.5
10/13/2020	6:29:44 PM	38.39	55.6
10/13/2020	6:34:44 PM	38.52	55.8
10/13/2020	6:39:44 PM	38.64	55.9
10/13/2020	6:44:44 PM	38.77	56.3
10/13/2020	6:49:44 PM	38.87	56.8
10/13/2020	6:54:44 PM	38.98	57.2
10/13/2020	6:59:44 PM	39.07	57.5
10/13/2020	7:04:44 PM	39.17	57.7
10/13/2020	7:09:44 PM	39.26	58
10/13/2020	7:14:44 PM	39.36	58.2
10/13/2020	7:19:44 PM	39.45	58.4
10/13/2020	7:24:44 PM	39.53	58.6
10/13/2020	7:29:44 PM	39.62	58.8
10/13/2020	7:34:44 PM	39.71	59
10/13/2020	7:39:44 PM	39.79	59.2
10/13/2020	7:44:44 PM	39.87	59.3
10/13/2020	7:49:44 PM	39.96	59.4
10/13/2020	7:54:44 PM	40.04	59.5
10/13/2020	7:59:44 PM	40.12	59.7
10/13/2020	8:04:44 PM	40.2	59.8
10/13/2020	8:09:44 PM	40.28	59.9
10/13/2020	8:14:44 PM	40.36	59.9
10/13/2020	8:19:44 PM	40.43	60
10/13/2020	8:24:44 PM	40.51	60.1


10/13/2020	8:29:44 PM	40.58	60.2
10/13/2020	8:34:44 PM	40.65	60.2
10/13/2020	8:39:44 PM	40.72	60.3
10/13/2020	8:44:44 PM	40.79	60.4
10/13/2020	8:49:44 PM	40.86	60.5
10/13/2020	8:54:44 PM	40.92	60.6
10/13/2020	8:59:44 PM	40.99	60.7
10/13/2020	9:04:44 PM	41.05	60.8
10/13/2020	9:09:44 PM	41.11	60.8
10/13/2020	9:14:44 PM	41.18	60.9
10/13/2020	9:19:44 PM	41.23	61
10/13/2020	9:24:44 PM	41.28	61.1
10/13/2020	9:29:44 PM	41.34	61.2
10/13/2020	9:34:44 PM	41.39	61.3
10/13/2020	9:39:44 PM	41.44	61.4
10/13/2020	9:44:44 PM	41.49	61.5
10/13/2020	9:49:44 PM	41.54	61.6
10/13/2020	9:54:44 PM	41.59	61.7
10/13/2020	9:59:44 PM	41.63	61.7
10/13/2020	10:04:44 PM	41.68	61.8
10/13/2020	10:09:44 PM	41.72	61.9
10/13/2020	10:14:44 PM	41.76	61.9
10/13/2020	10:19:44 PM	41.8	62
10/13/2020	10:24:44 PM	41.84	62.1
10/13/2020	10:29:44 PM	41.87	62.2
10/13/2020	10:34:44 PM	41.91	62.3
10/13/2020	10:39:44 PM	41.94	62.4
10/13/2020	10:44:44 PM	41.97	62.4
10/13/2020	10:49:44 PM	42	62.5
10/13/2020	10:54:44 PM	42.03	62.6
10/13/2020	10:59:44 PM	42.06	62.7
10/13/2020	11:04:44 PM	42.09	62.8
10/13/2020	11:09:44 PM	42.11	62.8
10/13/2020	11:14:44 PM	42.14	62.9
10/13/2020	11:19:44 PM	42.16	63
10/13/2020	11:24:44 PM	42.18	63
10/13/2020	11:29:44 PM	42.21	63.1
10/13/2020	11:34:44 PM	42.23	63.2
10/13/2020	11:39:44 PM	42.25	63.3
10/13/2020	11:44:44 PM	42.27	63.4
10/13/2020	11:49:44 PM	42.29	63.4
10/13/2020	11:54:44 PM	42.31	63.5
10/13/2020	11:59:44 PM	42.33	63.6
10/14/2020	12:04:44 AM	42.35	63.6
10/14/2020	12:09:44 AM	42.37	63.7
10/14/2020	12:14:44 AM	42.39	63.8
10/14/2020	12:19:44 AM	42.41	63.8
10/14/2020	12:24:44 AM	42.42	63.9
10/14/2020	12:29:44 AM	42.44	64
10/14/2020	12:34:44 AM	42.45	64
10/14/2020	12:39:44 AM	42.47	64.1
10/14/2020	12:44:44 AM	42.48	64.1
10/14/2020	12:49:44 AM	42.49	64.2
10/14/2020	12:54:44 AM	42.51	64.2
10/14/2020	12:59:44 AM	42.51	64.3
10/14/2020	1:04:44 AM	42.52	64.4
10/14/2020	1:09:44 AM	42.54	64.4
10/14/2020	1:14:44 AM	42.54	64.5
10/14/2020	1:19:44 AM	42.55	64.5
10/14/2020	1:24:44 AM	42.56	64.6
10/14/2020	1:29:44 AM	42.56	64.6
10/14/2020	1:34:44 AM	42.57	64.7
10/14/2020	1:39:44 AM	42.57	64.8
10/14/2020	1:44:44 AM	42.58	64.8
10/14/2020	1:49:44 AM	42.59	64.9
10/14/2020	1:54:44 AM	42.59	64.9
10/14/2020	1:59:44 AM	42.6	65
10/14/2020	2:04:44 AM	42.6	65
10/14/2020	2:09:44 AM	42.61	65.1
10/14/2020	2:14:44 AM	42.62	65.1
10/14/2020	2:19:44 AM	42.63	65.1
10/14/2020	2:24:44 AM	42.64	65.2
10/14/2020	2:29:44 AM	42.65	65.2
10/14/2020	2:34:44 AM	42.67	65.3

10/14/2020	2:39:44 AM	42.69	65.3
10/14/2020	2:44:44 AM	42.7	65.3
10/14/2020	2:49:44 AM	42.72	65.4
10/14/2020	2:54:44 AM	42.74	65.4
10/14/2020	2:59:44 AM	42.76	65.4
10/14/2020	3:04:44 AM	42.78	65.4
10/14/2020	3:09:44 AM	42.8	65.4
10/14/2020	3:14:44 AM	42.82	65.4
10/14/2020	3:19:44 AM	42.85	65.4
10/14/2020	3:24:44 AM	42.87	65.4
10/14/2020	3:29:44 AM	42.88	65.4
10/14/2020	3:34:44 AM	42.91	65.4
10/14/2020	3:39:44 AM	42.93	65.4
10/14/2020	3:44:44 AM	42.95	65.4
10/14/2020	3:49:44 AM	42.97	65.4
10/14/2020	3:54:44 AM	42.99	65.4
10/14/2020	3:59:44 AM	43.01	65.4
10/14/2020	4:04:44 AM	43.03	65.4
10/14/2020	4:09:44 AM	43.05	65.4
10/14/2020	4:14:44 AM	43.06	65.4
10/14/2020	4:19:44 AM	43.08	65.4
10/14/2020	4:24:44 AM	43.09	65.5
10/14/2020	4:29:44 AM	43.11	65.5
10/14/2020	4:34:44 AM	43.12	65.5
10/14/2020	4:39:44 AM	43.13	65.5
10/14/2020	4:44:44 AM	43.14	65.5
10/14/2020	4:49:44 AM	43.15	65.5
10/14/2020	4:54:44 AM	43.17	65.5
10/14/2020	4:59:44 AM	43.18	65.5
10/14/2020	5:04:44 AM	43.2	65.6
10/14/2020	5:09:44 AM	43.22	65.6
10/14/2020	5:14:44 AM	43.22	65.6
10/14/2020	5:19:44 AM	43.24	65.6
10/14/2020	5:24:44 AM	43.25	65.6
10/14/2020	5:29:44 AM	43.26	65.6
10/14/2020	5:34:44 AM	43.26	65.6
10/14/2020	5:39:44 AM	43.27	65.7
10/14/2020	5:44:44 AM	43.28	65.7
10/14/2020	5:49:44 AM	43.29	65.7
10/14/2020	5:54:44 AM	43.3	65.7
10/14/2020	5:59:44 AM	43.3	65.7
10/14/2020	6:04:44 AM	43.3	65.8
10/14/2020	6:09:44 AM	43.31	65.8
10/14/2020	6:14:44 AM	43.31	65.8
10/14/2020	6:19:44 AM	43.32	65.8
10/14/2020	6:24:44 AM	43.32	65.8
10/14/2020	6:29:44 AM	43.33	65.8
10/14/2020	6:34:44 AM	43.33	65.9
10/14/2020	6:39:44 AM	43.34	65.9
10/14/2020	6:44:44 AM	43.34	65.9
10/14/2020	6:49:44 AM	43.34	66
10/14/2020	6:54:44 AM	43.34	66
10/14/2020	6:59:44 AM	43.34	66
10/14/2020	7:04:44 AM	43.35	66
10/14/2020	7:09:44 AM	43.35	66
10/14/2020	7:14:44 AM	43.35	66
10/14/2020	7:19:44 AM	43.35	66.1
10/14/2020	7:24:44 AM	43.36	66.1
10/14/2020	7:29:44 AM	43.37	66.1
10/14/2020	7:34:44 AM	43.37	66.1
10/14/2020	7:39:44 AM	43.38	66.1
10/14/2020	7:44:44 AM	43.38	66.1
10/14/2020	7:49:44 AM	43.39	66.1
10/14/2020	7:54:44 AM	43.4	66.1
10/14/2020	7:59:44 AM	43.4	66.2
10/14/2020	8:04:44 AM	43.41	66.2
10/14/2020	8:09:44 AM	43.42	66.2
10/14/2020	8:14:44 AM	43.42	66.2
10/14/2020	8:19:44 AM	43.42	66.2
10/14/2020	8:24:44 AM	43.43	66.2
10/14/2020	8:29:44 AM	43.44	66.2
10/14/2020	8:34:44 AM	43.44	66.2
10/14/2020	8:39:44 AM	43.45	66.2
10/14/2020	8:44:44 AM	43.45	66.2

10/14/2020	8:49:44 AM	43.46	66.2
10/14/2020	8:54:44 AM	43.46	66.3
10/14/2020	8:59:44 AM	43.47	66.3
10/14/2020	9:04:44 AM	43.47	66.3
10/14/2020	9:09:44 AM	43.48	66.3
10/14/2020	9:14:44 AM	43.49	66.3
10/14/2020	9:19:44 AM	43.49	66.3
10/14/2020	9:24:44 AM	43.5	66.3
10/14/2020	9:29:44 AM	43.51	66.3
10/14/2020	9:34:44 AM	43.51	66.3
10/14/2020	9:39:44 AM	43.52	66.3
10/14/2020	9:44:44 AM	43.52	66.3

10/14/2020	9:49:44 AM	43.54	66.3
10/14/2020	9:54:44 AM	43.54	66.3
10/14/2020	9:59:44 AM	43.55	66.3
10/14/2020	10:04:44 AM	43.56	66.3
10/14/2020	10:09:44 AM	43.57	66.3
10/14/2020	10:14:44 AM	43.57	66.3
10/14/2020	10:19:44 AM	43.57	66.2
10/14/2020	10:24:44 AM	43.55	66.1
10/14/2020	10:29:44 AM	43.48	65.8
10/14/2020	10:34:44 AM	43.4	65.9
10/14/2020	10:39:44 AM	43.31	66
10/14/2020	10:44:44 AM	43.21	66.1
10/14/2020	10:49:44 AM	43.11	66.3
10/14/2020	10:54:44 AM	43.02	66.3
10/14/2020	10:59:44 AM	42.94	66.3
10/14/2020	11:04:44 AM	42.88	66.4
10/14/2020	11:09:44 AM	42.81	66.5
10/14/2020	11:14:44 AM	42.75	66.5
10/14/2020	11:19:44 AM	42.7	66.5
10/14/2020	11:24:44 AM	42.66	66.5
10/14/2020	11:29:44 AM	42.62	66.5
10/14/2020	11:34:44 AM	42.58	66.5
10/14/2020	11:39:44 AM	42.55	66.4
10/14/2020	11:44:44 AM	42.52	66.4
10/14/2020	11:49:44 AM	42.5	66.5
10/14/2020	11:54:44 AM	42.47	66.4
10/14/2020	11:59:44 AM	42.45	66.3
10/14/2020	12:04:44 PM	42.43	66.2
10/14/2020	12:09:44 PM	42.41	66.3
10/14/2020	12:14:44 PM	42.39	66.2
10/14/2020	12:19:44 PM	42.38	66.1
10/14/2020	12:24:44 PM	42.37	66.1
10/14/2020	12:29:44 PM	42.35	66.1
10/14/2020	12:34:44 PM	42.34	66
10/14/2020	12:39:44 PM	42.33	65.9
10/14/2020	12:44:44 PM	42.32	65.8
10/14/2020	12:49:44 PM	42.3	65.8
10/14/2020	12:54:44 PM	42.29	65.7
10/14/2020	12:59:44 PM	42.29	65.6
10/14/2020	1:04:44 PM	42.27	65.6
10/14/2020	1:09:44 PM	42.26	65.6
10/14/2020	1:14:44 PM	42.25	65.4
10/14/2020	1:19:44 PM	42.26	65.4
10/14/2020	1:24:44 PM	42.24	64.7
10/14/2020	1:29:44 PM	42.24	64.7
10/14/2020	1:34:44 PM	42.25	64.8
10/14/2020	1:39:44 PM	42.24	64.8
10/14/2020	1:44:44 PM	42.23	64.8
10/14/2020	1:49:44 PM	42.22	64.7
10/14/2020	1:54:44 PM	42.23	64.6
10/14/2020	1:59:44 PM	42.23	64.6
10/14/2020	2:04:44 PM	42.22	64.6
10/14/2020	2:09:44 PM	42.22	64.4
10/14/2020	2:14:44 PM	42.23	64.4
10/14/2020	2:19:44 PM	42.22	64.3
10/14/2020	2:24:44 PM	42.22	64.6
10/14/2020	2:29:44 PM	42.15	57.3
		40.92510563	62.92077465

R17768 MultiChannel 170E21



Device Name:	RHTemp1000IS	RHTemp1000IS
Device Description:	Intrinsically Safe Temperature and Humidity Data Logger	Intrinsically Safe Temperature and Humidity Data Logger
Serial Number:	R17768	R17768
Device ID:	MultiChannel	MultiChannel

Date	Time	Channel 1 Temperature (°C)	Channel 2 Humidity (% RH)
10/13/2020	2:54:43 PM	24.87	63.2
10/13/2020	2:59:43 PM	25.35	64.7
10/13/2020	3:04:43 PM	26.23	64.3
10/13/2020	3:09:43 PM	27.41	61.3
10/13/2020	3:14:43 PM	28.79	59.1
10/13/2020	3:19:43 PM	30.31	57.2
10/13/2020	3:24:43 PM	31.88	55.3
10/13/2020	3:29:43 PM	33.43	53.7
10/13/2020	3:34:43 PM	34.93	52.3
10/13/2020	3:39:43 PM	36.37	51.2
10/13/2020	3:44:43 PM	37.73	50.2
10/13/2020	3:49:43 PM	39.03	49.3
10/13/2020	3:54:43 PM	40.24	48.5
10/13/2020	3:59:43 PM	41.37	47.9
10/13/2020	4:04:43 PM	42.42	47.4
10/13/2020	4:09:43 PM	43.37	46.9
10/13/2020	4:14:43 PM	44.26	46.7
10/13/2020	4:19:43 PM	45.07	46.4
10/13/2020	4:24:43 PM	45.8	46.3
10/13/2020	4:29:43 PM	46.47	46.2
10/13/2020	4:34:43 PM	47.07	46.2
10/13/2020	4:39:43 PM	47.6	46.3
10/13/2020	4:44:43 PM	48.09	46.4
10/13/2020	4:49:43 PM	48.52	46.6
10/13/2020	4:54:43 PM	48.9	46.9
10/13/2020	4:59:43 PM	49.25	47.1
10/13/2020	5:04:43 PM	49.56	47.4
10/13/2020	5:09:43 PM	49.84	47.8
10/13/2020	5:14:43 PM	50.1	48.2
10/13/2020	5:19:43 PM	50.33	48.6
10/13/2020	5:24:43 PM	50.53	49
10/13/2020	5:29:43 PM	50.71	49.4
10/13/2020	5:34:43 PM	50.87	49.9
10/13/2020	5:39:43 PM	51.02	50.4
10/13/2020	5:44:43 PM	51.15	51
10/13/2020	5:49:43 PM	51.28	51.9
10/13/2020	5:54:43 PM	51.4	52.8
10/13/2020	5:59:43 PM	51.51	53.6
10/13/2020	6:04:43 PM	51.61	54.3
10/13/2020	6:09:43 PM	51.69	54.9
10/13/2020	6:14:43 PM	51.77	55.4
10/13/2020	6:19:43 PM	51.84	55.9
10/13/2020	6:24:43 PM	51.9	56.2
10/13/2020	6:29:43 PM	51.96	56.6
10/13/2020	6:34:43 PM	52.01	56.9
10/13/2020	6:39:43 PM	52.06	57.2
10/13/2020	6:44:43 PM	52.1	57.6
10/13/2020	6:49:43 PM	52.13	57.9
10/13/2020	6:54:43 PM	52.16	58.2
10/13/2020	6:59:43 PM	52.19	58.4
10/13/2020	7:04:43 PM	52.21	58.5
10/13/2020	7:09:43 PM	52.24	58.7
10/13/2020	7:14:43 PM	52.27	58.9
10/13/2020	7:19:43 PM	52.28	59.1
10/13/2020	7:24:43 PM	52.3	59.3
10/13/2020	7:29:43 PM	52.32	59.5
10/13/2020	7:34:43 PM	52.33	59.6
10/13/2020	7:39:43 PM	52.34	59.8
10/13/2020	7:44:43 PM	52.36	59.9
10/13/2020	7:49:43 PM	52.37	60.2
10/13/2020	7:54:43 PM	52.38	60.3
10/13/2020	7:59:43 PM	52.38	60.5
10/13/2020	8:04:43 PM	52.39	60.6

10/13/2020	8:09:43 PM	52.39	60.7
10/13/2020	8:14:43 PM	52.4	60.9
10/13/2020	8:19:43 PM	52.4	61
10/13/2020	8:24:43 PM	52.4	61.1
10/13/2020	8:29:43 PM	52.4	61.2
10/13/2020	8:34:43 PM	52.4	61.4
10/13/2020	8:39:43 PM	52.4	61.6
10/13/2020	8:44:43 PM	52.4	61.8
10/13/2020	8:49:43 PM	52.39	61.9
10/13/2020	8:54:43 PM	52.39	61.9
10/13/2020	8:59:43 PM	52.39	62
10/13/2020	9:04:43 PM	52.39	62
10/13/2020	9:09:43 PM	52.39	62.1
10/13/2020	9:14:43 PM	52.4	62.1
10/13/2020	9:19:43 PM	52.39	62.2
10/13/2020	9:24:43 PM	52.39	62.3
10/13/2020	9:29:43 PM	52.38	62.5
10/13/2020	9:34:43 PM	52.37	62.6
10/13/2020	9:39:43 PM	52.37	62.6
10/13/2020	9:44:43 PM	52.37	62.7
10/13/2020	9:49:43 PM	52.37	62.8
10/13/2020	9:54:43 PM	52.36	62.8
10/13/2020	9:59:43 PM	52.36	62.8
10/13/2020	10:04:43 PM	52.36	62.9
10/13/2020	10:09:43 PM	52.36	62.9
10/13/2020	10:14:43 PM	52.35	62.9
10/13/2020	10:19:43 PM	52.34	62.9
10/13/2020	10:24:43 PM	52.34	63
10/13/2020	10:29:43 PM	52.34	63
10/13/2020	10:34:43 PM	52.33	63
10/13/2020	10:39:43 PM	52.33	63.1
10/13/2020	10:44:43 PM	52.33	63.1
10/13/2020	10:49:43 PM	52.32	63.2
10/13/2020	10:54:43 PM	52.32	63.2
10/13/2020	10:59:43 PM	52.31	63.3
10/13/2020	11:04:43 PM	52.3	63.3
10/13/2020	11:09:43 PM	52.3	63.4
10/13/2020	11:14:43 PM	52.3	63.4
10/13/2020	11:19:43 PM	52.29	63.4
10/13/2020	11:24:43 PM	52.29	63.4
10/13/2020	11:29:43 PM	52.28	63.4
10/13/2020	11:34:43 PM	52.28	63.5
10/13/2020	11:39:43 PM	52.27	63.5
10/13/2020	11:44:43 PM	52.27	63.5
10/13/2020	11:49:43 PM	52.26	63.5
10/13/2020	11:54:43 PM	52.26	63.5
10/13/2020	11:59:43 PM	52.26	63.5
10/14/2020	12:04:43 AM	52.26	63.5
10/14/2020	12:09:43 AM	52.26	63.5
10/14/2020	12:14:43 AM	52.25	63.5
10/14/2020	12:19:43 AM	52.24	63.5
10/14/2020	12:24:43 AM	52.24	63.5
10/14/2020	12:29:43 AM	52.23	63.5
10/14/2020	12:34:43 AM	52.23	63.6
10/14/2020	12:39:43 AM	52.22	63.6
10/14/2020	12:44:43 AM	52.22	63.5
10/14/2020	12:49:43 AM	52.22	63.6
10/14/2020	12:54:43 AM	52.22	63.6
10/14/2020	12:59:43 AM	52.22	63.6
10/14/2020	1:04:43 AM	52.22	63.6
10/14/2020	1:09:43 AM	52.21	63.6
10/14/2020	1:14:43 AM	52.21	63.7
10/14/2020	1:19:43 AM	52.2	63.7
10/14/2020	1:24:43 AM	52.2	63.7
10/14/2020	1:29:43 AM	52.19	63.7
10/14/2020	1:34:43 AM	52.19	63.7
10/14/2020	1:39:43 AM	52.19	63.7
10/14/2020	1:44:43 AM	52.19	63.7
10/14/2020	1:49:43 AM	52.19	63.7
10/14/2020	1:54:43 AM	52.18	63.7

10/14/2020	1:59:43 AM	52.19	63.7
10/14/2020	2:04:43 AM	52.18	63.7
10/14/2020	2:09:43 AM	52.17	63.7
10/14/2020	2:14:43 AM	52.17	63.7
10/14/2020	2:19:43 AM	52.17	63.7
10/14/2020	2:24:43 AM	52.17	63.6
10/14/2020	2:29:43 AM	52.17	63.6
10/14/2020	2:34:43 AM	52.17	63.6
10/14/2020	2:39:43 AM	52.17	63.6
10/14/2020	2:44:43 AM	52.17	63.6
10/14/2020	2:49:43 AM	52.17	63.6
10/14/2020	2:54:43 AM	52.17	63.6
10/14/2020	2:59:43 AM	52.16	63.6
10/14/2020	3:04:43 AM	52.17	63.6
10/14/2020	3:09:43 AM	52.16	63.6
10/14/2020	3:14:43 AM	52.16	63.6
10/14/2020	3:19:43 AM	52.15	63.6
10/14/2020	3:24:43 AM	52.15	63.6
10/14/2020	3:29:43 AM	52.15	63.6
10/14/2020	3:34:43 AM	52.14	63.6
10/14/2020	3:39:43 AM	52.15	63.6
10/14/2020	3:44:43 AM	52.15	63.6
10/14/2020	3:49:43 AM	52.15	63.6
10/14/2020	3:54:43 AM	52.15	63.6
10/14/2020	3:59:43 AM	52.15	63.6
10/14/2020	4:04:43 AM	52.15	63.6
10/14/2020	4:09:43 AM	52.15	63.6
10/14/2020	4:14:43 AM	52.14	63.6
10/14/2020	4:19:43 AM	52.14	63.6
10/14/2020	4:24:43 AM	52.13	63.6
10/14/2020	4:29:43 AM	52.12	63.7
10/14/2020	4:34:43 AM	52.12	63.7
10/14/2020	4:39:43 AM	52.11	63.7
10/14/2020	4:44:43 AM	52.11	63.6
10/14/2020	4:49:43 AM	52.12	63.6
10/14/2020	4:54:43 AM	52.12	63.6
10/14/2020	4:59:43 AM	52.12	63.6
10/14/2020	5:04:43 AM	52.12	63.6
10/14/2020	5:09:43 AM	52.12	63.6
10/14/2020	5:14:43 AM	52.11	63.6
10/14/2020	5:19:43 AM	52.11	63.6
10/14/2020	5:24:43 AM	52.11	63.6
10/14/2020	5:29:43 AM	52.1	63.6
10/14/2020	5:34:43 AM	52.1	63.6
10/14/2020	5:39:43 AM	52.09	63.6
10/14/2020	5:44:43 AM	52.09	63.6
10/14/2020	5:49:43 AM	52.08	63.6
10/14/2020	5:54:43 AM	52.07	63.6
10/14/2020	5:59:43 AM	52.07	63.6
10/14/2020	6:04:43 AM	52.06	63.6
10/14/2020	6:09:43 AM	52.06	63.5
10/14/2020	6:14:43 AM	52.05	63.5
10/14/2020	6:19:43 AM	52.05	63.5
10/14/2020	6:24:43 AM	52.05	63.5
10/14/2020	6:29:43 AM	52.04	63.5
10/14/2020	6:34:43 AM	52.04	63.5
10/14/2020	6:39:43 AM	52.03	63.5
10/14/2020	6:44:43 AM	52.03	63.5
10/14/2020	6:49:43 AM	52.02	63.5
10/14/2020	6:54:43 AM	52.02	63.5
10/14/2020	6:59:43 AM	52.02	63.5
10/14/2020	7:04:43 AM	52.01	63.6
10/14/2020	7:09:43 AM	52.01	63.6
10/14/2020	7:14:43 AM	52.01	63.6
10/14/2020	7:19:43 AM	52.01	63.6
10/14/2020	7:24:43 AM	52	63.6
10/14/2020	7:29:43 AM	52	63.5
10/14/2020	7:34:43 AM	52	63.5
10/14/2020	7:39:43 AM	52	63.5
10/14/2020	7:44:43 AM	52	63.5

10/14/2020	7:49:43 AM	52	63.5
10/14/2020	7:54:43 AM	52	63.5
10/14/2020	7:59:43 AM	52	63.5
10/14/2020	8:04:43 AM	52	63.5
10/14/2020	8:09:43 AM	52	63.5
10/14/2020	8:14:43 AM	52	63.5
10/14/2020	8:19:43 AM	52	63.5
10/14/2020	8:24:43 AM	52	63.5
10/14/2020	8:29:43 AM	52	63.4
10/14/2020	8:34:43 AM	52	63.4
10/14/2020	8:39:43 AM	52	63.4
10/14/2020	8:44:43 AM	52	63.4
10/14/2020	8:49:43 AM	52	63.4
10/14/2020	8:54:43 AM	51.99	63.4
10/14/2020	8:59:43 AM	51.99	63.3
10/14/2020	9:04:43 AM	51.99	63.3
10/14/2020	9:09:43 AM	51.99	63.3
10/14/2020	9:14:43 AM	51.99	63.3
10/14/2020	9:19:43 AM	51.99	63.3
10/14/2020	9:24:43 AM	51.99	63.2
10/14/2020	9:29:43 AM	52	63.2
10/14/2020	9:34:43 AM	52	63.2
10/14/2020	9:39:43 AM	52	63.1
10/14/2020	9:44:43 AM	52.01	63.1

10/14/2020	9:49:43 AM	52.01	63.1
10/14/2020	9:54:43 AM	52.01	63.1
10/14/2020	9:59:43 AM	52.01	63.1
10/14/2020	10:04:43 AM	52.01	63.1
10/14/2020	10:09:43 AM	52.01	63.1
10/14/2020	10:14:43 AM	52	63.1
10/14/2020	10:19:43 AM	52	63.1
10/14/2020	10:24:43 AM	51.95	61.8
10/14/2020	10:29:43 AM	51.78	51.2
10/14/2020	10:34:43 AM	51.7	44.9
10/14/2020	10:39:43 AM	51.67	43.1
10/14/2020	10:44:43 AM	51.7	41.8
10/14/2020	10:49:43 AM	51.76	39.4
10/14/2020	10:54:43 AM	51.84	38.2
10/14/2020	10:59:43 AM	51.93	37.5
10/14/2020	11:04:43 AM	52.02	36.8
10/14/2020	11:09:43 AM	52.1	36.1
10/14/2020	11:14:43 AM	52.18	35.6
10/14/2020	11:19:43 AM	52.24	35.3
10/14/2020	11:24:43 AM	52.3	34.8
10/14/2020	11:29:43 AM	52.35	34.4
10/14/2020	11:34:43 AM	52.39	33.9
10/14/2020	11:39:43 AM	52.43	33.9
10/14/2020	11:44:43 AM	52.46	33.5
10/14/2020	11:49:43 AM	52.49	32.9
10/14/2020	11:54:43 AM	52.52	32.6
10/14/2020	11:59:43 AM	52.54	32.3
10/14/2020	12:04:43 PM	52.56	31.9
10/14/2020	12:09:43 PM	52.58	31.5
10/14/2020	12:14:43 PM	52.59	31.1
10/14/2020	12:19:43 PM	52.6	31.1
10/14/2020	12:24:43 PM	52.6	30.8
10/14/2020	12:29:43 PM	52.61	30.4
10/14/2020	12:34:43 PM	52.62	30.2
10/14/2020	12:39:43 PM	52.62	30
10/14/2020	12:44:43 PM	52.63	29.7
10/14/2020	12:49:43 PM	52.64	29.5
10/14/2020	12:54:43 PM	52.65	29.1
10/14/2020	12:59:43 PM	52.66	29
10/14/2020	1:04:43 PM	52.66	28.9
10/14/2020	1:09:43 PM	52.66	28.7
10/14/2020	1:14:43 PM	52.66	28.4
10/14/2020	1:19:43 PM	52.66	28.3
10/14/2020	1:24:43 PM	52.66	28.9
10/14/2020	1:29:43 PM	52.66	28.4
10/14/2020	1:34:43 PM	52.66	28.4
10/14/2020	1:39:43 PM	52.67	28.1
10/14/2020	1:44:43 PM	52.67	27.7
10/14/2020	1:49:43 PM	52.67	27.4
10/14/2020	1:54:43 PM	52.67	27.3
10/14/2020	1:59:43 PM	52.66	27.2
10/14/2020	2:04:43 PM	52.67	27
10/14/2020	2:09:43 PM	52.66	26.9
10/14/2020	2:14:43 PM	52.67	26.7
10/14/2020	2:19:43 PM	52.68	26.4
10/14/2020	2:24:43 PM	52.68	25.6
10/14/2020	2:29:43 PM	52.59	24.8
		50.94647887	55.7778169

EO Gas Release Information

Tote #	Gas Pre-Weight (g)	Gas Post-Weight (g)	Net Weight Release (g)	Tote #	Gas Pre-Weight (g)	Gas Post-Weight (g)	Net Weight Release (g)
Tote #1	104.98	76.09	28.89	Tote #19			
Tote #2	106.81	75.45	31.36	Tote #20			
Tote #3	101.68	76.00	25.68	Tote #21			
Tote #4	102.59	75.79	26.80	Tote #22			
Tote #5	102.33	76.42	25.91	Tote #23			
Tote #6	102.88	75.85	27.03	Tote #24			
Tote #7	103.32	75.97	27.35	Tote #25			
Tote #8	104.00	75.93	28.07	Tote #26			
Tote #9	105.40	75.91	29.49	Tote #27			
Tote #10	106.10	75.43	30.67	Tote #28			
Tote #11	105.89	75.68	30.21	Tote #29			
Tote #12	104.59	75.78	28.81	Tote #30			
Tote #13	105.16	76.19	28.97	Tote #31			
Tote #14	104.86	76.12	28.74	Tote #32			
Tote #15	102.99	75.98	27.01	Tote #33			
Tote #16	102.46	76.03	26.43	Tote #34			
Tote #17	104.57	75.91	28.66	Tote #35			
Tote #18	104.28	75.83	28.45	Tote #36			

N/A
ER 13OCT2020

All EO Gas Post Weights Pass (22 g – 45 g) ☒ Yes ☐ No (If no, notify Supervisor)

Operator Signature/Date: Marcia Beltrán 14OCT2020 Checked By/Date: Bruce Stiller 16 OCT 2020

pre E

001:N	+	104.988	9
002:N	+	106.819	9
003:N	+	101.682	9
004:N	+	102.596	9
005:N	+	102.331	9
006:N	+	102.881	9
007:N	+	103.321	9
008:N	+	104.007	9
009:N	+	105.404	9
010:N	+	106.107	9
011:N	+	105.894	9
012:N	+	104.593	9
013:N	+	105.164	9
014:N	+	104.864	9
015:N	+	102.995	9
016:N	+	102.462	9
017:N	+	104.575	9
018:N	+	104.283	9

post E

021:N	+	76.097	9
022:N	+	75.458	9
023:N	+	76.000	9
024:N	+	75.797	9
025:N	+	76.426	9
026:N	+	75.855	9
027:N	+	75.975	9
028:N	+	75.933	9
029:N	+	75.910	9
030:N	+	75.631	9
031:N	+	75.688	9
032:N	+	75.780	9
033:N	+	76.191	9
034:N	+	76.126	9
035:N	+	75.989	9
036:N	+	76.036	9
037:N	+	75.911	9
038:N	+	75.839	9

EO Sterilization Router
2.0

Tote #	Dosimeter Travel Length		Tote #	Dosimeter Travel Length	
Tote #1	▲	(Pass)	Tote # 19	▲	Pass Fail
Tote # 2	▲	(Pass)	Tote # 20	▲	Pass Fail
Tote # 3	▲	(Pass)	Tote # 21	▲	Pass Fail
Tote # 4	▲	(Pass)	Tote # 22	▲	Pass Fail
Tote # 5	▲	(Pass)	Tote # 23	▲	Pass Fail
Tote # 6	▲	(Pass)	Tote # 24	▲	Pass Fail
Tote # 7	▲	(Pass)	Tote # 25	▲	Pass Fail
Tote # 8	▲	(Pass)	Tote # 26	▲	Pass Fail
Tote # 9	▲	(Pass)	Tote # 27	▲	Pass Fail
Tote # 10	▲	(Pass)	Tote # 28	▲	Pass Fail
Tote # 11	▲	(Pass)	Tote # 29	▲	Pass Fail
Tote # 12	▲	(Pass)	Tote # 30	▲	Pass Fail
Tote # 13	▲	(Pass)	Tote # 31	▲	Pass Fail
Tote # 14	▲	(Pass)	Tote # 32	▲	Pass Fail
Tote # 15	▲	(Pass)	Tote # 33	▲	Pass Fail
Tote # 16	▲	(Pass)	Tote # 34	▲	Pass Fail
Tote # 17	▲	(Pass)	Tote # 35	▲	Pass Fail
Tote # 18	▲	(Pass)	Tote # 36	▲	Pass Fail
Operator Signature/Date	Maria Beltran 16 OCT 2020		Checked By/Date	Peter Stahlke 16 OCT 2020	