



ENVIRONMENTAL PROTECTION DIVISION

Richard E. Dunn, Director

EPD Director's Office

2 Martin Luther King, Jr. Drive
Suite 1456, East Tower
Atlanta, Georgia 30334
404-656-4713

June 7, 2021

MEMORANDUM

To: Karen Hays, Air Branch Chief *KH*

Through: James Boylan, Assistant Branch Chief

From: DeAnna Oser, Ambient Monitoring Program Manager

RE: **Revised:** General Coffee July 2, 2020 Sampling Results for Ethylene Oxide

A sample of Ethylene Oxide (EtO) was collected by the Ambient Monitoring Program at our General Coffee monitoring station on July 2, 2020. The sample was sent to Eastern Research Group (ERG) for analysis.

EPA has posted two memorandums discussing the ethylene oxide concentrations in the collection canisters and potential biases that may be caused by the lining of the canisters. These memorandums are posted on our website for reference.

We are updating the data results for the sample collected on July 2, 2020 based on additional information that has become available concerning the canister used for this sample. When this canister was used a second time in our study, the concentration was also significantly high as compared to the other samples collected that sample day. The code SC indicates that the invalidated due to sampler contamination.

Location	Date	Qualifier	EtO Concentration (ug/m ³)
General Coffee	07/02/2020		SC

If you have any questions, please let me know.

DGO/do