



UNITED STATES ENVIRONMENTAL PROTECTION AGENCY
 Region 4 Laboratory Services and Applied Science Division
 980 College Station Road, Athens, Georgia 30605-2700
 D.A.R.T. Id: 19-0372
 Project: 19-0372, GA EPD Can Check - Reported by Sallie Hale

July 16, 2019

4LSASD-LSB

MEMORANDUM

SUBJECT: FINAL Analytical Report
 Project: 19-0372, GA EPD Can Check

FROM: Sallie Hale
 OCS Analyst

THRU: Jeffrey Hendel, Chief
 LSB Organic Chemistry Section

TO: Stacie Masters

Attached are the final results for the analytical groups listed below. This report shall not be reproduced except in full without approval of the Region 4 laboratory. These analyses were performed in accordance with the Laboratory Services Branch's Laboratory Operations and Quality Assurance Manual (LSB LOQAM) found at www.epa.gov/region4/sesd/asbsop. Any unique project data quality objectives specified in writing by the data requestor have also been incorporated into the data unless otherwise noted in the Report Narrative. Chemistry data have been verified based on the LSB LOQAM specifications and have been qualified by this laboratory if the applicable quality control criteria were not met. Verification is defined in Chapter 5 of the LSB LOQAM. For a listing of specific data qualifiers and explanations, please refer to the Data Qualifier Definitions included in this report. The reported results are accurate within the limits of the method(s) and are representative only of the samples as received by the laboratory.

Analyses Included in this report:	Method Used:	Accreditations:
Volatile Organics (VOA)		
Volatile organic compounds	EPA TO-15 (Air)	ISO



UNITED STATES ENVIRONMENTAL PROTECTION AGENCY
Region 4 Laboratory Services and Applied Science Division
980 College Station Road, Athens, Georgia 30605-2700
D.A.R.T. Id: 19-0372
Project: 19-0372, GA EPD Can Check - Reported by Sallie Hale

Report Narrative for Work Order: E192601 Analysis: VOA

7/16/19 VOC SJH: This sample was analyzed against two curves made with standards from different sources. The ethylene oxide curve check standard failed high evaluated against the 6/25 curve and low against the 6/30 curve. We are officially reporting the higher sample value (6/25), but with J and QC-3 qualifiers because "analyte calibration criteria not met" was the situation on both days. (In case you are interested, the lower value (6/30) was 0.23 ug/m3.) All other QC samples easily passed on both days.

If you have any questions, call Kristin Trapp at 706-355-8868.

Sample Disposal Policy

Due to limited space for long term sample storage, LSB's policy is to dispose of samples on a periodic schedule. Air samples collected in summa canisters will be disposed of 30 days following the issuance of this report. All other sample media including original samples, sample extracts and or digestates will be disposed of, in accordance with applicable regulations, 60 days from the date of this report.

This sample disposal policy does not apply to criminal samples which are held until the laboratory is notified by the criminal investigators that case development and litigation are complete.

These samples may be held in the laboratory's custody for a longer period of time. If samples require storage beyond the 60-day period, please contact the Sample Control Coordinator by e-mail at R4SampleCustody@epa.gov.



UNITED STATES ENVIRONMENTAL PROTECTION AGENCY
Region 4 Laboratory Services and Applied Science Division
980 College Station Road, Athens, Georgia 30605-2700
D.A.R.T. Id: 19-0372
Project: 19-0372, GA EPD Can Check - Reported by Sallie Hale

SAMPLES INCLUDED IN THIS REPORT

Project: 19-0372, GA EPD Can Check

Sample ID	Laboratory ID	Matrix	Date Collected	Date Received
Silonite 6 Liter Canister # 32492	E192601-01	Air	6/14/19 00:00	6/21/19 10:15



UNITED STATES ENVIRONMENTAL PROTECTION AGENCY
Region 4 Laboratory Services and Applied Science Division
980 College Station Road, Athens, Georgia 30605-2700
D.A.R.T. Id: 19-0372
Project: 19-0372, GA EPD Can Check - Reported by Sallie Hale

DATA QUALIFIER DEFINITIONS

J The identification of the analyte is acceptable; the reported value is an estimate.
QC-3 Analyte calibration criteria not met

ACRONYMS AND ABBREVIATIONS

CAS Chemical Abstracts Service

Note: Analytes with no known CAS identifiers have been assigned codes beginning with "E", the EPA ID as assigned by the EPA Substance Registry System (www.epa.gov/srs), or beginning with "R4-", a unique identifier assigned by the EPA Region 4 laboratory.

MDL Method Detection Limit - The minimum concentration of a substance (an analyte) that can be measured and reported with a 99% confidence that the analyte concentration is greater than zero.

MRL Minimum Reporting Limit - Analyte concentration that corresponds to the lowest demonstrated level of acceptable quantitation. The MRL is sample-specific and accounts for preparation weights and volumes, dilutions, and moisture content of soil/sediments.

TIC Tentatively Identified Compound - An analyte identified based on a match with the instrument software's mass spectral library. A calibration standard has not been analyzed to confirm the compound's identification or the estimated concentration reported.

ACCREDITATIONS:

ISO ASB is accredited by ISO/IEC 17025, including an amplification for forensic accreditation through ANSI-ASQ National Accreditation Board.

Refer to the certificate and scope of accreditation AT-1644 at:
<http://www.epa.gov/aboutepa/about-region-4s-science-and-ecosystem-support-division-sesd>

NR The EPA Region 4 Laboratory has not requested accreditation for this test.



UNITED STATES ENVIRONMENTAL PROTECTION AGENCY
 Region 4 Laboratory Services and Applied Science Division
 980 College Station Road, Athens, Georgia 30605-2700
 D.A.R.T. Id: 19-0372
 Project: 19-0372, GA EPD Can Check - Reported by Sallie Hale

Volatile Organics

Project: 19-0372, GA EPD Can Check

Sample ID: Silonite 6 Liter Canister # 32492

Lab ID: E192601-01

Station ID:

Matrix: Air

Date Collected: 6/14/19 0:00

<i>CAS Number</i>	<i>Analyte</i>	<i>Results</i>	<i>Qualifiers</i>	<i>Units</i>	<i>MRL</i>	<i>Prepared</i>	<i>Analyzed</i>	<i>Method</i>
75-21-8	Ethylene oxide	0.309	J, QC-3	ug/m3	0.0390	6/14/19 0:00	6/26/19 6:57	EPA TO-15



UNITED STATES ENVIRONMENTAL PROTECTION AGENCY
 Region 4 Laboratory Services and Applied Science Division
 980 College Station Road, Athens, Georgia 30605-2700
 D.A.R.T. Id: 19-0372
 Project: 19-0372, GA EPD Can Check - Reported by Sallie Hale

Volatile Organics (VOA) - Quality Control
US-EPA, Region 4, LSASD

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
Batch 1906082 - V TO-15 Air Canister										
Blank (1906082-BLK1)					Prepared: 06/13/19 Analyzed: 06/26/19					
EPA TO-15										
Ethylene oxide	U	0.0390	ug/m3							QC-6, U
LCS (1906082-BS1)					Prepared: 06/20/19 Analyzed: 06/26/19					
EPA TO-15										
Ethylene oxide	2.0919		ppbv	2.1560		97.0	70-130			QC-6
LCS Dup (1906082-BSD1)					Prepared: 06/20/19 Analyzed: 06/26/19					
EPA TO-15										
Ethylene oxide	2.0986		ppbv	2.1560		97.3	70-130	0.317	25	QC-6
Duplicate (1906082-DUP1)		Source: E192601-01			Prepared: 06/14/19 Analyzed: 06/26/19					
EPA TO-15										
Ethylene oxide	0.30453	0.0390	ug/m3		0.30857			1.32	25	J, QC-3
MRL Verification (1906082-PS1)					Prepared: 06/20/19 Analyzed: 06/26/19					
EPA TO-15										
Ethylene oxide	0.030360		ppbv	0.021560		141	50-150			MRL-5, QC-6



UNITED STATES ENVIRONMENTAL PROTECTION AGENCY

Region 4 Laboratory Services and Applied Science Division

980 College Station Road, Athens, Georgia 30605-2700

D.A.R.T. Id: 19-0372

Project: 19-0372, GA EPD Can Check - Reported by Sallie Hale

Notes and Definitions for QC Samples

- U The analyte was not detected at or above the reporting limit.
- J The identification of the analyte is acceptable; the reported value is an estimate.
- MRL-5 MRL verification for Air matrix
- QC-3 Analyte calibration criteria not met
- QC-6 Calibration check standard greater than method control limits.