

Georgia Department of Natural Resources

Environmental Protection Division · Air Protection Branch

4244 International Parkway · Suite 120 · Atlanta · Georgia 30354

Telephone: 404/363-7000 · Fax: 404/363-7100

Judson H. Turner, Director

JUL 18 2016

MEMORANDUM:

TO: Michael Odom
THROUGH: Ross Winne, Marcus Cureton
FROM: Anna Gray
SUBJECT: SOURCE TEST REPORT REVIEW

The following test has been reviewed and was conducted in an acceptable fashion for the purpose intended.

COMPANY NAME	Sterigenics U.S. LLC
COMPANY LOCATION	Atlanta, GA
SOURCE TESTED	10-Sterilizer Chambers/Ceilcote Packed Tower Scrubber
POLLUTANT DETERMINED	Ethylene Oxide
REPORT REVIEWED BY	Anna Gray
TEST WITNESSED BY	Ross Winne
DATE(S) OF TEST	March 17, 2016 to March 18, 2016
DATE RECEIVED BY APB	May 24, 2016
APPLICABLE REGULATION	Permit No. 7389-067-0093-S-05-3, Condition 2.3

MEMORANDUM

The reduction efficiency test of the Ceilcote packed tower scrubber emission control system was conducted in accordance with USEPA CFR 40, Part 63.365. At least one of the test runs was performed using emissions from the new 30-pallet sterilizer (Chamber#11). The other two chambers were: #3 and #8.

Run #1 Inlet EtO= 44.9 lbs
Outlet EtO= 0.0000065 lbs
EtO Control Efficiency= 99.9999%
Minutes/cycle= 24

Run #2 Inlet EtO=61.9 lbs
Outlet EtO= 0.0000141 lbs
EtO Control Efficiency= 99.9999%
Minutes/cycle= 23

Run #3 Inlet EtO= 76.1 lbs
Outlet EtO= 0.0000076 lbs
EtO Control Efficiency= 99.9999%
Minutes/cycle= 20

Average EtO Control Efficiency: 99.9999%
Required EtO Control Efficiency: 99%

The amount of ethylene oxide used during each sterilization cycle was calculated by the Gas law and the conditions at the beginning and the end of every exhaust phase. All exhaust phase testing was conducted during normal process load conditions, but with an empty sterilization chamber to facilitate inlet mass calculation and the performance of multiple test runs.

Parameters during test:

Ceilcote scrubber readings: pH= 1.2 ; Storage tank level= 186 inches ;Glycol concentration=36.3%

AAT scrubber readings: pH=0.9; Tank level= 105 inches; Glycol concentration=36.8%

**REPORT OF
AIR POLLUTION SOURCE TESTING
OF AN ETHYLENE OXIDE EMISSION-CONTROL SYSTEM
OPERATED BY STERIGENICS, INC.
IN ATLANTA, GEORGIA
ON MARCH 17-18, 2016**

RECEIVED

MAY 25 2016

ISMP

Submitted to:

**GEORGIA DEPARTMENT OF NATURAL RESOURCES
Environmental Protection Division
4244 International Parkway, Suite 120
Atlanta, Georgia 30354**

Submitted by:

**STERIGENICS U.S., LLC.
2971 Olympic Industrial Boulevard
Atlanta, Georgia 30339**

GDNR Permit Number 7839-067-0093-S-05-3

Prepared by:

**ECSI, INC.
PO Box 848
San Clemente, California 92674-0848**

April 21, 2016

ECSi

CONTACT SUMMARY

CLIENT

Ms. Sue Reinhardt
EHS Manager
STERIGENICS U.S., LLC.
2015 Spring Road, Suite 650
Oak Brook, Illinois 60523

Phone: (630)928-1768
FAX: (630)928-1701
Email: sreinhardt@sterigenics.com

FACILITY

Mr. Daryl Mosby
General Manager
STERIGENICS U.S., LLC.
2973 Olympic Industrial Boulevard
Atlanta, Georgia 30080

Phone: (404)355-4485
FAX: (404)355-4852
Email: dmosby@sterigenics.com

TEST DATE

March 17-18, 2016

REGULATORY AGENCY

Mr. Don Holder
Principal Environmental Engineer
Stationary Source Compliance Program
GEORGIA DEPARTMENT OF NATURAL RESOURCES
Environmental Protection Division, Air Protection Branch
4244 International Parkway, Suite 120
Atlanta, Georgia 30354

Phone: (404)363-7000
FAX: (404)363-7100
Email: don_holder@dnr.state.ga.us

TESTING CONTRACTOR

Daniel P. Kremer
Project Manager
ECSi, Inc.
PO Box 848
San Clemente, California 92674-0848

Phone: (949)400-9145
FAX: (949)281-2169
email: dankremer@ecsi1.com

TABLE OF CONTENTS

	<u>PAGE NO.</u>
CONTACT SUMMARY	i
TABLE OF CONTENTS	ii
LIST OF TABLES	iii
LIST OF APPENDICES	iv
1.0 INTRODUCTION	1
2.0 EQUIPMENT	2
3.0 TESTING	3
4.0 RULE/COMPLIANCE REQUIREMENTS	4
5.0 TEST METHOD REFERENCE	5
5.1 Introduction	5
5.2 Volumetric Flow Measurement	6
5.3 Mass-Emissions Measurement	6
5.4 Sample Transport	7
5.5 GC Injection	7
5.6 GC Conditions	7
5.7 Calibration Standards	7
5.8 Sampling Duration	8
5.9 Control Efficiency/Mass-Emissions Calculations	8
6.0 TEST SCENARIO	10
7.0 QA/QC	11
7.1 Field Testing Quality Assurance	11
7.2 Calibration Procedures	11
8.0 TEST RESULTS	12
TABLES	13
APPENDICES	15

LIST OF TABLES

<u>TABLE</u>	<u>DESCRIPTION</u>	<u>PAGE NO.</u>
1	Ethylene Oxide Control Efficiency	14

LIST OF APPENDICES

<u>APPENDIX</u>	<u>DESCRIPTION</u>	<u>PAGE NO.</u>
A	Calibration Data	A-1
B	Run#1 Chromatograms	B-1
C	Run#2 Chromatograms	C-1
D	Run#3 Chromatograms	D-1
E	Field Data and Calculation Worksheets	E-1
F	Gas Certifications	F-1
G	Process/Parametric Monitoring Data	G-1

1.0 INTRODUCTION

On Thursday and Friday, March 17-18, 2016, ECSi, Inc. performed air pollution source testing of an ethylene oxide (EtO) emission-control device operated by Sterigenics, Inc. in Atlanta, Georgia. The control device tested was a Ceilcote packed tower scrubber emission-control system, which is currently used to control emissions from ten EtO sterilizer vacuum pumps. The purpose of the testing program was to demonstrate continued compliance with the conditions established in the Air Quality Permit granted to Sterigenics by the Georgia Department of Natural Resources, Environmental Protection Division (GDNR).

2.0 EQUIPMENT

The EtO gas-sterilization system is comprised of ten commercial sterilizers, which are discharged through liquid-ring vacuum pumps to a Ceilcote packed tower scrubber emission-control system, ten sterilizer exhaust vents (backvents) and one aeration room, which are discharged to a two-stage Advanced Air Technologies (AAT) Safe Cell emission-control system. As an alternative emission-control scenario, the facility also has the capability to discharge the sterilization chamber vacuum pumps to the AAT Safe Cell system. The gas-sterilization and emission-control equipment consist of the following:

- Ten Gas Sterilizers, two 5-pallet, two 6-pallet, four 13-pallet, and two 30-pallet capacity, each comprised of a steam-heated sterilization chamber, a recirculating vacuum pump chamber evacuation system, a backvent valve, and a fugitive emissions exhaust hood;
- One aeration room (AR-1), 152,400 cubic feet capacity, comprised of a heated aeration chamber and a chamber exhaust system.

Sterilizer vacuum pump emissions are controlled by:

- One Ceilcote packed tower chemical scrubber, equipped with: a reaction/interface column, 27' 4" high, 42" in diameter, with a 20' bed of #1 Tellerette packing; a 115 GPM scrubber fluid recirculation system; and two 28,000 gallon reaction/storage tanks.

Sterilizer backvent and aeration emissions are controlled by:

- One two-stage Advanced Air Technologies Safe Cell emission-control system, comprised of a packed-tower chemical scrubber (SC1), equipped with a packed reaction/interface column, a scrubber fluid recirculation system, and a scrubber fluid reaction/storage tank, and a dry bed reactor/scrubber (SC2), comprised of a bank of solid-bed reaction vessels, connected in parallel, installed downstream of SC1 and upstream of a dedicated blower exhaust system.

3.0 TESTING

EtO source testing was conducted in accordance with the procedures outlined in USEPA CFR40, Part 63.365. EtO emissions monitoring was conducted simultaneously at the inlet and outlet of the packed tower scrubber during the first chamber evacuation of the sterilizer exhaust phase of one of the ten currently operating sterilizers, and at least one of the test runs was performed using emissions from the new 30-pallet sterilizer. A total of three exhaust-phase test runs were performed.

During the first chamber evacuation of the exhaust phase, EtO emissions to the inlet of the packed tower scrubber were determined using the Ideal Gas Law and the chamber conditions at the beginning and at the end of the first chamber evacuation. During the first chamber evacuation of the exhaust phase, EtO emissions from the outlet of the packed tower scrubber were determined using direct source sample injection into the GC.

All exhaust phase testing was conducted during normal process load conditions, but with an empty sterilization chamber to facilitate inlet mass calculation and the performance of multiple test runs. The testing program was conducted in accordance with the procedures outlined in the following sections.

4.0 RULE/COMPLIANCE REQUIREMENTS

The EtO gas-sterilization system at Sterigenics was tested to demonstrate compliance with the EPA requirements, as specified in the GDNR Air Quality Permit. The following requirements must be met:

- The sterilizer exhaust phase (post exposure vacuum pulses) emissions must be vented to control equipment with an EtO emission-reduction efficiency of at least 99 % by weight.

Testing is required to demonstrate compliance with these requirements. Source testing of the packed tower scrubber emission-control device is required initially, and may be required periodically thereafter.

5.0 TEST METHOD REFERENCE

5.1 INTRODUCTION

EtO source testing was conducted in accordance with the procedures outlined in USEPA CFR40, Part 63.365. EtO emissions monitoring was conducted simultaneously at the inlet and outlet of the packed tower scrubber during the first chamber evacuation of the sterilizer exhaust phase of one of the ten currently operating sterilizers, and at least one of the test runs was performed using emissions from the new 30-pallet sterilizer. A total of three exhaust-phase test runs were performed.

Exhaust phase testing with one sterilizer discharging to the scrubber at a time represents worst-case conditions for demonstration of control efficiency compliance. At this lower inlet loading, the scrubber must perform at its maximum efficiency to achieve outlet EtO concentrations low enough to demonstrate compliance. One of the larger sterilizers was tested for each of the test runs to provide a realistic operational scenario.

During the first chamber evacuation of the exhaust phase, EtO emissions to the inlet of the packed tower scrubber were determined using the Ideal Gas Law and the chamber conditions at the beginning and at the end of the first chamber evacuation. During the first chamber evacuation of the exhaust phase, EtO emissions from the outlet of the packed tower scrubber were determined using direct source sample injection into the GC.

All exhaust phase testing was conducted during normal process load conditions, but with an empty sterilization chamber to facilitate inlet mass calculation and the performance of multiple test runs. The testing program was conducted in accordance with the procedures outlined in the following sections.

Operation and documentation of process conditions was performed by personnel from Sterigenics, Inc. using existing monitoring instruments installed by the manufacturer on the equipment to be tested. In accordance with the procedures established in USEPA CFR40, Part 63, Subpart O, scrubber liquor level was recorded. This parametric monitoring data is attached as Appendix G.

VOLUMETRIC FLOW MEASUREMENT

Exhaust gas flow at the outlet of the scrubber was determined by 40 CFR 60, Appendix A, Method 2, using an s-type pitot tube and an inclined-oil manometer. Sampling ports were located in accordance with 40 CFR 60, Appendix A, Method 1. The test ports were located far enough from any flow disturbances to permit accurate flow measurement.

Temperature measurements were obtained from a type K thermocouple and thermometer attached to the sampling probe. Exhaust gas composition was assumed to be air and small amounts of water vapor. Water vapor was negligible and, based on previous test data, a value of 2 percent was used for flow calculations.

CONTROL EFFICIENCY AND MASS EMISSIONS MEASUREMENT

During the first chamber evacuation of the sterilizer exhaust phase, the mass emissions of EtO vented to the inlet of the scrubber were determined using the procedures outlined in CFR40, Part 63.365. This method allows the determination of the mass of EtO vented to the inlet of the scrubber through calculations based on the Ideal Gas Law and using the conditions (pressure, temperature, volume) of the sterilization chamber immediately after it has been charged with sterilant gas, and upon conclusion of the first chamber evacuation of the exhaust phase.

The mass of EtO vented to the inlet of the scrubber during the first chamber evacuation of the exhaust phase was determined by calculating the mass of EtO present in the chamber after the first chamber evacuation and subtracting it from the mass of EtO present in the chamber after it had been charged with sterilant gas. The mass of EtO present in the chamber was calculated using Equation 1, shown below in Section 5.9.

During the first chamber evacuation of the sterilizer exhaust phase, EtO emissions from the outlet were determined using direct source sample injection into the GC. The mass of EtO emitted from the outlet was determined using Equation 2, shown below in Section 5.9. Mass-mass control-efficiency of EtO during the sterilizer exhaust phase was calculated by comparing the mass of EtO vented to the system inlet to the mass of EtO vented from the system outlet.

During the sterilization chamber exhaust phase, vented gas was analyzed by an SRI, Model 8610, portable gas chromatograph (GC), equipped with the following: dual, heated sample loops and injectors; dual

columns; and dual detectors. A photoionization detector (PID) was used to quantify low-level EtO emissions at the packed tower scrubber outlet.

5.4 SAMPLE TRANSPORT

Source gas was pumped to the GC at approximately 500-1000 cubic centimeters per minute (cc/min) from the sampling ports through two lengths of Teflon[®] sample line, each with a nominal volume of approximately 75 cubic centimeters (cc) and an outer diameter of 0.25 inch. At the outlet of the scrubber the sampling ports were located in the exhaust stack.

5.5 GC INJECTION

Source-gas samples were then injected into the GC which was equipped with two heated sampling loops, each containing a volume of approximately 2cc and maintained at 100 degrees Celsius (C). Injections occurred at approximately one-minute intervals during the sterilization chamber exhaust phase. Helium was the carrier gas for the PID.

5.6 GC CONDITIONS

The packed columns for the GC were both operated at 80 degrees C. The columns were stainless steel, 6 feet long, 0.125 inch outer diameter, packed with 1 percent SP-1000 on 60/80 mesh Carbopack B.

Any unused sample gas was vented from the GC system back to the inlet of the scrubber.

5.7 CALIBRATION STANDARDS

The PID was calibrated for low-range ppmv level analyses using gas proportions similar to the following:

- 1) 100 ppmv EtO, balance nitrogen
- 2) 50 ppmv EtO, balance nitrogen (audit gas)
- 3) 10 ppmv EtO, balance nitrogen
- 4) 1 ppmv EtO, balance nitrogen

Each of these calibration standards was in a separate, certified manufacturer's cylinder. Copies of the calibration gas laboratory certificates are attached as Appendix F.

5.8 SAMPLING DURATION

Exhaust phase EtO measurements were taken for the entire duration of the first chamber evacuation, which was approximately 20-30 minutes. This encompassed a total sampling duration of approximately 20-30 minutes for each exhaust phase test run.

5.9 CONTROL-EFFICIENCY/MASS-EMISSIONS CALCULATIONS

The following equation was used to calculate mass of EtO discharged to the inlet of the emission-control system during the first chamber evacuation of the sterilizer exhaust phase:

EQUATION 1:

$$W_c = W_{ci} - W_{cf}$$

Where:

W_c = Weight of EtO discharged from the sterilization chamber to the emission-control system during the first chamber evacuation, pounds

$$W_{ci} = (mw)(p)(P)(V)/(R)(T)$$

(and W_{cf})

Where:

W_{ci} = Weight of EtO present in the sterilization chamber before the first chamber evacuation, pounds

W_{cf} = Weight of EtO present in the sterilization chamber after the first chamber evacuation, pounds

MW = Molecular weight of EtO, 44.05 lb/mol

p = Percent of EtO in chamber

$$= W_s/W_i$$

Where:

W_s = Scale-measured weight of EtO charged into sterilization chamber

W_i = Calculated weight of EtO charged into sterilization chamber (@ 100%)

P = Sterilization chamber pressure (after charging/at the end of the 1st evac), psia

- V = Sterilization chamber volume, ft³
 R = Gas constant, 10.73 psia·ft³/mol·°R
 T = Sterilization chamber temperature (after charging/at the end of the 1st evac), °R

Note: Standard conditions are 68°F and 1 atm.

Mass emissions of EtO during the exhaust phase were calculated using the following equation:

EQUATION 2:

$$\text{MassRate} = (\text{VolFlow})(\text{MolWt})(\text{ppmv EtO}/10^6)/(\text{MolVol})$$

Where:

MassRate = EtO mass flow rate, pounds per minute

VolFlow = Corrected volumetric flow rate, standard cubic feet per minute at 68 degrees F

MolWt = 44.05 pounds EtO per pound mole

ppmv EtO = EtO concentration, parts per million by volume

10⁶ = Conversion factor, ppmv per "cubic foot per cubic foot"

MolVol = 385.32 cubic feet per pound mole at one atmosphere and 68 degrees F

Results of the control-efficiency testing are presented in Section 8.0 and in Table 1.

6.0 TEST SCENARIO

During exhaust phase testing, each sterilizer was tested during normal process load conditions, but with an empty sterilization chamber to facilitate the performance of multiple test runs. A total of three exhaust-phase test runs were performed to verify the performance of the emission-control device. Testing was conducted with an effort to offer minimal disruption to the Sterigenics production schedule. The testing schedule was as follows:

- 1) Testing equipment was set up and calibrated.
- 2) An empty-chamber cycle was started in one of the larger sterilizers. This sterilizer was isolated for test use and designated as a test chamber.
- 3) Exhaust Phase Test Run #1 was conducted. Sampling was performed at outlet of the scrubber during the first chamber evacuation of the test chamber. During the performance of the test, only the sterilizer used for the test was allowed to discharge to the Ceilcote scrubber.
- 4) An empty-chamber cycle was started in another of the larger sterilizers. This sterilizer was isolated for test use and designated as a test chamber.
- 5) Exhaust Phase Test Run #2 was conducted. Sampling was performed at outlet of the scrubber during the first chamber evacuation of the test chamber. During the performance of the test, only the sterilizer used for the test was allowed to discharge to the Ceilcote scrubber.
- 6) An empty-chamber cycle was started in the new 30-pallet sterilizer. This sterilizer was isolated for test use and designated as a test chamber.
- 7) Exhaust Phase Test Run #3 was conducted. Sampling was performed at outlet of the scrubber during the first chamber evacuation of the test chamber. During the performance of the test, only the sterilizer used for the test was allowed to discharge to the Ceilcote scrubber.
- 8) Post calibration check was performed, testing equipment was packed.

7.0 QA/QC

7.1 FIELD TESTING QUALITY ASSURANCE

At the beginning of the test, the sampling system was leak checked at a vacuum of 15 inches of mercury. The sampling system was considered leak free when the flow indicated by the rotameters fell to zero.

At the beginning of the test, a system blank was analyzed to ensure that the sampling system was free of EtO. Ambient air was introduced at the end of the heated sampling line and drawn through the sampling system line to the GC for analysis. The resulting chromatogram also provided a background level for non-EtO components (i.e. ambient air, carbon dioxide, water vapor) which are present in the source gas stream due to the ambient dilution air which is drawn into the emission-control device, and due to the destruction of EtO by the emission-control device which produces carbon dioxide and water vapor. This chromatogram, designated AMB, is included with the calibration data in Appendix A.

7.2 CALIBRATION PROCEDURES

The GC system was calibrated at the beginning and conclusion of each day's testing. Using the Peaksimple II analytical software, a point-to-point calibration curve was constructed for each detector. A gas cylinder of similar composition as the calibration gases, but certified by a separate supplier, was used to verify calibration gas composition and GC performance.

All calibration gases and support gases used were of the highest purity and quality available. A copy of the laboratory certification for each calibration gas is attached as Appendix F.

8.0 TEST RESULTS

The Ceilcote scrubber demonstrated an EtO control efficiency of 99.99998 percent. In accordance with EPA requirements, as specified in the GDNR Air Quality Permit, this control equipment must have an EtO control efficiency of 99 percent or more during the sterilizer exhaust phase (vacuum pump emissions). The emission-control device met this requirement.

The test results are summarized in Table 1. These tables include results for EtO control efficiency of the emission-control device. Chromatograms and chromatographic supporting data are attached as Appendices A through D. Copies of field data and calculation worksheets are attached as Appendix E.

TABLES

TABLE 1
ETHYLENE OXIDE CONTROL EFFICIENCY
OF A CEILCOTE PACKED TOWER SCRUBBER EMISSION CONTROL DEVICE
OPERATED BY STERIGENICS, INC.
IN ATLANTA, GEORGIA
ON MARCH 17-18, 2016


<u>Run #</u>	<u>Stack Flow (dscfm) (2)</u>	<u>Average Outlet Conc. (ppm) (1)</u>	<u>Outlet EtO Mass Flow (lbs/min) (3)</u>	<u>Minutes/ Cycle</u>	<u>Outlet EtO Mass Emissions (lbs)</u>	<u>Inlet EtO Mass Emissions (lbs)</u>	<u>EtO Control Efficiency (%)</u>
#1	235	0.0100	0.0000003	24	0.0000065	44.9	99.999986
#2	262	0.0214	0.0000006	23	0.0000141	61.9	99.999977
#3	332	0.0100	0.0000004	20	0.0000076	76.1	99.999990
Average EtO Control Efficiency:							99.999984
Required EtO Control Efficiency:							99

- Notes: (1) - PPM = parts per million by volume
(2) - DSCFM = dry standard cubic feet per minute
(3) - LBS/MIN = EtO emissions, pounds per minute

APPENDICES

APPENDIX A
Calibration Data


pk	Name	Start	End	Calibration	Int.Std	Units
	Dead Vol / Air	0.000	0.350		0.000	
	Ambient H2O	0.350	0.500		0.000	
	Ethylene Oxide	0.500	0.600	C:\peak359\1Ster	0.00016	ppm
	Acetaldehyde	0.600	0.800		0.000	
	CO2	0.800	1.000		0.000	



slope of curve: 0.20
 y-axis intercept: 0.00
 repeatability: 1.00
 number of levels: 4
 1/rel SD of CF's: 0.1/66.8
 0.2007X
 1.0000
 1 calibrated: Thu Mar 17 17:12:29 2016

Area/ht	Amount	CF	Current	Previous #1	Previous #2
0.000	0.000	0.000	0.000	N/A	N/A
0.230	1.100	0.209	0.230	N/A	N/A
2.010	10.100	0.199	2.010	N/A	N/A
19.400	100.000	0.194	19.400	N/A	N/A

Name	Start	End	Calibration	Int.Std	Units
Dead Vol / Air	0.000	0.350		0.000	
Ambient H2O	0.350	0.490		0.000	
Ethylene Oxide	0.490	0.600	C:\peak359\2Ster	0.00016	ppm
Acetaldehyde	0.600	0.800		0.000	
CO2	0.800	1.000		0.000	




slope of curve: 1.34
 y-axis intercept: 0.00
 r-squared: 1.00
 number of levels: 4
 (rel SD of CF's: 0.7/67.0
 1.3350X
 1.0000

calibrated: Thu Mar 17 17:11:52 2016


Area/ht.	Amount	CF	Current	Previous #1	Previous #2
0.000	0.000	0.000	0.000	N/A	N/A
1.560	1.100	1.418	1.560	N/A	N/A
13.300	10.100	1.317	13.300	N/A	N/A
127.000	100.000	1.270	127.000	N/A	N/A

Client: Sterigenics - Atlanta
 Client ID: PreCal
 Analysis date: 03/17/2016 16:29:46
 Method: Direct Injection
 Description: CHANNEL 1 - FID
 Column: 1% SP-1000, Carbopack B
 Carrier: HELIUM
 Temp. prog: eto-100.tem
 Components: eto1-100.cpt
 Data file: 1SterAti-2016-Amb.CHR (c:\peak359)
 Sample: Ambient Background
 Operator: D. Kremer

Client: Sterigenics - Atlanta
 Client ID: PreCal
 Analysis date: 03/17/2016 16:29:46
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carbopack B
 Carrier: HELIUM
 Temp. prog: eto-100.tem
 Components: eto2-100.cpt
 Data file: 2SterAti-2016-Amb.CHR (c:\peak359)
 Sample: Ambient Background
 Operator: D. Kremer




Component	Retention	Area	External Units
Dead Vol / Air	0.250	2.5560	0.0000
Ethylene Oxide	0.550	1.9905	9.7206 ppm
		4.5465	9.7206




Component	Retention	Area	External Units
Dead Vol / Air	0.150	17.9840	0.0000
Ambient H2O	0.400	13.9590	0.0000
Ethylene Oxide	0.550	10.4880	10.2891 ppm
Acetaldehyde	0.733	0.1590	0.0000
CO2	0.833	0.0790	0.0000
		42.6690	10.2891

Client: Sterigenics - Atlanta
 Client ID: PreCal
 Analysis date: 03/17/2016 16:36:23
 Method: Direct Injection
 Description: CHANNEL 1 - FID
 Column: 1% SP-1000, Carbopack B
 Carrier: HELIUM
 Temp. prog: eto-100.tem
 Components: eto1-100.cpt
 Data file: 1SterAti-2016-C01.CHR (c:\peak359)
 Sample: 1.10 ppm EtO std
 Operator: D. Kremer

Client: Sterigenics - Atlanta
 Client ID: PreCal
 Analysis date: 03/17/2016 16:36:23
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carbopack B
 Carrier: HELIUM
 Temp. prog: eto-100.tem
 Components: eto2-100.cpt
 Data file: 2SterAti-2016-C01.CHR (c:\peak359)
 Sample: 1.10 ppm EtO std
 Operator: D. Kremer




Component	Retention	Area	External Units
Dead Vol / Air	0.250	1.0360	0.0000
Ethylene Oxide	0.566	0.2290	0.0000 ppm
		1.2650	0.0000




Component	Retention	Area	External Units
Dead Vol / Air	0.050	5.3230	0.0000
Ethylene Oxide	0.516	1.5405	0.0000 ppm
Acetaldehyde	0.766	0.2910	0.0000
CO2	0.933	0.1620	0.0000
		7.3165	0.0000

Client: Sterigenics - Atlanta
 Client ID: PreCal
 Analysis date: 03/17/2016 16:42:02
 Method: Direct Injection
 Description: CHANNEL 1 - FID
 Column: 1% SP-1000, Carboxpack B
 Carrier: HELIUM
 Temp. prog: eto-100.tem
 Components: eto1-100.cpt
 Data file: 1SterAti-2016-C02.CHR (c:\peak359)
 Sample: 1.10 ppm EtO std
 Operator: D. Kremer

Client: Sterigenics - Atlanta
 Client ID: PreCal
 Analysis date: 03/17/2016 16:42:02
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carboxpack B
 Carrier: HELIUM
 Temp. prog: eto-100.tem
 Components: eto2-100.cpt
 Data file: 2SterAti-2016-C02.CHR (c:\peak359)
 Sample: 1.10 ppm EtO std
 Operator: D. Kremer




Component	Retention	Area	External Units
Dead Vol / Air	0.216	0.8510	0.0000
Ethylene Oxide	0.550	0.2150	0.0000 ppm
		1.0660	0.0000




Component	Retention	Area	External Units
Dead Vol / Air	0.050	3.3250	0.0000
Ambient H2O	0.466	1.1770	0.0000
Ethylene Oxide	0.500	1.5680	0.0000 ppm
Acetaldehyde	0.700	0.2875	0.0000
CO2	0.866	0.1555	0.0000
		6.5130	0.0000

Client: Sterigenics - Atlanta
 Client ID: PreCal
 Analysis date: 03/17/2016 16:48:30
 Method: Direct Injection
 Description: CHANNEL 1 - FID
 Column: 1% SP-1000, Carbopack B
 Carrier: HELIUM
 Temp. prog: eto-100.tem
 Components: eto1-100.cpt
 Data file: 1SterAti-2016-C03.CHR (c:\peak359)
 Sample: 10.1 ppm EtO std
 Operator: D. Kremer

Client: Sterigenics - Atlanta
 Client ID: PreCal
 Analysis date: 03/17/2016 16:48:30
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carbopack B
 Carrier: HELIUM
 Temp. prog: eto-100.tem
 Components: eto2-100.cpt
 Data file: 2SterAti-2016-C03.CHR (c:\peak359)
 Sample: 10.1 ppm EtO std
 Operator: D. Kremer





Component	Retention	Area	External Units
Dead Vol / Air	0.216	0.4950	0.0000
Ethylene Oxide	0.550	1.9860	0.0000 ppm
CO2		2.4810	0.0000



Component	Retention	Area	External Units
Dead Vol / Air	0.166	7.5970	0.0000
Ethylene Oxide	0.550	13.1100	0.0000 ppm
CO2	0.816	0.2135	0.0000
		20.9205	0.0000

Client: Sterigenics - Atlanta
 Client ID: PreCal
 Analysis date: 03/17/2016 16:50:18
 Method: Direct Injection
 Description: CHANNEL 1 - FID
 Column: 1% SP-1000, Carbopack B
 Carrier: HELIUM
 Temp. prog: eto-100.tem
 Components: eto1-100.cpt
 Data file: 1SterAtt-2016-C04.CHR (c:\peak359)
 Sample: 10.1 ppm EtO std
 Operator: D. Kremer

Client: Sterigenics - Atlanta
 Client ID: PreCal
 Analysis date: 03/17/2016 16:50:18
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carbopack B
 Carrier: HELIUM
 Temp. prog: eto-100.tem
 Components: eto2-100.cpt
 Data file: 2SterAtt-2016-C04.CHR (c:\peak359)
 Sample: 10.1 ppm EtO std
 Operator: D. Kremer




Component	Retention	Area	External Units
Dead Vol / Air	0.233	0.3865	0.0000
Ethylene Oxide	0.566	2.0200	0.0000 ppm
		2.4065	0.0000


Component	Retention	Area	External Units
Dead Vol / Air	0.183	3.2555	0.0000
Ambient H2O	0.433	1.1720	0.0000
Ethylene Oxide	0.550	13.3885	0.0000 ppm
Acetaldehyde	0.666	0.6740	0.0000
CO2	0.816	0.1585	0.0000
		18.6485	0.0000

Client: Sterigenics - Atlanta
 Client ID: PreCal
 Analysis date: 03/17/2016 16:55:19
 Method: Direct Injection
 Description: CHANNEL 1 - FID
 Column: 1% SP-1000, Carbopack B
 Carrier: HELIUM
 Temp. prog: eto-100.tem
 Components: eto1-100.cpt
 Data file: 1SterAti-2016-C05.CHR (c:\peak359)
 Sample: 100 ppm EtO std
 Operator: D. Kremer

Client: Sterigenics - Atlanta
 Client ID: PreCal
 Analysis date: 03/17/2016 16:55:19
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carbopack B
 Carrier: HELIUM
 Temp. prog: eto-100.tem
 Components: eto2-100.cpt
 Data file: 2SterAti-2016-C05.CHR (c:\peak359)
 Sample: 100 ppm EtO std
 Operator: D. Kremer




Component	Retention	Area	External Units
Dead Vol / Air	0.216	0.3200	0.0000
Ethylene Oxide	0.550	19.1230	0.0000 ppm
		19.4430	0.0000




Component	Retention	Area	External Units
Dead Vol / Air	0.100	3.0115	0.0000
Ethylene Oxide	0.533	125.7335	0.0000 ppm
Acetaldehyde	0.766	0.1620	0.0000
CO2	0.883	0.1930	0.0000
		129.1000	0.0000

Client: Sterigenics - Atlanta
 Client ID: PreCal
 Analysis date: 03/17/2016 16:57:53
 Method: Direct Injection
 Description: CHANNEL 1 - FID
 Column: 1% SP-1000, Carbopack B
 Carrier: HELIUM
 Temp. prog: eto-100.tem
 Components: eto1-100.cpt
 Data file: 1SterAtt-2016-C06.CHR (c:\peak359)
 Sample: 100 ppm EtO std
 Operator: D. Kremer

Client: Sterigenics - Atlanta
 Client ID: PreCal
 Analysis date: 03/17/2016 16:57:53
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carbopack B
 Carrier: HELIUM
 Temp. prog: eto-100.tem
 Components: eto2-100.cpt
 Data file: 2SterAtt-2016-C06.CHR (c:\peak359)
 Sample: 100 ppm EtO std
 Operator: D. Kremer




Component	Retention	Area	External Units
Dead Vol / Air	0.216	0.3735	0.0000
Ethylene Oxide	0.566	19.6340	0.0000 ppm
		20.0075	0.0000




Component	Retention	Area	External Units
Dead Vol / Air	0.166	3.9360	0.0000
Ambient H2O	0.400	1.9310	0.0000
Ethylene Oxide	0.550	128.3350	0.0000 ppm
Acetaldehyde	0.783	0.1810	0.0000
CO2	0.883	0.3060	0.0000
		134.6890	0.0000

Client: Sterigenics - Atlanta
 Client ID: PreCal
 Analysis date: 03/17/2016 17:10:35
 Method: Direct Injection
 Description: CHANNEL 1 - FID
 Column: 1% SP-1000, Carbopack B
 Carrier: HELIUM
 Temp. prog: eto-100.tem
 Components: eto1-100.cpt
 Data file: 1SterAtt-2016-C07.CHR (c:\peak359)
 Sample: 48.8 ppm EtO std
 Operator: D. Kremer

Client: Sterigenics - Atlanta
 Client ID: PreCal
 Analysis date: 03/17/2016 17:10:35
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carbopack B
 Carrier: HELIUM
 Temp. prog: eto-100.tem
 Components: eto2-100.cpt
 Data file: 2SterAtt-2016-C07.CHR (c:\peak359)
 Sample: 48.8 ppm EtO std
 Operator: D. Kremer




Component	Retention	Area	External Units
Ethylene Oxide	10.550	10.0020	49.8355 ppm
		10.0020	49.8355




Component	Retention	Area	External Units
Dead Vol / Air	0.066	2.9000	0.0000
Ethylene Oxide	0.533	64.6810	48.4500 ppm
Acetaldehyde	0.716	0.1790	0.0000
unknown	0.800	0.1840	0.0000
CO2	0.883	0.1290	0.0000
		67.8890	48.4500

Client: Sterigenics - Atlanta
 Client ID: PostCal
 Analysis date: 03/18/2016 12:36:57
 Method: Direct Injection
 Description: CHANNEL 1 - FID
 Column: 1% SP-1000, Carbopack B
 Carrier: HELIUM
 Temp. prog: eto-100.tem
 Components: eto1-100.cpt
 Data file: 1SterAti-2016-C08.CHR (c:\peak359)
 Sample: 48.8 ppm EtO std
 Operator: D. Kremer

Client: Sterigenics - Atlanta
 Client ID: PostCal
 Analysis date: 03/18/2016 12:36:57
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carbopack B
 Carrier: HELIUM
 Temp. prog: eto-100.tem
 Components: eto2-100.cpt
 Data file: 2SterAti-2016-C08.CHR (c:\peak359)
 Sample: 48.8 ppm EtO std
 Operator: D. Kremer




Component	Retention	Area	External Units
Dead Vol / Air	0.233	0.3180	0.0000
Ethylene Oxide	0.566	9.8670	49.1629 ppm
		10.1850	49.1629



Component	Retention	Area	External Units
Dead Vol / Air	0.016	2.1460	0.0000
Ethylene Oxide	0.550	64.8120	48.5482 ppm
Acetaldehyde	0.766	0.1670	0.0000
CO2	0.883	0.1620	0.0000
		67.2870	48.5482


APPENDIX B
Run#1 Chromatograms

Client: Sterigenics - Atlanta
 Client ID: Run#1Exh
 Analysis date: 03/17/2016 17:40:36
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carbopack B
 Carrier: HELIUM
 Temp. prog: eto-100.tem
 Components: eto2-100.cpt
 Data file: 2SterAtl-2016-1E01.CHR (c:\peak359)
 Sample: Ceicote Scrubber Outlet
 Operator: D. Kremer




Component	Retention	Area	External Units
ad Vol / Air	0.033	5.0240	0.0000
ibient H2O	0.400	23.7900	0.0000
aldehyde	0.600	20.5880	0.0000
		49.4020	0.0000

Client: Sterigenics - Atlanta
 Client ID: Run#1Exh
 Analysis date: 03/17/2016 17:42:26
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carbopack B
 Carrier: HELIUM
 Temp. prog: eto-100.tem
 Components: eto2-100.cpt
 Data file: 2SterAti-2016-1E02.CHR (c:\peak359)
 Sample: Ceilcote Scrubber Outlet
 Operator: D. Kremer




Component	Retention	Area	External Units
ad Vol / Air	0.283	0.1320	0.0000
lient H2O	0.400	22.9070	0.0000
etaldehyde	0.600	35.6490	0.0000
		58.6880	0.0000

Client: Sterigenics - Atlanta
 Client ID: Run#1Exh
 Analysis date: 03/17/2016 17:43:40
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carbopack B
 Carrier: HELIUM
 Temp. prog: eto-100.tem
 Components: eto2-100.cpt
 Data file: 2SterAtt-2016-1E03.CHR (c:\peak359)
 Sample: Ceilcote Scrubber Outlet
 Operator: D. Kremer




Component	Retention	Area	External Units
Water	0.400	21.9240	0.0000
Formaldehyde	0.600	4.8350	0.0000
O2	0.833	227.6240	0.0000
		254.3830	0.0000

Client: Sterigenics - Atlanta
 Client ID: Run#1Exh
 Analysis date: 03/17/2016 17:45:25
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carbopack B
 Carrier: HELIUM
 Temp. prog: eto-100.tem
 Components: eto2-100.cpt
 Data file: 2SterAtl-2016-1E04.CHR (c:\peak359)
 Sample: Ceilcote Scrubber Outlet
 Operator: D. Kremer




Component	Retention	Area	External Units
Vol / Air	0.050	27.8150	0.0000
ent H2O	0.400	29.9365	0.0000
aldehyde	0.600	5.1715	0.0000
	0.916	105.6440	0.0000
		168.5670	0.0000

Client: Sterigenics - Atlanta
 Client ID: Run#1Exh
 Analysis date: 03/17/2016 17:48:12
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carbopack B
 Carrier: HELIUM
 Temp: prog: eto-100.tem
 Components: eto2-100.cpt
 Data file: 2SterAtt-2016-1E05.CHR (c:\peak359)
 Sample: Ceilcote Scrubber Outlet
 Operator: D. Kremer




Component	Retention	Area	External Units
ad Vol / Air	0.150	84.9220	0.0000
lient H2O	0.400	109.3620	0.0000
		194.2840	0.0000

Client: Sterigenics - Atlanta
 Client ID: Run#1Exh
 Analysis date: 03/17/2016 17:50:14
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carbopack B
 Carrier: HELIUM
 Sample prog: eto-100.tem
 Components: eto2-100.cpt
 Data file: 2SterAti-2016-1E06.CHR (c:\peak359)
 Sample: Ceilcote Scrubber Outlet
 Operator: D. Kremer




Component	Retention	Area	External Units
aldehyde	0.683	254.9205	0.0000
		254.9205	0.0000

Client: Sterigenics - Atlanta
 Run ID: Run#1Exh
 Date: 03/17/2016 17:53:04
 Method: Direct Injection
 Injection: CHANNEL 2 - PID
 Column: 1% SP-1000, Carbopack B
 Carrier: HELIUM
 Temp: eto-100.tem
 Program: eto2-100.cpt
 Data file: 2SterAtt-2016-1E07.CHR (c:\peak359)
 Sample: Ceilcote Scrubber Outlet
 Operator: D. Kremer




Component	Retention	Area	External Units
Water / Air	0.016	13.0010	0.0000
Water H2O	0.400	47.5260	0.0000
Acetylene	0.583	9.9500	0.0000
	70.4770	0.0000	0.0000

Client: Sterigenics - Atlanta
 Client ID: Run#1Exh
 Date: 03/17/2016 17:57:16
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carbopack B
 Carrier: HELIUM
 Rep. prog: eto-100.tem
 Components: eto2-100.cpt
 Data file: 2SterAtt-2016-1E08.CHR (c:\peak359)
 Sample: Ceilcote Scrubber Outlet
 Operator: D. Kremer




Component	Retention	Area	External Units
Water H2O	0.400	42.6290	0.0000
Acetaldehyde	0.583	14.4550	0.0000
		57.0840	0.0000

Client: Sterigenics - Atlanta
 Run ID: Run#1Exh
 Date: 03/17/2016 17:58:30
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carbopack B
 Carrier: HELIUM
 Sample: Ceilcote Scrubber Outlet
 Operator: D. Kremer



Component	Retention	Area	External Units
Cl / Air	0.066	2.4960	0.0000
dehyde	0.600	113.3410	0.0000
	0.916	95.7085	0.0000
		211.5455	0.0000


Client: Sterigenics - Atlanta
 Client ID: Run#1Exh
 Analysis date: 03/17/2016 17:59:51
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carbopack B
 Carrier: HELIUM
 Temp. prog: eto-100.tem
 Components: eto2-100.cpt
 Data file: 2SterAti-2016-1E10.CHR (c:\peak359)
 Sample: Ceilcote Scrubber Outlet
 Operator: D. Kremer



Component	Retention	Area	External Units
Water H2O	0.450	174.1750	0.0000
		174.1750	0.0000

Client: Sterigenics - Atlanta
 Run ID: Run#1Exh
 Date: 03/17/2016 18:01:23
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carbopack B
 Carrier: HELIUM
 File: eto-100.tem
 Components: eto2-100.cpt
 Data file: 2SterAtI-2016-1E11.CHR (c:\peak359)
 Sample: Cellicote Scrubber Outlet
 Operator: D. Kremer


256.000 External Unit



Component	Retention	Area	External Units
Vol / Air	0.266	76.9670	0.0000
	0.966	5.8720	0.0000
		82.8390	0.0000


APPENDIX C
Run#2 Chromatograms

Client: Sterigenics - Atlanta
 Client ID: Run#2Exh
 Analysis date: 03/18/2016 11:20:04
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, CarboPack B
 Carrier: HELIUM
 Temp: prog: eto-100.tem
 Components: eto2-100.cpt
 Data file: 2SterAtl-2016-2E01.CHR (c:\peak359)
 Sample: Ceilcote Scrubber Outlet
 Operator: D. Kremer




Component	Retention	Area	External Units
ed Vol / Air	0.016	5.7145	0.0000
ient H2O	0.400	72.0530	0.0000
aldehyde	0.600	82.3875	0.0000
		160.1550	0.0000

Client: Sterigenics - Atlanta
 Client ID: Run#2Exh
 Analysis date: 03/18/2016 11:21:21
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carbopack B
 Carrier: HELIUM
 Temp. prog: eto-100.tem
 Components: eto2-100.cpt
 Data file: 2SterAtt-2016-2E02.CHR (c:\peak359)
 Sample: Ceilcote Scrubber Outlet
 Operator: D. Kremer




Component	Retention	Area	External Units
Vol / Air	0.116	17.3180	0.0000
aldehyde	0.600	154.5910	0.0000
		171.9090	0.0000

Client: Sterigenics - Atlanta
 Client ID: Run#2Exh
 Analysis date: 03/18/2016 11:22:28
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carbopack B
 Carrier: HELIUM
 Sample prog: eto-100.tem
 Components: eto2-100.cpt
 Data file: 2SterAtl-2016-2E03.CHR (c:\peak359)
 Sample: Ceilcote Scrubber Outlet
 Operator: D. Kremer




Component	Retention	Area	External Units
Vol / Air	0.016	8.1035	0.0000
ent H2O	0.416	10.1950	0.0000
ene Oxide	0.483	0.2315	0.1734 ppm
dehyde	0.733	15.2105	0.0000
		33.7405	0.1734

Client: Sterigenics - Atlanta
 Client ID: Run#2Exh
 Analysis date: 03/18/2016 11:23:45
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carbopack B
 Carrier: HELIUM
 Amp. prog: eto-100.tem
 Components: eto2-100.cpt
 Data file: 2SterAti-2016-2E04.CHR (c:\peak359)
 Sample: Cellcote Scrubber Outlet
 Operator: D. Kremer




Component	Retention	Area	External Units
Vol / Air	0.033	0.0035	0.0000
ent H2O	0.416	397.9920	0.0000
		397.9955	0.0000

Client: Sterigenics - Atlanta
 Client ID: Run#2Exh
 Analysis date: 03/18/2016 11:24:48
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carbopack B
 Carrier: HELIUM
 Sample prog: eto-100.tem
 Components: eto2-100.cpt
 Data file: 2SterAti-2016-2E05.CHR (c:\peak359)
 Sample: Celicote Scrubber Outlet
 Operator: D. Kremer




Component	Retention	Area	External Units
Acetaldehyde	0.016	6.9300	0.0000
Acetaldehyde	0.600	473.1440	0.0000
		480.0740	0.0000

Client: Sterigenics - Atlanta
 Client ID: Run#2Exh
 Date: 03/18/2016 11:27:04
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carboxpack B
 Carrier: HELIUM
 Sample: Ceilcote Scrubber Outlet
 Operator: D. Kremer



Component	Retention	Area	External Units
aldehyde	0.733	317.9725	0.0000
		317.9725	0.0000


Client: Sterigenics - Atlanta
 Client ID: Run#2Exh
 Analysis date: 03/18/2016 11:28:19
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carbopack B
 Carrier: HELIUM
 Sample prog: eto-100.tem
 Components: eto2-100.cpt
 Data file: 2SterAtt-2016-2E07.CHR (c:\peak359)
 Sample: Ceilcote Scrubber Outlet
 Operator: D. Kremer




Component	Retention	Area	External Units
Vol / Air	0.016	9.0340	0.0000
aldehyde	0.583	479.8670	0.0000
		488.9010	0.0000

Client: Stergenics - Atlanta
 Client ID: Run#2EXh
 Analysis date: 03/18/2016 11:29:33
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, CarboPack B
 Carrier: HELIUM
 Temp. prog: eto-100.tem
 Components: eto2-100.cpt
 Data file: 2SterAt-2016-2E08 CHR (c:\peak359)
 Sample: Cellcote Scrubber Outlet
 Operator: D. Kremer


Component	Retention	Area	External Units
aldehyde	0.583	551.0445	0.0000
		551.0445	0.0000



Client: Sterigenics - Atlanta
 Client ID: Run#2Exh
 Analysis date: 03/18/2016 11:30:48
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carbo-pack B
 Carrier: HELIUM
 Temp. prog: eto-100.tem
 Components: eto2-100.cpt
 Data file: 2SterAll-2016-2E09.CHR (c:\peak359)
 Sample: Cellcote Scrubber Outlet
 Operator: D. Kremer




Client: Sterigenics - Atlanta
 Client ID: Run#2EXh
 Analysis date: 03/18/2016 11:32:01
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carbowack B
 Carrier: HELIUM
 Temp. prog: eto-100.tem
 Components: eto2-100.cpt
 Data file: 2SterAt-2016-2E10.CHR (c:\peak359)
 Sample: Cellcote Scrubber Outlet
 Operator: D. Kremer




Component	Retention	Area	External Units
Vol / Air	0.016	0.0075	0.0000
Ident H2O	0.400	598.4050	0.0000
		598.4125	0.0000

Client: Stengenic - Atlanta
 Client ID: Run#2E11
 Analysis date: 03/18/2016 11:33:17
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carboxpack B
 Carrier: HELIUM
 Temp. prog: eto-100.tem
 Components: eto2-100.cpt
 Data file: 2SterAt-2016-2E11 CHR (c:\peak359)
 Sample: Cellcote Scrubber Outlet
 Operator: D. Kremer



Component	Retention	Area	External Units
Vol / Air	0.016	0.0875	0.0000
Water H2O	0.400	466.3270	0.0000
		466.4145	0.0000


Client: Stergenics - Atlanta
 Part ID: Run#25x
 Date: 03/18/2016 11:34:40
 Method: Direct Injection
 Injection: CHANNEL 2 - PID
 Column: 1% SP-1000, Carboxpack B
 Carrier: HELIUM
 Program: eto-100.tem
 Sample: eto2-100.cpt
 Data file: 2SterAt-2016-2E12.CHR (c:\peak359)
 Sample: Cellcoite Scrubber Outlet
 Operator: D. Kremer




Retention	Area	External Units
0.166	66.50300	0.0000
0.400	100.0960	0.0000
0.783	0.0945	0.0000
0.850	0.0635	0.0000
166.7570		0.0000

Client ID: Run#2EXh
 Client: Organics - Quantia
 Method date: 03/18/2016 11:35:50
 Method: Direct Injection
 Injection: CHANNEL 2 - PID
 Column: 1% SP-1000, Carboxpack B
 Carrier: HELIUM
 Temp. prog: eto-100.tem
 Components: eto2-100.cpt
 Data file: 2016-2E13.CHR (c:\peak359)
 Sample: Cellcote Scrubber Outlet
 Operator: D. Kremer

Component	Retention	Area	External Units
H2O	0.400	362.5100	0.0000
		362.5100	0.0000



Client: Stergenics - Atlanta
 Client ID: Run#2EXH
 Analysis date: 03/18/2016 11:37:03
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, CarboPack B
 Carrier: HELIUM
 Temp. prog: eto-100.tem
 Components: eto2-100.cpt
 Data file: 2SterAII-2016-2E14.CHR (c:\peak359)
 Sample: Cellcote Scrubber Outlet
 Operator: D. Kremer




Component	Retention	Area	External Units
Vol / Air	0.016	0.1240	0.0000
Went H2O	0.416	411.0930	0.0000
		411.2170	0.0000

ECSI

Run#3 Chromatograms

APPENDIX D

Client: Stengentics - Atlanta
 Client ID: Run#3EXh
 Analysis date: 03/18/2016 12:04:38
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carboxpack B
 Carrier: HELIUM
 Temp. prog: eto-100.tem
 Components: eto2-100.cpt
 Data file: 2SterA-2016-3E01 CHR (c:\peak359)
 Sample: Cellcote Scrubber Outlet
 Operator: D. Kremer




Component	Retention	Area	External Units
etol / Air	0.300	0.0470	0.0000
Water H2O	0.416	21.8050	0.0000
Acetaldehyde	0.616	21.0575	0.0000
		42.9095	0.0000


Client: Stengenic - Atlanta
 Client ID: Run#3Exh
 Analysis date: 03/18/2016 12:05:47
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carboxpack B
 Carrier: HELIUM
 Sample: Cellcote Scrubber Outlet
 Operator: D. Kremer

Component: eto-100.tem
 Data file: ZSterAt-2016-3E02.CHR (c:\peak359)

Component	Retention	Area	External Units
etoh / Air	0.016	2.4840	0.0000
H2O	0.450	23.9350	0.0000
aldehyde	0.616	42.2580	0.0000
		68.6770	0.0000




Client: Stergenics - Atlanta
 Client ID: Run#3EXh
 Analysis date: 03/18/2016 12:07:02
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carpopack B
 Carrier: HELIUM
 Temp. prog: etc-100.tem
 Components: etc2-100.cpt
 Data file: 2SterAt-2016-3E03.CHR (c:\peak359)
 Sample: Ceilcofe Scrubber Outlet
 Operator: D. Kremer



Component	Retention	Area	External Units
Hydro	0.600	111.6350	0.0000
Form H2O	0.400	59.5080	0.0000
Vol / Air	0.016	0.0300	0.0000


Client: Stergenics - Atlanta
 Client ID: Run#3ECh
 Sys date: 03/18/2016 12:08:21
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Caropack B
 Carrier: HELIUM
 Temp. prog: eto-100.tem
 Components: eto2-100.cpt
 Data file: 2SterAtt-2016-3E04.CHR (c:\peak59)
 Sample: Cellcote Scrubber Outlet
 Operator: D. Kremer



Component	Retention	Area	External Units
Component H2O	0.400	780.6905	0.0000
		780.6905	0.0000

Client: Stengentics - Atlanta
 Client ID: Run#3EXh
 Date: 03/18/2016 12:10:33
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carboxack B
 Carrier: HELIUM
 Inj. prog: eto-100.tem
 Components: eto2-100.cpt
 Data file: 2SterAtt-2016-3E05 CHR (c:\peak359)
 Sample: Cetecote Scrubber Outlet
 Operator: D. Kremer


Component	Retention	Area	External Units
N2 / Air	0.166	14.0700	0.0000
H2O	0.466	684.3665	0.0000
		698.4365	0.0000



Abundance
 256.000
 0.0000
 0.0000


Client: Stergenics - Atlanta
 Client ID: Run#33Exh
 Analysis date: 03/18/2016 12:12:21
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carboxpack B
 Carrier: HELIUM
 Temp. prog: eto2-100.tem
 Components: 2SterA-2016-3E06.CHR (c:\peak359)
 Data file: 2SterA-2016-3E06.CHR (c:\peak359)
 Sample: Cellcote Scrubber Outlet
 Operator: D. Kremer

Component	Retention	Area	External Units
Water H2O	0.383	20.3570	0.0000
Acetylene	0.600	2.8470	0.0000
	0.833	13.8355	0.0000
	37.0395		0.0000




Client: Stengentics - Atlanta
 Client ID: Run#3EXh
 Analysis date: 03/18/2016 12:14:17
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carboxpack B
 Carrier: HELIUM
 Temp. prog: eto-100.tem
 Components: eto2-100.cpt
 Data file: 2SterAt-2016-3E07 CHR (c:\peak359)
 Sample: Ceilcote Scrubber Outlet
 Operator: D. Kremer

Component	Retention	Area	External Units
et H2O	0.400	665.0985	0.0000
		665.0985	0.0000




Client: Stergenics - Atlanta
 Client ID: Run#3Exh
 Analysis date: 03/18/2016 12:15:53
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carbo-pack B
 Carrier: HELIUM
 Temp. prog: eto-100.tem
 Components: eto2-100.cpt
 Data file: ZSterA-2016-3E08.CHR (c:\peak359)
 Sample: Cellcote Scrubber Outlet
 Operator: D. Kremer

Component	Retention	Area	External	Units
Vol / Air	0.166	19.3480	0.0000	
		19.3480	0.0000	




Client: Sterigenics - Atlanta
 Client ID: Run#3E3h
 Analysis date: 03/18/2016 12:17:52
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carbo-pack B
 Carrier: HELIUM
 Temp. prog: eto-100.tem
 Components: eto2-100.cpt
 Data file: 2SterAtt-2016-3E09.CHR (c:\peak359)
 Sample: Celicote Scrubber Outlet
 Operator: D. Kremer

Component	Retention	Area	External Units
Water H2O	0.433	506.4860	0.0000
		506.4860	0.0000




Client: Sterigenics - Atlanta
 Client ID: Run#3Exh
 Analysis date: 03/18/2016 12:20:21
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carbo-pack B
 Carrier: HELIUM
 Temp. prog: eto-100.tem
 Components: eto2-100.cpt
 Data file: 2SterAtt-2016-3E10.CHR (c:\peak359)
 Sample: Cellcote Scrubber Outlet
 Operator: D. Kremer




Component	Retention	Area	External Units
H2O	0.400	31.2910	0.0000
Methyde	0.616	10.4720	0.0000
	0.816	3.9310	0.0000
	45.6940		0.0000

Client: Stergenics - Atlanta
 Client ID: Run#3Exh
 Analysis date: 03/18/2016 12:21:34
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carboxpack B
 Carrier: HELIUM
 Temp. prog: etc-100.tem
 Components: etc2-100.cpt
 Data file: 2SterAtt-2016-3E11 CHR (c:\peak359)
 Sample: Cellcote Scrubber Outlet
 Operator: D. Kremer



Component	Retention	Area	External Units
H2O	0.366	698.7945	0.0000
		698.7945	0.0000

Client: Stergenics - Atlanta
 Client ID: Run#3Exh
 Analysis date: 03/18/2016 12:22:52
 Method: Direct Injection
 Description: CHANNEL 2 - PID
 Column: 1% SP-1000, Carboxpack B
 Carrier: HELIUM
 Temp. prog: eto-100.tem
 Components: eto2-100.cpt
 Data file: ZSierAt-2016-3E12.CHR (c:\peak359)
 Sample: Cellcote Scrubber Outlet
 Operator: D. Kremer



Component	Retention	Area	External Units
Hyd / Air	0.166	29.2670	0.0000
Hydride	0.583	153.5495	0.0000
		182.8165	0.0000

Field Data and Calculation Worksheets

APPENDIX E

ECSI, Inc.

Ethylene Oxide Mass Emissions Data and Calculations

Run #2 (Cellcote Scrubber Outlet) - Chamber #8

Sterigenics, Inc. - Atlanta, GA

March 18, 2016

DeltaP	SqrDeltaP	Temp (F)	ppm Eto	1-%H2O =	mw =	stack area =	press =	Tstd =	Pstd =	Cp =	Kp =	Velocity =	Flow =	MWeto =	MolVol =	ppmv/ft3 =	Eto Mass Flow =	Average =	Eto Emissions =
0.01	0.1000	98	0.01	0.97	28.54	0.785	29.05	528	29.92	0.99	85.49	6.2	262	44.05	385.32	1000000	0.000006	0.0080	0.0000141
0.015	0.1225	88	0.01															0.0025	0.000141
0.02	0.1414	88	0.17															0.005	0.000141
0.015	0.1225	88	0.01															0.0025	0.000141
0.01	0.1000	97	0.01															0.0025	0.000141
0.0075	0.0866	99	0.01															0.0025	0.000141
0.005	0.0707	94	0.01															0.0025	0.000141
0.005	0.0707	92	0.01															0.0025	0.000141
0.005	0.0707	91	0.01															0.0025	0.000141
0.005	0.0707	92	0.01															0.0025	0.000141
0.005	0.0707	94	0.01															0.0025	0.000141
0.005	0.0707	99	0.01															0.0025	0.000141
0.005	0.0500	96	0.01															0.0025	0.000141
0.0025	0.0500	95	0.01															0.0025	0.000141
0.0080	0.0896	93.8	0.0214															0.0080	0.000141
																			0.000141

INLET CALCULATION:

Pre-Evac:	V =	1469	ft3	Post-Evac:	V =	1469	ft3
	P =	17.3	in Hg Abs		P =	2.0	in Hg Abs
	T =	129	degF		T =	129	degF
	R =	10.73			R =	10.73	
	mw =	44.05			mw =	44.05	
	lbs Eto @ 100% =	86.83	lbs		lbs Eto @ 100% =	10.04	lbs

Initial Eto = Scale Wt. = 70.0 lbs
 % Eto @ Chamber = Scale Wt. / lbs Eto @ 100% (Pre) = 80.6 %
 Final Eto = % Eto @ Chamber X lbs Eto @ 100% (Post) = 8.1 lbs
 INLET ETO = Initial Eto - Final Eto = 61.9 lbs

CONTROL EFFICIENCY = 99.999977 %

ECSI - VELOCITY TRAVERSE DATA

Client: Sterigenics, Inc. Run #: 1 Date: 3/17/2016 Port Sketch: _____
 Location: Atlanta, Georgia Probe Type: Std. Baro Press: 29.05
 Source: Cellcote Packed Tower Scrubber Outlet Stack I.D.: 12 in. Static Press: -0.01

Port 1

Port 2

Inches From Port	Point#	Delta P				Stack Temp (F)	Cyclonic Angle	Point#	Delta P				Stack Temp (F)	Cyclonic Angle
		Low	High	Average	Sq Root				Low	High	Average	Sq Root		
0.4	1	0.0025	0.0025	0.0025	0.0500	79	0	1	0.0025	0.0025	0.0025	0.0500	79	0
1.25	2	0.0025	0.005	0.00375	0.0612	79	0	2	0.005	0.005	0.005	0.0707	79	0
2.3	3	0.005	0.0075	0.00625	0.0791	79	0	3	0.005	0.0075	0.00625	0.0791	79	0
3.9	4	0.0075	0.0075	0.0075	0.0866	79	0	4	0.0075	0.01	0.00875	0.0935	80	0
8.1	5	0.0075	0.01	0.00875	0.0935	79	0	5	0.0075	0.0075	0.0075	0.0866	80	0
9.7	6	0.005	0.0075	0.00625	0.0791	79	0	6	0.005	0.0075	0.00625	0.0791	80	0
10.75	7	0.005	0.005	0.005	0.0707	80	0	7	0.0025	0.005	0.00375	0.0612	80	0
11.6	8	0.0025	0.0025	0.0025	0.0500	80	0	8	0.0025	0.0025	0.0025	0.0500	80	0
	9							9						
	10							10						
	11							11						
	12							12						
	13							13						
	14							14						
	15							15						
	16							16						
	17							17						
	18							18						
	19							19						
	20							20						
	21							21						
	22							22						
	23							23						
	24							24						

Average Values: 0.0053 0.0713 79.4 0.0

ETHYLENE OXIDE SOURCE TEST/CALIBRATION DATA

Client: Sterigenics - Atlanta

Source Tested: Cellcote scrubber/AT safe cell system Date: 3/17/16
3/18/16

PRE CALIBRATION									
		Audit Standard (48.8 ppmv) Result							
Inlet (FID)	Calibration Gas	1.10	10.1	100	1000	10080			
	Conc. (ppmv)	ppm EtO	ppm EtO	ppm EtO	ppm EtO	ppm EtO			
	Area Counts #1	23	199	19.1					
Area Counts #2	28	202	19.6						
Average Area	23	201	19.4						
		Audit Standard (48.8 ppmv) Result							
Outlet (PID)	Calibration Gas	1.10	10.1	100					
	Conc. (ppmv)	ppm EtO	ppm EtO	ppm EtO					
	Area Counts #1	154	131	126					
Area Counts #2	157	134	128						
Average Area	156	133	127						

48.8

48.8

Ch. 3 Ch. 8 Ch. 11
 Run #1 Run #2 Run #3
 179/1803 118/1140 225/2308
~~2229/2335~~ ~~210/2325~~ ~~210/2325~~
 Backflow stop Stop Stop Stop
 P: 29.05 %H₂O: 3
 EtO Usage (lbs/yr): Cycles Per Week:

POST CALIBRATION									
		Audit Standard (48.8 ppmv) Result							
Inlet (FID)	Calibration Gas	1.10	10.1	100	1000	10080			
	Conc. (ppmv)	ppm EtO	ppm EtO	ppm EtO	ppm EtO	ppm EtO			
	Area Counts #1								
Area Counts #2									
Average Area									
		Audit Standard (48.8 ppmv) Result							
Outlet (PID)	Calibration Gas	1.10	10.1	100					
	Conc. (ppmv)	ppm EtO	ppm EtO	ppm EtO					
	Area Counts #1								
Area Counts #2									
Average Area									

ECSI

Gas Certifications

APPENDIX F

CERTIFIED WORKING CLASS

Single-Certified Calibration Standard

Scott Specialty Gases

2500 CAJON BLVD., SAN BERNARDINO, CA 92411

Phone: 909-887-2577 Fax: 909-887-0549

CERTIFICATE OF ACCURACY: Certified Working Class Calibration Standard

Product Information

Part No. 02-E7184-001
Item No. 02020001310TCL

P.O. No. VBL-3 KREMER

Cylinder Number: CA44448

Cylinder Size: CL

Certification Date: 14Apr2014

CERTIFIED CONCENTRATION

Component Name

ETHYLENE OXIDE

NITROGEN

TRACEABILITY

Traceable To

Scott Reference Standard

Concentration

(Moles)

1.10 PPM

SALANCE

Accuracy

(+/-%)

5

Customer

ECSI, INC.

PO BOX 848

SAN CLEMENTE, CA 92672

APPROVED BY:

ME

DATE:

2-14-14

CERTIFIED WORKING CLASS

Single-Certified Calibration Standard

Scott Specialty Gases

2500 CALON BLVD, SAN BERNARDINO, CA 92411

PHONE: 909-887-2871 FAX: 909-887-0549

CERTIFICATE OF ACCURACY: Certified Working Class Calibration Standard

Product Information

Project No: 02-57164-008
Item No: 0203001320TCL
Mfg. No: VSL-B-KLEMER

Cylinder Number: 01M002232

Cylinder Size: CL

Expiration Date: 11/20/14

CERTIFIED CONCENTRATION

Component Name

Ethylene Oxide

Nitrogen

TRACEABILITY

Traceable To

Scott Reference Standard

Concentration

(Moles)

10.1

ppm

EALANCE

Accuracy

(±%)

5

Customer

ECSS, INC

PO BOX 848

SAN CLEMENTE, CA 92672

APPROVED BY

CERTIFIED WORKING CLASS

Single-Certified Calibration Standard

Scott Specialty Gases



100 CALON BLVD., SAN BERNARDINO, CA 92411 Phone: 909-887-2511 Fax: 909 887 0549

CERTIFICATE OF ACCURACY: Certified Working Class Calibration Standard

Product Information
Project No: 02-87184-004
Item No: 02020013801C
A.D. No: VBL-D-KRBER9

Cylinder Number: C1M011385
Cylinder Size: CL
Certification Date: 11/20/2011

CERTIFIED CONCENTRATION

Component Name	Concentration (Moles)	Accuracy (±%)
ETHYLENE OXIDE	100 PPM	± 5
NITROGEN	BALANCE	± 5

TRACEABILITY

Traceable To
Scott Reference Standard

APPROVED BY: E. M. Williams DATE: _____

CERTIFIED WORKING CLASS

Single-Certified Calibration Standard

Scott Specialty Gases



2500 CALON BLVD., SAN BRANDBINO, CA 92411

PHONE: 509-887-2877 FAX: 509-887-0549

CERTIFICATE OF ACCURACY: Certified Working Class Calibration Standard

Product Information

Project No.: 02-57164-005
Item No.: 02020001949TC
P.O. No.: VBL-D KREMER
Cylinder Number: CLM002810
Cylinder Size: CL
Certification Date: 4/14/2014

CERTIFIED CONCENTRATION

Component Name

ETHYLENE OXIDE
NITROGEN

TRACEABILITY

Traceable To

Scott Reference Standards

Concentration

1.0001

PPM

BALANCE

5

Accuracy
(+/-%)

Customer

EOSL INC
PO BOX 848
SAN CLEMENTE, CA 92672

APPROVED BY:

DATE

4-14-14

CERTIFIED WORKING CLASS

Single-Certified Calibration Standard

Scott Specialty Gases



1500 CALSON BLVD., SAN BERNARDINO, CA 92411

Phone: 909-887-2571 Fax: 909-887-0849

CERTIFICATE OF ACCURACY: Certified Working Class Calibration Standard

Product Information

Project No.: 02-57184-008

Form No.: 02020001S407C1

P.O. No.: V6L-D KREMER

Cylinder Number: CLM008787

Cylinder Size: G1

Certification Date: 1A9D2014

CERTIFIED CONCENTRATION

Component Name

ETHYLENE OXIDE

NITROGEN

TRACEABILITY

Traceable To

Scott Reference Standard

Concentration

(Moles)

10.080

PPM

SALANCE

5

Accuracy

(±1%)

Customer

ECOL, INC.

PO BOX 818

SAN CLEMENTE, CA 92672

APPROVED BY:

[Handwritten Signature]

DATE:

6-12-12

ECSI

G-1

Process/Parametric Monitoring Data

APPENDIX G

PRESS TEMP (DEG F) RH WVP * GAS
 ALARMS & MESSAGES
 ACTION TAKEN

TIME	INHG	Avg	% GAS	PHASE	ELAPSED	CYCLE	VALUE
13:14	13.6	131		PHASE	0:13		
13:16	15.7	131		PHASE	0:13		
13:17	17.2	131		PHASE	0:13		
MAX:	17.2	131		PHASE	0:13		
MIN:	3.5	130		PHASE	0:13		
13:18				PHASE	0:13		
13:19				PHASE	0:13		
13:20				PHASE	0:13		
13:21				PHASE	0:13		
13:22				PHASE	0:13		
13:23				PHASE	0:13		
13:24				PHASE	0:13		
13:25				PHASE	0:13		
13:26				PHASE	0:13		
13:27				PHASE	0:13		
13:28				PHASE	0:13		
13:29				PHASE	0:13		
13:30				PHASE	0:13		
13:31				PHASE	0:13		
13:32				PHASE	0:13		
13:33				PHASE	0:13		
13:34				PHASE	0:13		
13:35				PHASE	0:13		
13:36				PHASE	0:13		
13:37				PHASE	0:13		
13:38				PHASE	0:13		
13:39				PHASE	0:13		
13:40				PHASE	0:13		
13:41				PHASE	0:13		
13:42				PHASE	0:13		
13:43				PHASE	0:13		
13:44				PHASE	0:13		
13:45				PHASE	0:13		
13:46				PHASE	0:13		
13:47				PHASE	0:13		
13:48				PHASE	0:13		
13:49				PHASE	0:13		
13:50				PHASE	0:13		
13:51				PHASE	0:13		
13:52				PHASE	0:13		
13:53				PHASE	0:13		
13:54				PHASE	0:13		
13:55				PHASE	0:13		
13:56				PHASE	0:13		
13:57				PHASE	0:13		
13:58				PHASE	0:13		
13:59				PHASE	0:13		
14:00				PHASE	0:13		
14:01				PHASE	0:13		
14:02				PHASE	0:13		
14:03				PHASE	0:13		
14:04				PHASE	0:13		
14:05				PHASE	0:13		
14:06				PHASE	0:13		
14:07				PHASE	0:13		
14:08				PHASE	0:13		
14:09				PHASE	0:13		
14:10				PHASE	0:13		
14:11				PHASE	0:13		
14:12				PHASE	0:13		
14:13				PHASE	0:13		
14:14				PHASE	0:13		
14:15				PHASE	0:13		
14:16				PHASE	0:13		
14:17				PHASE	0:13		
14:18				PHASE	0:13		
14:19				PHASE	0:13		
14:20				PHASE	0:13		
14:21				PHASE	0:13		
14:22				PHASE	0:13		
14:23				PHASE	0:13		
14:24				PHASE	0:13		
14:25				PHASE	0:13		
14:26				PHASE	0:13		
14:27				PHASE	0:13		
14:28				PHASE	0:13		
14:29				PHASE	0:13		
14:30				PHASE	0:13		
14:31				PHASE	0:13		
14:32				PHASE	0:13		
14:33				PHASE	0:13		
14:34				PHASE	0:13		
14:35				PHASE	0:13		
14:36				PHASE	0:13		
14:37				PHASE	0:13		
14:38				PHASE	0:13		
14:39				PHASE	0:13		
14:40				PHASE	0:13		
14:41				PHASE	0:13		
14:42				PHASE	0:13		
14:43				PHASE	0:13		
14:44				PHASE	0:13		
14:45				PHASE	0:13		
14:46				PHASE	0:13		
14:47				PHASE	0:13		
14:48				PHASE	0:13		
14:49				PHASE	0:13		
14:50				PHASE	0:13		
14:51				PHASE	0:13		
14:52				PHASE	0:13		
14:53				PHASE	0:13		
14:54				PHASE	0:13		
14:55				PHASE	0:13		
14:56				PHASE	0:13		
14:57				PHASE	0:13		
14:58				PHASE	0:13		
14:59				PHASE	0:13		
15:00				PHASE	0:13		

STERILANT USED THIS PHASE: 50.8
 CYCLE TOTAL: 50.8
 GAS DWELL (ED) PHASE

TIME	INHG	Avg	% GAS	PHASE	ELAPSED	CYCLE	VALUE
17:18	17.2	131		PHASE	0:13		
17:19	17.1	130		PHASE	0:13		
17:20	17.1	130		PHASE	0:13		
17:21	17.1	130		PHASE	0:13		
17:22	17.1	130		PHASE	0:13		
17:23	17.1	130		PHASE	0:13		
17:24	17.1	130		PHASE	0:13		
17:25	17.1	130		PHASE	0:13		
17:26	17.1	130		PHASE	0:13		
17:27	17.1	130		PHASE	0:13		
17:28	17.1	130		PHASE	0:13		
17:29	17.1	130		PHASE	0:13		
17:30	17.1	130		PHASE	0:13		
17:31	17.1	130		PHASE	0:13		
17:32	17.1	130		PHASE	0:13		
17:33	17.1	130		PHASE	0:13		
17:34	17.1	130		PHASE	0:13		
17:35	17.1	130		PHASE	0:13		
17:36	17.1	130		PHASE	0:13		
17:37	17.1	130		PHASE	0:13		
17:38	17.1	130		PHASE	0:13		
17:39	17.1	130		PHASE	0:13		
17:40	17.1	130		PHASE	0:13		
17:41	17.1	130		PHASE	0:13		
17:42	17.1	130		PHASE	0:13		
17:43	17.1	130		PHASE	0:13		
17:44	17.1	130		PHASE	0:13		
17:45	17.1	130		PHASE	0:13		
17:46	17.1	130		PHASE	0:13		
17:47	17.1	130		PHASE	0:13		
17:48	17.1	130		PHASE	0:13		
17:49	17.1	130		PHASE	0:13		
17:50	17.1	130		PHASE	0:13		
17:51	17.1	130		PHASE	0:13		
17:52	17.1	130		PHASE	0:13		
17:53	17.1	130		PHASE	0:13		
17:54	17.1	130		PHASE	0:13		
17:55	17.1	130		PHASE	0:13		
17:56	17.1	130		PHASE	0:13		
17:57	17.1	130		PHASE	0:13		
17:58	17.1	130		PHASE	0:13		
17:59	17.1	130		PHASE	0:13		
18:00	17.1	130		PHASE	0:13		
18:01	17.1	130		PHASE	0:13		
18:02	17.1	130		PHASE	0:13		
18:03	17.1	130		PHASE	0:13		
18:04	17.1	130		PHASE	0:13		
18:05	17.1	130		PHASE	0:13		
18:06	17.1	130		PHASE	0:13		
18:07	17.1	130		PHASE	0:13		
18:08	17.1	130		PHASE	0:13		
18:09	17.1	130		PHASE	0:13		
18:10	17.1	130		PHASE	0:13		
18:11	17.1	130		PHASE	0:13		
18:12	17.1	130		PHASE	0:13		
18:13	17.1	130		PHASE	0:13		
18:14	17.1	130		PHASE	0:13		
18:15	17.1	130		PHASE	0:13		
18:16	17.1	130		PHASE	0:13		
18:17	17.1	130		PHASE	0:13		
18:18	17.1	130		PHASE	0:13		
18:19	17.1	130		PHASE	0:13		
18:20	17.1	130		PHASE	0:13		
18:21	17.1	130		PHASE	0:13		
18:22	17.1	130		PHASE	0:13		
18:23	17.1	130		PHASE	0:13		
18:24	17.1	130		PHASE	0:13		
18:25	17.1	130		PHASE	0:13		
18:26	17.1	130		PHASE	0:13		
18:27	17.1	130		PHASE	0:13		
18:28	17.1	130		PHASE	0:13		
18:29	17.1	130		PHASE	0:13		
18:30	17.1	130		PHASE	0:13		
18:31	17.1	130		PHASE	0:13		
18:32	17.1	130		PHASE	0:13		
18:33	17.1						

TIME INHGA PRESS TEMP (DEG F) RH VAP % GAS ALARMS & MESSAGES ACTION TAKEN

HUMIDIFICATION (PRESSURE) PHASE

4:15	2.0	131	112
4:16	3.5	131	112
MAX:	3.5	131	112
MIN:	2.0	131	112
PHASE 0:00 PHASE ELAPSED 0:00 CYCLE 0:37			

HUMIDITY DWELL (PRESS) PHASE

4:16	3.6	131	112
4:18	3.8	131	112
4:20	3.8	130	111
4:22	3.8	129	110
4:24	3.8	128	109
4:26	3.8	128	108
4:26	3.5	128	108
MAX:	3.8	131	112
MIN:	3.5	128	108
PHASE 0:10 PHASE ELAPSED 0:10 CYCLE 0:47			

GAS A (EO) PHASE

4:26	3.5	128	108
4:28	5.5	128	107
4:30	7.5	128	107
4:32	9.5	129	107
4:34	11.5	129	106
4:36	13.6	130	106
4:38	15.5	130	106
4:39	17.2	129	106
MAX:	17.2	130	108
MIN:	3.5	128	106
PHASE 0:13 PHASE ELAPSED 0:13 CYCLE 1:00			

GAS DWELL (EO) PHASE

STERILANT USED THIS PHASE: 70.0% CYCLE TOTAL: 70.0%

4:40	17.1	129	106
4:42	17.1	128	107
4:44	17.2	128	108
4:46	17.2	128	108

11. Phase » EO Inject by Pressure

Date / Time	Avg. Chamber Gas Temp	Current Pressure	AT-EO [mg/l]	EO Weight [lbs]	Used
3/18/2016 13:59:04	132.2	6.4	173.0	70	
14:00:04	132.3	7.3	224.8	86	
14:00:53	132.4	8.0	263.8	98	
14:00:57	132.4	8.1	267.1	99	

Step Gas Usage

ID	Gas	Drum S/N	Lot Number	Usage Value
56	EO	E000086	U TLX902056B16	99lbs

Device	Summary	Summary Value	Specified Range
Current Phase Time	Final	00:05:54	00:01:00 - 00:30:00
Elapsed Phase Time	Final	00:05:54	n/a - N/A
Tempave	Min	131.6	120.0 - 140.0 °F
Tempave	Max	132.4	120.0 - 140.0 °F
PCurr	Final	8.1	7.0 - 9.5 inHg
PCurr	Final	8.1	n/a - 9.0 inHg
AT-EO	Final	267.1	N/A - N/A mg/l
EO Weight Used	Final	99	N/A - N/A lbs

12. Phase » Static Dwell

Date / Time	Avg. Chamber Gas Temp	Current Pressure
3/18/2016 14:00:58	132.4	8.1
14:01:58	132.4	8.0
14:02:58	132.4	8.0
14:03:58	132.2	8.0
14:04:58	132.1	8.0
14:05:57	132.1	8.0
14:05:58	132.1	8.0
14:05:58	132.1	8.0

Device	Summary	Summary Value	Specified Range
Current Phase Time	Final	00:05:01	00:04:00 - 00:06:00
Elapsed Phase Time	Final	00:05:01	n/a - N/A
Tempave	Min	132.1	120.0 - 140.0 °F
Tempave	Max	132.4	120.0 - 140.0 °F
PCurr	Min	8.0	7.0 - 9.0 inHg
PCurr	Min	8.0	n/a - N/A inHg
PCurr	Max	8.1	7.0 - 9.0 inHg
PCurr	Max	8.1	n/a - N/A inHg

13. Phase » EO Inject by Pressure

Device	Summary	Summary Value	Specified Range
Current Phase Time	Final	00:05:01	00:04:00 - 00:06:00
Elapsed Phase Time	Final	00:05:01	n/a - N/A
Tempave	Min	132.1	120.0 - 140.0 °F
Tempave	Max	132.4	120.0 - 140.0 °F
PCurr	Min	8.0	7.0 - 9.0 inHg
PCurr	Min	8.0	n/a - N/A inHg
PCurr	Max	8.1	7.0 - 9.0 inHg
PCurr	Max	8.1	n/a - N/A inHg

(Phase No.12)

14. Phase » EO Dwell with Nitrogen Makeup

Date / Time	Avg. Chamber Pressure	Current [inHg]	[mg/l] AT-EO	[mg/l] AT-AH	RH Sensor [%]
3/18/2016 16:01:40	132.0	13.0	578.4	34.5	N/A
16:02:40	132.1	13.0	578.4	34.5	N/A
16:03:26	132.1	13.0	578.4	34.5	N/A

Operator has requested abort

16:03:26	132.1	13.0	578.4	34.5	N/A
----------	-------	------	-------	------	-----

Device	Summary	Summary Value	Specified Range
Current Phase Time	Final	01:51:46	01:59:00 - 02:15:00
Elapsed Phase Time	Final	01:51:46	n/a - N/A
Tempave	Min	131.8	125.0 - 135.0 °F
Tempave	Max	132.8	125.0 - 135.0 °F
PCurr	Min	13.0	12.0 - 14.5 inHg
PCurr	Min	13.0	n/a - 14.0 inHg
PCurr	Max	13.1	12.0 - 14.5 inHg
PCurr	Max	13.1	n/a - 14.0 inHg
RH from AH Calc	Min	N/A	N/A - N/A %
RH from AH Calc	Max	N/A	N/A - N/A %
AT-AH	Min	33.6	N/A - N/A mg/l
AT-AH	Max	35.4	N/A - N/A mg/l
AT-EO	Min	545.5	N/A - N/A mg/l
AT-EO	Max	578.4	N/A - N/A mg/l

DVN Current Phase Time has a measurement value of 01:51:46. The min specified value is 01:59:00

15. Phase » Vacuum

Date / Time	Avg. Chamber Pressure	Current [inHg]	Gas Temp
3/18/2016 16:03:26	132.1	13.0	

SYS Cycle has entered EO Abort sequence while running Run ID 301852 at phase no. 14 - 0

OPR Chamber was placed in stop mode while running Run ID 301852 at phase no. 14 - 0

OPR Chamber was placed in run mode while running Run ID 301852 at phase no. 14 - 0

OPR Operator has requested run mode

16:09:26	131.5	8.9	
----------	-------	-----	--

16:09:19	131.5	8.9	
----------	-------	-----	--

16:09:07	131.5	8.9	
----------	-------	-----	--

16:09:02	131.5	8.9	
----------	-------	-----	--

16:08:26	131.5	8.9	
----------	-------	-----	--

16:08:11	131.5	9.0	
----------	-------	-----	--

16:07:26	131.6	9.9	
----------	-------	-----	--

16:06:26	131.7	10.7	
----------	-------	------	--

16:05:26	131.9	11.5	
----------	-------	------	--

16:04:26	132.0	12.5	
----------	-------	------	--

16:03:26	132.1	13.0	
----------	-------	------	--

Process Run Record

Run ID 301852

5. Phase » Vacuum

Date/Time	Avg. Current [F]	Chamber Pressure [inHg]	Gas Temp
16:09:29	131.5	9.0	

Operator has requested run mode

Chamber was placed in stop mode while running Run ID 301852 at phase no. 14 - 0

Operator has requested run mode

Chamber was placed in run mode while running Run ID 301852 at phase no. 14 - 0

Operator has requested run mode

Chamber was placed in stop mode while running Run ID 301852 at phase no. 14 - 0

Operator has requested run mode

Chamber was placed in run mode while running Run ID 301852 at phase no. 14 - 0

Operator has requested run mode

Chamber was placed in stop mode while running Run ID 301852 at phase no. 14 - 0

Operator has requested run mode

Chamber was placed in run mode while running Run ID 301852 at phase no. 14 - 0

Operator has requested run mode

Chamber was placed in stop mode while running Run ID 301852 at phase no. 14 - 0

Operator has requested run mode

Chamber was placed in run mode while running Run ID 301852 at phase no. 14 - 0

Operator has requested run mode

Chamber was placed in stop mode while running Run ID 301852 at phase no. 14 - 0

Operator has requested run mode

Chamber was placed in run mode while running Run ID 301852 at phase no. 14 - 0

Operator has requested run mode

Chamber was placed in stop mode while running Run ID 301852 at phase no. 14 - 0

Operator has requested run mode

Chamber was placed in run mode while running Run ID 301852 at phase no. 14 - 0

Operator has requested run mode

Chamber was placed in stop mode while running Run ID 301852 at phase no. 14 - 0

Operator has requested run mode

Chamber was placed in run mode while running Run ID 301852 at phase no. 14 - 0

Operator has requested run mode

Chamber was placed in stop mode while running Run ID 301852 at phase no. 14 - 0

Operator has requested run mode

Chamber was placed in run mode while running Run ID 301852 at phase no. 14 - 0

Operator has requested run mode

Chamber was placed in stop mode while running Run ID 301852 at phase no. 14 - 0

Operator has requested run mode

Chamber was placed in run mode while running Run ID 301852 at phase no. 14 - 0

Operator has requested run mode

Chamber was placed in stop mode while running Run ID 301852 at phase no. 14 - 0

Operator has requested run mode

Chamber was placed in run mode while running Run ID 301852 at phase no. 14 - 0

Operator has requested run mode

Chamber was placed in stop mode while running Run ID 301852 at phase no. 14 - 0

Cellcote Readings:

T2 = 186 - Inches

PH = 1.2


Glycol = 36.3%

AAT Readings:

T2 = 105 - Inches

PH = 0.9

Glycol = 36.8%

Recorded By: 
Date: 18-MAR-16